



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB104700

Analyst By : JIGNESH

Supervisor Review By : apatel

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	08/20/2019	16:10
2	CAL2	1	Water	NA	NA	20.1	7.01	08/20/2019	16:11
3	CAL3	1	Water	NA	NA	20.1	10.04	08/20/2019	16:12
4	ICV	1	Water	NA	NA	20.1	6.99	08/20/2019	16:14
5	CCV1	1	Water	NA	NA	20.2	2.01	08/20/2019	16:15
6	K4418-02	1	Water	NA	NA	24.8	7.45	08/20/2019	16:35
7	K4418-02DUP	1	Water	NA	NA	24.9	7.46	08/20/2019	16:36
8	CCV2	1	Water	NA	NA	20.1	12.02	08/20/2019	16:37

2B 10400

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K4418 WorkList ID : 130034 Department : Wet-Chemistry Date : 08-20-2019 16:22:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/20/2019	Water	K4418-02	Corrosivity	Cool 4 deg C	PSEG01	J11	CFVP-228L	08/20/2019	9040C

Date/Time 08.20.19 16:25
Received by: jo
Relinquished by: gmm

Date/Time 08.20.19 17:00
Received by: gmm
Relinquished by: jo