



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 8/22/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 14:10
In Date: 08/21/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:30
Out Date: 08/22/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLIDS-OVEN

QC:LB104714

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K4445-01	MH10K6	1	1.00	2.00	2.00	100.0	100% SOLIDS
K4445-02	MH10K7	2	1.00	2.00	2.00	100.0	100% SOLIDS
K4445-03	MH10K8	3	1.00	2.00	2.00	100.0	100% SOLIDS
K4445-04	MH10K9	4	1.00	2.00	2.00	100.0	100% SOLIDS
K4445-05	MH10M4	5	1.00	2.00	2.00	100.0	100% SOLIDS
K4445-06	MH10M5	6	1.00	2.00	2.00	100.0	100% SOLIDS
K4445-07	MH10M6	7	1.00	2.00	2.00	100.0	100% SOLIDS
K4445-08	MH10N4	8	1.00	2.00	2.00	100.0	100% SOLIDS
K4445-09	MH10N4D	9	1.00	2.00	2.00	100.0	100% SOLIDS
K4445-10	MH10N4S	10	1.00	2.00	2.00	100.0	100% SOLIDS
K4445-11	MH1150	11	1.00	2.00	2.00	100.0	100% SOLIDS
K4445-12	MH1151	12	1.00	2.00	2.00	100.0	100% SOLIDS
K4445-13	MH1152	13	1.00	2.00	2.00	100.0	100% SOLIDS
K4445-14	MH1153	14	1.00	2.00	2.00	100.0	100% SOLIDS
K4445-15	MH1172	15	1.00	2.00	2.00	100.0	100% SOLIDS
K4445-16	MH1173	16	1.00	2.00	2.00	100.0	100% SOLIDS
K4445-17	MH1174	17	1.00	2.00	2.00	100.0	100% SOLIDS
K4445-18	MH1175	18	1.00	2.00	2.00	100.0	100% SOLIDS
K4445-19	MH1176	19	1.00	2.00	2.00	100.0	100% SOLIDS
K4445-20	MH1177	20	1.00	2.00	2.00	100.0	100% SOLIDS
K4445-21	MH1178	21	1.00	2.00	2.00	100.0	100% SOLIDS
K4445-22	MH1179	22	1.00	2.00	2.00	100.0	100% SOLIDS

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$