



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 8/22/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 14:25
In Date: 08/21/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:35
Out Date: 08/22/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLIDS-OVEN

QC:LB104715

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K4446-01	MH10N5	1	1.00	2.00	2.00	100.0	100% SOLIDS
K4446-02	MH10N5D	2	1.00	2.00	2.00	100.0	100% SOLIDS
K4446-03	MH10N5S	3	1.00	2.00	2.00	100.0	100% SOLIDS
K4446-04	MH1180	4	1.00	2.00	2.00	100.0	100% SOLIDS
K4446-05	MH1181	5	1.00	2.00	2.00	100.0	100% SOLIDS
K4446-06	MH1182	6	1.00	2.00	2.00	100.0	100% SOLIDS
K4446-07	MH1183	7	1.00	2.00	2.00	100.0	100% SOLIDS
K4446-08	MH1184	8	1.00	2.00	2.00	100.0	100% SOLIDS
K4446-09	MH1186	9	1.00	2.00	2.00	100.0	100% SOLIDS
K4446-10	MH1187	10	1.00	2.00	2.00	100.0	100% SOLIDS
K4446-11	MH1188	11	1.00	2.00	2.00	100.0	100% SOLIDS
K4446-12	MH1189	12	1.00	2.00	2.00	100.0	100% SOLIDS
K4446-13	MH1190	13	1.00	2.00	2.00	100.0	100% SOLIDS
K4446-14	MH1191	14	1.00	2.00	2.00	100.0	100% SOLIDS
K4446-15	MH1192	15	1.00	2.00	2.00	100.0	100% SOLIDS
K4446-16	MH1193	16	1.00	2.00	2.00	100.0	100% SOLIDS
K4446-17	MH1194	17	1.00	2.00	2.00	100.0	100% SOLIDS
K4446-18	MH1195	18	1.00	2.00	2.00	100.0	100% SOLIDS
K4446-19	MH1196	19	1.00	2.00	2.00	100.0	100% SOLIDS
K4446-20	MH1197	20	1.00	2.00	2.00	100.0	100% SOLIDS
K4446-21	MH1198	21	1.00	2.00	2.00	100.0	100% SOLIDS

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$