



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104723

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	08/21/2019	15:15
2	CAL2	1	Water	NA	NA	20.1	7.00	08/21/2019	15:16
3	CAL3	1	Water	NA	NA	20.1	10.03	08/21/2019	15:17
4	ICV	1	Water	NA	NA	20.1	7.01	08/21/2019	15:20
5	CCV1	1	Water	NA	NA	20.1	2.01	08/21/2019	15:21
6	K4458-01	1	Water	NA	NA	23.3	5.21	08/21/2019	15:22
7	K4458-01DUP	1	Water	NA	NA	23.4	5.22	08/21/2019	15:24
8	CCV2	1	Water	NA	NA	20.1	12.02	08/21/2019	15:25

2B10423

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4458 WorkList ID : 130085 Department : Wet-Chemistry Date : 08-21-2019 15:06:13

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/19/2019	Water	K4458-01-17	pH	Cool 4 deg C	STER06	H13	POOL-WATER	08/19/2019	9040C

Date/Time 08.21.19 15:10
Received by: JA
Relinquished by: JA

Date/Time 08.21.19 16:00
Received by: cd
Relinquished by: JD