



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104726

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/21/2019	17:30
2	CAL2	1	Water	NA	NA	20.2	6.99	08/21/2019	17:31
3	CAL3	1	Water	NA	NA	20.2	10.02	08/21/2019	17:32
4	ICV	1	Water	NA	NA	20.2	7.01	08/21/2019	17:34
5	CCV1	1	Water	NA	NA	20.1	2.02	08/21/2019	18:15
6	K4465-01	1	Solid	20.02	20	24.4	5.47	08/21/2019	18:16
7	K4465-01DUP	1	Solid	20.03	20	24.5	5.48	08/21/2019	18:17
8	CCV2	1	Water	NA	NA	20.1	12.02	08/21/2019	18:20

LB 104726

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph k4465 WorkList ID : 130089 Department : Wet-Chemistry Date : 08-21-2019 16:19:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/21/2019	Solid	K4465-01, 15	pH	Cool 4 deg C	TULL02	H21	FK-PS-01-056	08/21/2019	9045D

Date/Time 08/21/19 16:20
Received by: Jn
Relinquished by: CB

Date/Time 08/21/19 17:25
Received by: CB
Relinquished by: Jn