



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104740

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	08/22/2019	12:20
2	CAL2	1	Water	NA	NA	20.3	7.00	08/22/2019	12:21
3	CAL3	1	Water	NA	NA	20.3	10.03	08/22/2019	12:22
4	ICV	1	Water	NA	NA	20.1	6.99	08/22/2019	12:24
5	CCV1	1	Water	NA	NA	20.1	2.01	08/22/2019	12:25
6	K4483-01	1	Water	NA	NA	25.5	5.32	08/22/2019	12:26
7	K4483-01DUP	1	Water	NA	NA	25.6	5.33	08/22/2019	12:27
8	CCV2	1	Water	NA	NA	20.4	12.02	08/22/2019	12:30

LB 104740

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4483 WorkList ID : 130131 Department : Wet-Chemistry Date : 08-22-2019 12:10:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/21/2019	Water	K4483-01-17	pH	Cool 4 deg C	STER01	I11	POOL-WATER	08/21/2019	SM 4500-H B

Date/Time: 08/22/2019 12:15
 Received by: Jn
 Relinquished by: Jn

Date/Time: 08/22/2019 13:00
 Received by: Jn
 Relinquished by: Jn