



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB104744

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP77554
Turbidity Calibration std, 0NTU	WP77548
Turbidity Calibration std, 20NTU	WP77550
Turbidity Calibration std, 40NTU	WP77549
Turbidity Calibration std, 1NTU	WP77553
Turbidity Calibration std, 10NTU	WP77551
Turbidity Calibration std, 5NTU	WP77552
10 NTU Standard 500 ml	W2297

Intercept: 0.0036

Slope: 0.964

Regression: 0.9997

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.280	08/22/2019	12:30
2	CAL2	1	1	0.996	08/22/2019	12:32
3	CAL3	5	1	4.870	08/22/2019	12:35
4	CAL4	10	1	9.640	08/22/2019	12:40
5	CAL5	20	1	18.600	08/22/2019	12:43
6	CAL6	40	1	38.900	08/22/2019	12:46



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB104744

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.550	9.903	08/22/2019	12:51
2	ICB		1	0.196	0.200	08/22/2019	12:59
3	CCV1	10	1	9.610	9.965	08/22/2019	13:04
4	CCB1		1	0.201	0.205	08/22/2019	13:08
5	LB104744BL		1	0.199	0.203	08/22/2019	13:13
6	K4478-01		1	13.600	14.104	08/22/2019	13:18
7	K4478-01DUP		1	14.000	14.519	08/22/2019	13:25
8	CCV2	10	1	9.620	9.976	08/22/2019	13:29
9	CCB2		1	0.195	0.199	08/22/2019	13:31

WORKLIST(Hardcopy Internal Chain)

LB10 4744

WorkList Name : TUR082219

WorkList ID : 130129

Department : Wet-Chemistry

Date : 08-22-2019 11:45:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/23/2019	Water	K4478-01	Turbidity	Cool 4 deg C	LIRO01	I11	UDS-6-01-82119	08/21/2019	SM2130 B

Date/Time 08-22-19 12:05

Received by: AL

Relinquished by: JP

Date/Time 08-22-19 13:35

Received by: JS

Relinquished by: AB