

7471B


LB104754

INSTRUMENT 15: CV1

Linear

 μ Abs.:

61477



A= 0.0000e+000

B= 1.6386e-004

C= -7.4791e-002

Rho= 0.9999065

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.055	-0.055	119	0.000	119				
0.2	0.200	0.232	0.032	1873	0.0 %	1873				
2.5	2.500	2.495	-0.005	15680	0.0 %	15680				
5.0	5.000	5.088	0.088	31507	0.0 %	31507				
7.5	7.500	7.442	-0.058	45870	0.0 %	45870				
10.0	10.000	9.999	-0.001	61477	0.0 %	61477				

LB104754 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Aug 2019 14:16:07

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	119	-	0.00007471B	PPB		22 Aug 2019 14:25:15	S	Std
0.2 - 1	50.2	1873	-	0.20007471B	PPB		22 Aug 2019 14:27:33	S	Std
2.5 - 1	52.5	15680	-	2.50007471B	PPB		22 Aug 2019 14:29:51	S	Std
5.0 - 1	55.0	31507	-	5.00007471B	PPB		22 Aug 2019 14:32:09	S	Std
7.5 - 1	57.5	45870	-	7.50007471B	PPB		22 Aug 2019 14:34:26	S	Std
10.0 - 1	50.0	61477	-	10.00007471B	PPB		22 Aug 2019 14:36:43	S	Std
ICV79 - 1	ICV79	24547	3.9475	-7471B	PPB		22 Aug 2019 14:40:58	U	SMPL
ICB79 - 1	ICB79	98	-0.0587	-7471B	PPB		22 Aug 2019 14:43:14	U	SMPL
CCV70 - 1	CCV70	31258	5.0472	-7471B	PPB		22 Aug 2019 14:45:30	U	SMPL
CCB70 - 1	CCB70	84	-0.0610	-7471B	PPB		22 Aug 2019 14:47:45	U	SMPL
CRA - 1	CRA	1853	0.2288	-7471B	PPB		22 Aug 2019 14:50:01	U	SMPL
HighStd - 1	HighStd	61767	10.0465	-7471B	PPB		22 Aug 2019 14:52:17	U	SMPL
ChkStd - 1	ChkStd	43179	7.0006	-7471B	PPB		22 Aug 2019 14:54:34	U	SMPL
PB122480BL - 1	PBS	82	-0.0614	-7471B	PPB		22 Aug 2019 14:56:51	U	SMPL
PB122480BS - 1	LCSS	21629	3.4694	-7471B	PPB		22 Aug 2019 14:59:08	U	SMPL
K4086-01 - 1	NB-08-082019	22597	3.6280	-7471B	PPB		22 Aug 2019 15:01:25	U	SMPL
K4087-01 - 1	HD-02-082119	1946	0.2441	-7471B	PPB		22 Aug 2019 15:03:42	U	SMPL
K4089-03 - 1	OK-01-082119	1268	0.1330	-7471B	PPB		22 Aug 2019 15:06:00	U	SMPL
K4095-09 - 1	SU-03-081919	1514	0.1733	-7471B	PPB		22 Aug 2019 15:08:18	U	SMPL
K4096-04 - 1	TR-05-081919	2528	0.3395	-7471B	PPB		22 Aug 2019 15:10:33	U	SMPL
CCV71 - 1	CCV71	30122	4.8611	-7471B	PPB		22 Aug 2019 15:12:50	U	SMPL
CCB71 - 1	CCB71	-28	-0.0794	-7471B	PPB		22 Aug 2019 15:15:05	U	SMPL
K4402-01 - 1	FRAC-TANK-MUD	3685	0.5290	-7471B	PPB		22 Aug 2019 15:17:22	U	SMPL
K4402-01DUP - 1	FRAC-TANK-MUDDUP	3658	0.5246	-7471B	PPB		22 Aug 2019 15:19:38	U	SMPL
K4402-01MS - 1	FRAC-TANK-MUDMS	27566	4.4422	-7471B	PPB		22 Aug 2019 15:21:54	U	SMPL
K4402-01MSD - 1	FRAC-TANK-MUDMSD	30349	4.8983	-7471B	PPB		22 Aug 2019 15:24:11	U	SMPL
K4408-01 - 1	SP-1	1380	0.1513	-7471B	PPB		22 Aug 2019 15:26:28	U	SMPL
K4418-04 - 1	CFVP-228L	260167	42.5568	-7471B	PPB		22 Aug 2019 15:28:45	U	SMPL
K4450-04 - 1	WC-(A-C)	17119	2.7304	-7471B	PPB		22 Aug 2019 15:31:02	U	SMPL
K4452-01 - 1	S-CORNER-3	3722	0.5351	-7471B	PPB		22 Aug 2019 15:33:22	U	SMPL
K4452-02 - 1	S-WALL	2151	0.2777	-7471B	PPB		22 Aug 2019 15:35:42	U	SMPL
K4465-01 - 1	FK-PS-01-056	3567	0.5097	-7471B	PPB		22 Aug 2019 15:37:58	U	SMPL
CCV72 - 1	CCV72	31022	5.0086	-7471B	PPB		22 Aug 2019 15:40:14	U	SMPL
CCB72 - 1	CCB72	-39	-0.0812	-7471B	PPB		22 Aug 2019 15:42:30	U	SMPL
K4402-01LX5 - 1		767	0.0509	-7471B	PPB		22 Aug 2019 15:44:46	U	SMPL
K4402-01A - 1		29006	4.6782	-7471B	PPB		22 Aug 2019 15:47:01	U	SMPL
K4418-04DLX10 - 1		28348	4.5704	-7471B	PPB		22 Aug 2019 15:49:17	U	SMPL
CCV73 - 1	CCV73	30589	4.9376	-7471B	PPB		22 Aug 2019 15:51:34	U	SMPL
CCB73 - 1	CCB73	109	-0.0569	-7471B	PPB		22 Aug 2019 15:53:51	U	SMPL