



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB104760

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	08/22/2019	16:30
2	CAL2	1	Water	NA	NA	20.1	6.98	08/22/2019	16:31
3	CAL3	1	Water	NA	NA	20.1	10.04	08/22/2019	16:33
4	ICV	1	Water	NA	NA	20.1	7.00	08/22/2019	16:35
5	CCV1	1	Water	NA	NA	20.2	2.02	08/22/2019	16:36
6	K4090-16	1	Solid	20.02	20	22.3	9.30	08/22/2019	16:37
7	K4487-01	1	Solid	20.03	20	22.7	8.79	08/22/2019	16:38
8	K4487-01DUP	1	Solid	20.04	20	22.6	8.80	08/22/2019	16:39
9	CCV2	1	Water	NA	NA	20.1	12.02	08/22/2019	16:40

20104960

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY K4090

WorkList ID : 130141

Department : Wet-Chemistry

Date : 08-22-2019 14:43:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/22/2019	Solid	K4090-16	Corrosivity	Cool 4 deg C	PSEG05		OR-03-0822219	08/22/2019	9045D
08/22/2019	Solid	K4487-01	Corrosivity	Cool 4 deg C	TETR16	111	WC-PILE-0822219	08/22/2019	9045D

Date/Time 08/22/2019 16:10
 Received by: JN
 Relinquished by:

Date/Time 08/22/2019 17:30
 Received by: CA
 Relinquished by: JN