



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104761

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	08/22/2019	16:30
2	CAL2	1	Water	NA	NA	20.1	6.98	08/22/2019	16:31
3	CAL3	1	Water	NA	NA	20.1	10.04	08/22/2019	16:33
4	ICV	1	Water	NA	NA	20.1	7.00	08/22/2019	16:35
5	CCV1	1	Water	NA	NA	20.2	2.02	08/22/2019	16:36
6	K4487-01	1	Solid	20.03	20	22.7	8.79	08/22/2019	16:38
7	K4487-01DUP	1	Solid	20.04	20	22.6	8.80	08/22/2019	16:39
8	CCV2	1	Water	NA	NA	20.1	12.02	08/22/2019	16:40

RD 104861

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4487 WorkList ID : 130153 Department : Wet-Chemistry Date : 08-22-2019 16:29:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/22/2019	Solid	K4487-01	pH	Cool 4 deg C	TETR16	I11	WC-PILE-082219	08/22/2019	9045D

Date/Time 08/22/2019 16:10
Received by: SA
Relinquished by: CD

Date/Time 08/22/2019 17:30
Received by: LM
Relinquished by: Se