

LB104767


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

61985



A= 0.0000e+000

B= 1.6298e-004

C= -6.9719e-003

Rho= 0.9997713

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.032	0.032	239	0.000	239				
0.2	0.200	0.228	0.028	1441	0.0 %	1441				
2.5	2.500	2.542	0.042	15637	0.0 %	15637				
5.0	5.000	4.850	-0.150	29800	0.0 %	29800				
7.5	7.500	7.453	-0.047	45773	0.0 %	45773				
10.0	10.000	10.095	0.095	61985	0.0 %	61985				

LB104767 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 22 Aug 2019 14:16:07				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50-0	239	-	0.00007471B	PPB		23 Aug 2019 12:08:08	S Std
0.2 - 1	50-2	1441	-	0.20007471B	PPB		23 Aug 2019 12:10:26	S Std
2.5 - 1	52-5	15637	-	2.50007471B	PPB		23 Aug 2019 12:12:44	S Std
5.0 - 1	55-0	29800	-	5.00007471B	PPB		23 Aug 2019 12:15:02	S Std
7.5 - 1	57-5	45773	-	7.50007471B	PPB		23 Aug 2019 12:17:20	S Std
10.0 - 1	510-0	61985	-	10.00007471B	PPB		23 Aug 2019 12:19:38	S Std
ICV81 - 1	ICV81	24669	4.0136	-7471B	PPB		23 Aug 2019 12:26:22	U SMPL
ICB81 - 1	ICB81	250	0.0338	-7471B	PPB		23 Aug 2019 12:28:39	U SMPL
CCV77 - 1	CCV77	29917	4.8690	-7471B	PPB		23 Aug 2019 12:30:55	U SMPL
CCB77 - 1	CCB77	229	0.0304	-7471B	PPB		23 Aug 2019 12:33:11	U SMPL
CRA - 1	CRA	1388	0.2192	-7471B	PPB		23 Aug 2019 12:35:26	U SMPL
HighStd - 1	HighStd	62248	10.1383	-7471B	PPB		23 Aug 2019 12:37:43	U SMPL
ChkStd - 1	ChkStd	41963	6.8322	-7471B	PPB		23 Aug 2019 12:39:59	U SMPL
PB122502BL - 1	PBS	184	0.0230	-7471B	PPB		23 Aug 2019 12:42:17	U SMPL
PB122502BS - 1	LCSS	21920	3.5656	-7471B	PPB		23 Aug 2019 12:44:33	U SMPL
K4090-13 - 1	OR-03-082219	3502	0.5638	-7471B	PPB		23 Aug 2019 12:46:52	U SMPL
K4090-13DUP - 1	OR-03-082219DUP	3480	0.5602	-7471B	PPB		23 Aug 2019 12:49:10	U SMPL
K4090-13MS - 1	OR-03-082219MS	26140	4.2534	-7471B	PPB		23 Aug 2019 12:51:27	U SMPL
K4090-13MSD - 1	OR-03-082219MSD	28962	4.7133	-7471B	PPB		23 Aug 2019 12:53:45	U SMPL
K4474-09 - 1	T1-2-F1-F4	1555	0.2465	-7471B	PPB		23 Aug 2019 12:56:01	U SMPL
CCV78 - 1	CCV78	30723	5.0003	-7471B	PPB		23 Aug 2019 12:58:17	U SMPL
CCB78 - 1	CCB78	244	0.0328	-7471B	PPB		23 Aug 2019 13:00:33	U SMPL
K4474-18 - 1	T1-2-K1-K4	32234	5.2466	-7471B	PPB		23 Aug 2019 13:02:49	U SMPL
K4475-01 - 1	T1-2A	29384	4.7821	-7471B	PPB		23 Aug 2019 13:05:05	U SMPL
K4475-04 - 1	T1-2E	33777	5.4981	-7471B	PPB		23 Aug 2019 13:07:22	U SMPL
K4475-07 - 1	T1-2D	11844	1.9234	-7471B	PPB		23 Aug 2019 13:09:39	U SMPL
K4475-10 - 1	T1-2C	29136	4.7417	-7471B	PPB		23 Aug 2019 13:11:57	U SMPL
K4475-13 - 1	T1-2B	34863	5.6751	-7471B	PPB		23 Aug 2019 13:14:13	U SMPL
K4477-01 - 1	TP-1(1-3)	1144	0.1795	-7471B	PPB		23 Aug 2019 13:16:31	U SMPL
K4477-02 - 1	TP-1(3-6)	702	0.1074	-7471B	PPB		23 Aug 2019 13:18:48	U SMPL
K4477-11 - 1	TP-2(1-3)	943	0.1467	-7471B	PPB		23 Aug 2019 13:21:05	U SMPL
K4477-12 - 1	TP-2(3-4)	1249	0.1966	-7471B	PPB		23 Aug 2019 13:23:22	U SMPL
CCV79 - 1	CCV79	31788	5.1739	-7471B	PPB		23 Aug 2019 13:25:38	U SMPL
CCB79 - 1	CCB79	252	0.0341	-7471B	PPB		23 Aug 2019 13:27:54	U SMPL
K4490-03 - 1	STUCCO	19522	3.1748	-7471B	PPB		23 Aug 2019 13:30:10	U SMPL
K4490-05 - 1	STUCCO-AND-BRICK	913	0.1418	-7471B	PPB		23 Aug 2019 13:32:27	U SMPL
K4490-08 - 1	BRICK-AND-PAINT	2242	0.3584	-7471B	PPB		23 Aug 2019 13:34:44	U SMPL
K4090-13LX5 - 1		859	0.1330	-7471B	PPB		23 Aug 2019 13:37:00	U SMPL
K4090-13A - 1		27797	4.5234	-7471B	PPB		23 Aug 2019 13:39:17	U SMPL
CCV80 - 1	CCV80	30719	4.9997	-7471B	PPB		23 Aug 2019 13:41:34	U SMPL
CCB80 - 1	CCB80	261	0.0356	-7471B	PPB		23 Aug 2019 13:43:51	U SMPL