



PERCENT SOLID

Supervisor: apatel  
Analyst: JIGNESH  
Date: 8/26/2019

OVENTEMP IN Celsius(°C): 107  
Time IN: 16:07  
In Date: 08/23/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 07:45  
Out Date: 08/24/2019  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB104770

| Lab ID   | Client SampleID | Dish # | Dish Wt(g) (A) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments |
|----------|-----------------|--------|----------------|-------------------------|---------------------------|---------|----------|
| K4503-01 | ETQ95           | 1      | 1.15           | 9.67                    | 8.27                      | 83.6    |          |
| K4503-02 | ETQ96           | 2      | 1.14           | 9.77                    | 7.97                      | 79.1    |          |
| K4503-03 | ETQ96MS         | 3      | 1.14           | 9.77                    | 7.97                      | 79.1    |          |
| K4503-04 | ETQ96MSD        | 4      | 1.14           | 9.77                    | 7.97                      | 79.1    |          |
| K4503-05 | ETQ97           | 5      | 1.13           | 9.66                    | 8.27                      | 83.7    |          |
| K4503-06 | ETQ98           | 6      | 1.15           | 9.93                    | 8.08                      | 78.9    |          |
| K4503-07 | ETQ99           | 7      | 1.13           | 9.81                    | 8.06                      | 79.8    |          |
| K4503-08 | ETQA0           | 8      | 1.14           | 9.75                    | 8.04                      | 80.1    |          |
| K4503-09 | ETQA1           | 9      | 1.13           | 9.99                    | 8.81                      | 86.7    |          |
| K4503-10 | ETQA2           | 10     | 1.14           | 9.81                    | 8.53                      | 85.2    |          |
| K4503-11 | ETQA3           | 11     | 1.14           | 9.55                    | 8.21                      | 84.1    |          |
| K4503-12 | ETQA4           | 12     | 1.14           | 9.88                    | 8.38                      | 82.8    |          |
| K4503-13 | ETQA5           | 13     | 1.14           | 9.81                    | 8.36                      | 83.3    |          |
| K4503-14 | ETQA6           | 14     | 1.15           | 9.84                    | 8.18                      | 80.9    |          |
| K4503-15 | ETQA7           | 15     | 1.15           | 9.78                    | 8.17                      | 81.3    |          |
| K4503-16 | ETQA8           | 16     | 1.13           | 9.78                    | 8.54                      | 85.7    |          |
| K4503-17 | ETQA9           | 17     | 1.14           | 9.98                    | 8.86                      | 87.3    |          |
| K4503-18 | ETQB0           | 18     | 1.15           | 9.63                    | 8.44                      | 86.0    |          |
| K4503-19 | ETQB5           | 19     | 1.15           | 9.66                    | 8.02                      | 80.7    |          |
| K4503-20 | ETQB6           | 20     | 1.14           | 9.83                    | 8.1                       | 80.1    |          |
| K4503-21 | ETQB7           | 21     | 1.14           | 9.61                    | 7.88                      | 79.6    |          |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

# WORKLIST(Hardcopy Internal Chain) LB104770

WorkList Name : %1-K4503      WorkList ID : 130182      Department : Wet-Chemistry      Date : 08-23-2019 13:14:01

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|------------------|-----------------|--------------|--------------|
| 08/28/2019 | Solid  | K4503-01 | Percent Solids | Cool 4 deg C | USEP04   | A61              | ETQ95           | 08/22/2019   | Chemtech -SO |
| 08/28/2019 | Solid  | K4503-02 | Percent Solids | Cool 4 deg C | USEP04   | A61              | ETQ96           | 08/22/2019   | Chemtech -SO |
| 08/28/2019 | Solid  | K4503-03 | Percent Solids | Cool 4 deg C | USEP04   | A61              | ETQ96MS         | 08/22/2019   | Chemtech -SO |
| 08/28/2019 | Solid  | K4503-04 | Percent Solids | Cool 4 deg C | USEP04   | A61              | ETQ96MSD        | 08/22/2019   | Chemtech -SO |
| 08/28/2019 | Solid  | K4503-05 | Percent Solids | Cool 4 deg C | USEP04   | A61              | ETQ97           | 08/22/2019   | Chemtech -SO |
| 08/28/2019 | Solid  | K4503-06 | Percent Solids | Cool 4 deg C | USEP04   | A61              | ETQ98           | 08/22/2019   | Chemtech -SO |
| 08/28/2019 | Solid  | K4503-07 | Percent Solids | Cool 4 deg C | USEP04   | A61              | ETQ99           | 08/22/2019   | Chemtech -SO |
| 08/28/2019 | Solid  | K4503-08 | Percent Solids | Cool 4 deg C | USEP04   | A61              | ETQA0           | 08/22/2019   | Chemtech -SO |
| 08/28/2019 | Solid  | K4503-09 | Percent Solids | Cool 4 deg C | USEP04   | A61              | ETQA1           | 08/22/2019   | Chemtech -SO |
| 08/28/2019 | Solid  | K4503-10 | Percent Solids | Cool 4 deg C | USEP04   | A61              | ETQA2           | 08/22/2019   | Chemtech -SO |
| 08/28/2019 | Solid  | K4503-11 | Percent Solids | Cool 4 deg C | USEP04   | A61              | ETQA3           | 08/22/2019   | Chemtech -SO |
| 08/28/2019 | Solid  | K4503-12 | Percent Solids | Cool 4 deg C | USEP04   | A61              | ETQA4           | 08/22/2019   | Chemtech -SO |
| 08/28/2019 | Solid  | K4503-13 | Percent Solids | Cool 4 deg C | USEP04   | A61              | ETQA5           | 08/22/2019   | Chemtech -SO |
| 08/28/2019 | Solid  | K4503-14 | Percent Solids | Cool 4 deg C | USEP04   | A61              | ETQA6           | 08/22/2019   | Chemtech -SO |
| 08/28/2019 | Solid  | K4503-15 | Percent Solids | Cool 4 deg C | USEP04   | A61              | ETQA7           | 08/22/2019   | Chemtech -SO |
| 08/28/2019 | Solid  | K4503-16 | Percent Solids | Cool 4 deg C | USEP04   | A61              | ETQA8           | 08/22/2019   | Chemtech -SO |
| 08/28/2019 | Solid  | K4503-17 | Percent Solids | Cool 4 deg C | USEP04   | A61              | ETQA9           | 08/22/2019   | Chemtech -SO |
| 08/28/2019 | Solid  | K4503-18 | Percent Solids | Cool 4 deg C | USEP04   | A61              | ETQB0           | 08/22/2019   | Chemtech -SO |
| 08/28/2019 | Solid  | K4503-19 | Percent Solids | Cool 4 deg C | USEP04   | A61              | ETQB5           | 08/22/2019   | Chemtech -SO |
| 08/28/2019 | Solid  | K4503-20 | Percent Solids | Cool 4 deg C | USEP04   | A61              | ETQB6           | 08/22/2019   | Chemtech -SO |
| 08/28/2019 | Solid  | K4503-21 | Percent Solids | Cool 4 deg C | USEP04   | A61              | ETQB7           | 08/22/2019   | Chemtech -SO |

Date/Time 8/23/19 14:35      Date/Time 8/23/19 15:00  
 Received by: DR.      Received by: DR.  
 Relinquished by: DR.      Relinquished by: DR.