


LB104713

7470A

INSTRUMENT ID : CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.016	0.016	82	0.000	82				
0.2	0.200	0.183	-0.017	869	0.0 %	869				
2.5	2.500	2.612	0.112	12340	0.0 %	12340				
5.0	5.000	4.767	-0.233	22517	0.0 %	22517				
7.5	7.500	7.668	0.168	36214	0.0 %	36214				
10.0	10.000	9.953	-0.047	47005	0.0 %	47005				

LB104773 INSTRUMENT ID: CV1

Method: 7470A		Operator: Admin		Date of Analysis: 23 Aug 2019 14:39:09				
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50-0	82	-	0.0000	7470A	PPB	23 Aug 2019 14:41:34	S Std
0.2 - 1	50-2	869	-	0.2000	7470A	PPB	23 Aug 2019 14:43:43	S Std
2.5 - 1	52-5	12340	-	2.5000	7470A	PPB	23 Aug 2019 14:45:52	S Std
5.0 - 1	55-0	22517	-	5.0000	7470A	PPB	23 Aug 2019 14:48:01	S Std
7.5 - 1	57-5	36214	-	7.5000	7470A	PPB	23 Aug 2019 14:50:09	S Std
10.0 - 1	510-0	47005	-	10.0000	7470A	PPB	23 Aug 2019 14:52:17	S Std
ICV82 - 1	ICV82	17454	3.6952	-	7470A	PPB	23 Aug 2019 15:10:55	U SMPL
ICB82 - 1	ICB82	-32	-0.0077	-	7470A	PPB	23 Aug 2019 15:15:36	U SMPL
CCV81 - 1	CCV81	22000	4.6579	-	7470A	PPB	23 Aug 2019 15:20:21	U SMPL
CCB81 - 1	CCB81	37	0.0069	-	7470A	PPB	23 Aug 2019 15:22:27	U SMPL
CRA - 1	CRA	838	0.1766	-	7470A	PPB	23 Aug 2019 15:24:33	U SMPL
HighStd - 1	HighStd	46860	9.9223	-	7470A	PPB	23 Aug 2019 15:26:40	U SMPL
ChkStd - 1	ChkStd	31810	6.7352	-	7470A	PPB	23 Aug 2019 15:28:46	U SMPL
PB122515BL - 1	PBW	189	0.0391	-	7470A	PPB	23 Aug 2019 15:30:53	U SMPL
PB122515BS - 1	LCSW	16461	3.4849	-	7470A	PPB	23 Aug 2019 15:33:01	U SMPL
K4087-03 - 1	HD-02-082119	54	0.0105	-	7470A	PPB	23 Aug 2019 15:35:08	U SMPL
K4087-03DUP - 1	HD-02-082119DUP	76	0.0152	-	7470A	PPB	23 Aug 2019 15:37:16	U SMPL
K4087-03MS - 1	HD-02-082119MS	16830	3.5631	-	7470A	PPB	23 Aug 2019 15:39:24	U SMPL
K4087-03MSD - 1	HD-02-082119MSD	15750	3.3344	-	7470A	PPB	23 Aug 2019 15:41:33	U SMPL
K4089-05 - 1	OK-01-082119	164	0.0338	-	7470A	PPB	23 Aug 2019 15:43:39	U SMPL
CCV82 - 1	CCV82	22156	4.6909	-	7470A	PPB	23 Aug 2019 15:45:46	U SMPL
CCB82 - 1	CCB82	35	0.0065	-	7470A	PPB	23 Aug 2019 15:47:53	U SMPL
K4450-07 - 1	WC-(A-C)	89	0.0180	-	7470A	PPB	23 Aug 2019 15:49:59	U SMPL
K4461-21 - 1	SB-01-COMP	53	0.0103	-	7470A	PPB	23 Aug 2019 15:52:06	U SMPL
K4461-22 - 1	SB-02-COMP	58	0.0114	-	7470A	PPB	23 Aug 2019 15:54:12	U SMPL
K4461-23 - 1	SB-03-COMP	167	0.0345	-	7470A	PPB	23 Aug 2019 15:56:19	U SMPL
K4461-24 - 1	SB-04-COMP	79	0.0158	-	7470A	PPB	23 Aug 2019 15:58:26	U SMPL
K4461-25 - 1	SB-05-COMP	21	0.0036	-	7470A	PPB	23 Aug 2019 16:00:34	U SMPL
K4461-26 - 1	SB-06-COMP	41	0.0078	-	7470A	PPB	23 Aug 2019 16:02:42	U SMPL
K4461-27 - 1	SB-07-COMP	68	0.0135	-	7470A	PPB	23 Aug 2019 16:04:50	U SMPL
K4461-28 - 1	SB-08-COMP	82	0.0165	-	7470A	PPB	23 Aug 2019 16:06:59	U SMPL
K4461-29 - 1	SB-09-COMP	88	0.0177	-	7470A	PPB	23 Aug 2019 16:09:05	U SMPL
CCV83 - 1	CCV83	22233	4.7072	-	7470A	PPB	23 Aug 2019 16:11:12	U SMPL
CCB83 - 1	CCB83	34	0.0063	-	7470A	PPB	23 Aug 2019 16:13:19	U SMPL
K4461-30 - 1	SB-10-COMP	66	0.0131	-	7470A	PPB	23 Aug 2019 16:15:25	U SMPL
K4462-05 - 1	SB-46-COMP	63	0.0125	-	7470A	PPB	23 Aug 2019 16:17:32	U SMPL
K4462-06 - 1	SB-47-COMP	33	0.0061	-	7470A	PPB	23 Aug 2019 16:19:39	U SMPL
PB122516BL - 1	PBW	79	0.0158	-	7470A	PPB	23 Aug 2019 16:21:46	U SMPL
PB122516BS - 1	LCSW	16506	3.4944	-	7470A	PPB	23 Aug 2019 16:23:54	U SMPL
K4090-16 - 1	OR-03-082219	57	0.0112	-	7470A	PPB	23 Aug 2019 16:26:02	U SMPL
K4090-16DUP - 1	OR-03-082219DUP	86	0.0173	-	7470A	PPB	23 Aug 2019 16:28:10	U SMPL
K4090-16MS - 1	OR-03-082219MS	16572	3.5084	-	7470A	PPB	23 Aug 2019 16:30:18	U SMPL
K4090-16MSD - 1	OR-03-082219MSD	16700	3.5355	-	7470A	PPB	23 Aug 2019 16:32:26	U SMPL
K4474-29 - 1	T1-2-F1-F4	45	0.0086	-	7470A	PPB	23 Aug 2019 16:34:34	U SMPL
CCV84 - 1	CCV84	22551	4.7745	-	7470A	PPB	23 Aug 2019 16:36:41	U SMPL
CCB84 - 1	CCB84	26	0.0046	-	7470A	PPB	23 Aug 2019 16:38:49	U SMPL
K4474-38 - 1	T1-2-K1-K4	53	0.0103	-	7470A	PPB	23 Aug 2019 16:40:56	U SMPL
K4475-14 - 1	T1-2A	49	0.0095	-	7470A	PPB	23 Aug 2019 16:43:03	U SMPL
K4475-17 - 1	T1-2E	67	0.0133	-	7470A	PPB	23 Aug 2019 16:45:11	U SMPL
K4475-20 - 1	T1-2D	94	0.0190	-	7470A	PPB	23 Aug 2019 16:47:18	U SMPL
K4475-23 - 1	T1-2C	122	0.0249	-	7470A	PPB	23 Aug 2019 16:49:25	U SMPL
K4475-26 - 1	T1-2B	107	0.0218	-	7470A	PPB	23 Aug 2019 16:51:33	U SMPL
K4487-01 - 1	WC-PILE-082219	89	0.0180	-	7470A	PPB	23 Aug 2019 16:53:41	U SMPL
K4490-04 - 1	STUCCO	80	0.0161	-	7470A	PPB	23 Aug 2019 16:55:49	U SMPL
K4490-06 - 1	STUCCO-AND-BRICK	73	0.0146	-	7470A	PPB	23 Aug 2019 16:57:57	U SMPL
K4490-09 - 1	BRICK-AND-PAINT	426	0.0893	-	7470A	PPB	23 Aug 2019 17:00:04	U SMPL
CCV85 - 1	CCV85	22165	4.6928	-	7470A	PPB	23 Aug 2019 17:02:11	U SMPL
CCB85 - 1	CCB85	37	0.0069	-	7470A	PPB	23 Aug 2019 17:04:18	U SMPL
K4493-23 - 1	T9-12-13-H-(0-1.9)	107	0.0218	-	7470A	PPB	23 Aug 2019 17:06:25	U SMPL
K4493-26 - 1	T9-12-13-I-(0-1.9)	100	0.0203	-	7470A	PPB	23 Aug 2019 17:08:32	U SMPL
K4493-29 - 1	T9-12-13-J-(0-1.9)	208	0.0432	-	7470A	PPB	23 Aug 2019 17:10:41	U SMPL
K4493-32 - 1	T9-12-13-K1-K4-(0-1.9)	114	0.0233	-	7470A	PPB	23 Aug 2019 17:12:49	U SMPL
PB122515TB - 1	PB122443TB	81	0.0163	-	7470A	PPB	23 Aug 2019 17:14:57	U SMPL
PB122516TB - 1	PB122481TB	69	0.0137	-	7470A	PPB	23 Aug 2019 17:17:04	U SMPL
K4087-03LX5 - 1		31	0.0057	-	7470A	PPB	23 Aug 2019 17:19:12	U SMPL
K4087-03A - 1		16218	3.4335	-	7470A	PPB	23 Aug 2019 17:21:21	U SMPL
K4090-16LX5 - 1		10	0.0012	-	7470A	PPB	23 Aug 2019 17:23:29	U SMPL
K4090-16A - 1		16682	3.5317	-	7470A	PPB	23 Aug 2019 17:25:37	U SMPL
CCV86 - 1	CCV86	22205	4.7013	-	7470A	PPB	23 Aug 2019 17:27:44	U SMPL
CCB86 - 1	CCB86	58	0.0114	-	7470A	PPB	23 Aug 2019 17:29:51	U SMPL