



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 8/27/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 11:12
In Date: 08/26/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:33
Out Date: 08/27/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB104785

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K4518-01	ETQB9	1	1.14	9.56	8.26	84.6	
K4518-02	ETQC0	2	1.15	9.74	8.68	87.7	
K4518-03	ETQC1	3	1.13	9.77	8.76	88.3	
K4518-04	ETQC2	4	1.15	9.76	8.59	86.4	
K4518-05	ETQC3	5	1.14	9.68	8.49	86.1	
K4518-06	ETQC4	6	1.15	9.88	8.43	83.4	
K4518-07	ETQC5	7	1.13	9.86	8.35	82.7	
K4518-08	ETQC6	8	1.13	9.85	8.36	82.9	
K4518-09	ETQC7	9	1.14	9.78	8.75	88.1	
K4518-10	ETQC8	10	1.14	9.9	8.76	87.0	
K4518-11	ETQC9	11	1.12	9.9	8.44	83.4	
K4518-12	ETQD0	12	1.14	9.73	7.67	76.0	
K4518-13	ETQD1	13	1.14	9.83	8.62	86.1	
K4518-14	ETQD2	14	1.13	9.64	8.48	86.4	
K4518-15	ETQD3	15	1.14	9.73	8.54	86.1	
K4518-16	ETQD4	16	1.12	9.88	8.1	79.7	
K4518-17	ETQD5	17	1.12	9.55	8.38	86.1	
K4518-18	ETQD6	18	1.14	9.66	8.29	83.9	
K4518-19	ETQD7	19	1.13	9.6	8.36	85.4	
K4518-20	ETQD8	20	1.12	9.96	8.69	85.6	
K4518-21	ETQD8MS	21	1.12	9.96	8.69	85.6	
K4518-22	ETQD8MSD	22	1.12	9.96	8.69	85.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$