



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104802

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	08/26/2019	16:25
2	CAL2	1	Water	NA	NA	20.1	7.01	08/26/2019	16:26
3	CAL3	1	Water	NA	NA	20.1	10.03	08/26/2019	16:27
4	ICV	1	Water	NA	NA	20.1	6.99	08/26/2019	16:30
5	CCV1	1	Water	NA	NA	20.1	2.02	08/26/2019	16:32
6	K4523-02	1	Water	NA	NA	20.5	9.47	08/26/2019	16:34
7	K4543-17	1	Water	NA	NA	19.8	12.05	08/26/2019	16:35
8	K4543-17DUP	1	Water	NA	NA	19.9	12.06	08/26/2019	16:36
9	CCV2	1	Water	NA	NA	20.1	12.02	08/26/2019	16:37

WORKLIST(Hardcopy Internal Chain)

104802

WorkList Name : PH K4523 WorkList ID : 130255 Department : Wet-Chemistry Date : 08-26-2019 16:12:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/23/2019	Water	K4523-02	pH	Cool 4 deg C	TETR06	I13	TT-005-IDWGW-20190823	08/23/2019	9040C
08/26/2019	Water	K4543-17	pH	Cool 4 deg C	PSEG01	J12	TP-2-GW	08/26/2019	9040C

Date/Time 08.26.19 16:15
 Received by: 
 Relinquished by:

Date/Time 08.26.19 17:30
 Received by: 
 Relinquished by: 