



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB104804

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].


Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	08/26/2019	16:25
2	CAL2	1	Water	NA	NA	20.1	7.01	08/26/2019	16:26
3	CAL3	1	Water	NA	NA	20.1	10.03	08/26/2019	16:27
4	ICV	1	Water	NA	NA	20.1	6.99	08/26/2019	16:30
5	CCV1	1	Water	NA	NA	20.2	2.02	08/26/2019	17:00
6	K4086-08	1	Solid	20.02	20	23.2	7.90	08/26/2019	17:01
7	K4088-14	1	Solid	20.04	20	23.1	8.88	08/26/2019	17:02
8	K4092-11	1	Solid	20.03	20	23.3	7.90	08/26/2019	17:05
9	K4093-10	1	Solid	20.02	20	22.2	8.80	08/26/2019	17:06
10	K4506-02	1	Solid	20.03	20	22.1	7.05	08/26/2019	17:09
11	K4543-04	1	Solid	20.04	20	20.5	7.73	08/26/2019	17:10
12	K4543-08	1	Solid	20.02	20	22.2	7.54	08/26/2019	17:12
13	K4543-12	1	Solid	20.03	20	22.5	7.20	08/26/2019	17:15
14	K4543-16	1	Solid	20.04	20	22.7	7.53	08/26/2019	17:16
15	K4543-16DUP	1	Solid	20.03	20	22.6	7.54	08/26/2019	17:20
16	CCV2	1	Water	NA	NA	20.1	12.02	08/26/2019	17:25

207104804

WORKLIST(Hardcopy Internal Chain)

WorkList Name: CORROSIONITY K4086 WorkList ID: 130254 Department: Wet-Chemistry Date: 08-26-2019 16:11:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/26/2019	Solid	K4086-08	Corrosivity	Cool 4 deg C	PSEG05		NB-07-082619	08/26/2019	9045D
08/26/2019	Solid	K4088-14	Corrosivity	Cool 4 deg C	PSEG05		CL-02-082619	08/26/2019	9045D
08/26/2019	Solid	K4092-11	Corrosivity	Cool 4 deg C	PSEG05		EO-02-082619	08/26/2019	9045D
08/26/2019	Solid	K4093-10	Corrosivity	Cool 4 deg C	PSEG05		PL-02-082619	08/26/2019	9045D
08/26/2019	Solid	K4506-02	Corrosivity	Cool 4 deg C	PSEG01	J12	RO-COMP	08/26/2019	9045D
08/26/2019	Solid	K4543-04	Corrosivity	Cool 4 deg C	PSEG01	J12	TP-2-S	08/26/2019	9045D
08/26/2019	Solid	K4543-08	Corrosivity	Cool 4 deg C	PSEG01	J12	TP-2-D	08/26/2019	9045D
08/26/2019	Solid	K4543-12	Corrosivity	Cool 4 deg C	PSEG01	J12	TP-3-S	08/26/2019	9045D
08/26/2019	Solid	K4543-16	Corrosivity	Cool 4 deg C	PSEG01	J12	TP-9-S	08/26/2019	9045D
08/26/2019	Water	K4543-18	Corrosivity	Cool 4 deg C	PSEG01	J12	TP-2-GW	08/26/2019	9040C

Date/Time: 08.26.19 16:10
 Received by: 
 Relinquished by: 

Date/Time: 08.26.19 18:30
 Received by: 
 Relinquished by: 