


LB104816

7470A

INSTRUMENT ID : CV1

Linear ▼



A= 0.0000e+000

B= 2.3923e-004

C= -4.7369e-002

Rho= 0.9998008

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.036	-0.036	49	0.000	49				
0.2	0.200	0.162	-0.038	877	0.0 %	877				
2.5	2.500	2.593	0.093	11038	0.0 %	11038				
5.0	5.000	5.073	0.073	21403	0.0 %	21403				
7.5	7.500	7.378	-0.122	31037	0.0 %	31037				
10.0	10.000	10.030	0.030	42123	0.0 %	42123				

LB104816 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 27 Aug 2019 11:09:11					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	49	-	0.0000	7470A	PPB	27 Aug 2019 11:13:14	S	Std
0.2 - 1	50-2	877	-	0.2000	7470A	PPB	27 Aug 2019 11:15:23	S	Std
2.5 - 1	52-5	11038	-	2.5000	7470A	PPB	27 Aug 2019 11:17:32	S	Std
5.0 - 1	55-0	21403	-	5.0000	7470A	PPB	27 Aug 2019 11:19:41	S	Std
7.5 - 1	57-5	31037	-	7.5000	7470A	PPB	27 Aug 2019 11:21:49	S	Std
10.0 - 1	510-0	42123	-	10.0000	7470A	PPB	27 Aug 2019 11:30:12	S	Std
ICV86 - 1	ICV86	17024	4.0252	-	7470A	PPB	27 Aug 2019 11:38:42	U	SMPL
ICB86 - 1	ICB86	2	-0.0469	-	7470A	PPB	27 Aug 2019 11:40:49	U	SMPL
CCV93 - 1	CCV93	21014	4.9798	-	7470A	PPB	27 Aug 2019 11:42:55	U	SMPL
CCB93 - 1	CCB93	-23	-0.0529	-	7470A	PPB	27 Aug 2019 11:45:02	U	SMPL
CRA - 1	CRA	856	0.1574	-	7470A	PPB	27 Aug 2019 11:47:08	U	SMPL
HighStd - 1	HighStd	41770	9.9452	-	7470A	PPB	27 Aug 2019 11:49:15	U	SMPL
ChkStd - 1	ChkStd	29111	6.9168	-	7470A	PPB	27 Aug 2019 11:51:22	U	SMPL
PB122583BL - 1	PBW	-9	-0.0495	-	7470A	PPB	27 Aug 2019 11:53:30	U	SMPL
PB122583BS - 1	LCSW	17144	4.0540	-	7470A	PPB	27 Aug 2019 11:55:39	U	SMPL
K4509-11 - 1	SB-48-COMP	7	-0.0457	-	7470A	PPB	27 Aug 2019 11:57:49	U	SMPL
K4509-11DUP - 1	SB-48-COMPDUP	46	-0.0364	-	7470A	PPB	27 Aug 2019 11:59:57	U	SMPL
K4509-11MS - 1	SB-48-COMPMS	18462	4.3693	-	7470A	PPB	27 Aug 2019 12:02:05	U	SMPL
K4509-11MSD - 1	SB-48-COMPMSD	16757	3.9614	-	7470A	PPB	27 Aug 2019 12:04:14	U	SMPL
K4509-12 - 1	SB-49-COMP	-27	-0.0538	-	7470A	PPB	27 Aug 2019 12:06:20	U	SMPL
CCV94 - 1	CCV94	21096	4.9994	-	7470A	PPB	27 Aug 2019 12:08:27	U	SMPL
CCB94 - 1	CCB94	-8	-0.0493	-	7470A	PPB	27 Aug 2019 12:10:34	U	SMPL
K4509-13 - 1	SB-50-COMP	29	-0.0404	-	7470A	PPB	27 Aug 2019 12:12:41	U	SMPL
K4509-14 - 1	SB-51-COMP	54	-0.0345	-	7470A	PPB	27 Aug 2019 12:14:48	U	SMPL
K4509-15 - 1	SB-45-COMP	61	-0.0328	-	7470A	PPB	27 Aug 2019 12:16:55	U	SMPL
K4510-15 - 1	SB-44-COMP	52	-0.0349	-	7470A	PPB	27 Aug 2019 12:19:02	U	SMPL
K4510-16 - 1	SB-41-COMP	30	-0.0402	-	7470A	PPB	27 Aug 2019 12:21:09	U	SMPL
K4510-17 - 1	SB-42-COMP	-1	-0.0476	-	7470A	PPB	27 Aug 2019 12:23:17	U	SMPL
K4510-18 - 1	SB-43-COMP	25	-0.0414	-	7470A	PPB	27 Aug 2019 12:25:25	U	SMPL
K4510-19 - 1	SB-14-COMP	30	-0.0402	-	7470A	PPB	27 Aug 2019 12:27:33	U	SMPL
K4510-20 - 1	SB-13-COMP	58	-0.0335	-	7470A	PPB	27 Aug 2019 12:29:42	U	SMPL
K4510-21 - 1	SB-11-COMP	62	-0.0325	-	7470A	PPB	27 Aug 2019 12:31:48	U	SMPL
CCV95 - 1	CCV95	21491	5.0939	-	7470A	PPB	27 Aug 2019 12:33:55	U	SMPL
CCB95 - 1	CCB95	-6	-0.0488	-	7470A	PPB	27 Aug 2019 12:36:02	U	SMPL
PB122584BL - 1	PBW	13	-0.0443	-	7470A	PPB	27 Aug 2019 12:38:09	U	SMPL
PB122584BS - 1	LCSW	16901	3.9958	-	7470A	PPB	27 Aug 2019 12:40:15	U	SMPL
K4508-01 - 1	BLK-A4	581	0.0916	-	7470A	PPB	27 Aug 2019 12:42:22	U	SMPL
K4508-01DUP - 1	BLK-A4DUP	551	0.0844	-	7470A	PPB	27 Aug 2019 12:44:29	U	SMPL
K4515-16 - 1	EB-1-082319	-41	-0.0572	-	7470A	PPB	27 Aug 2019 12:50:52	U	SMPL
K4523-02 - 1	TT-005-IDWGW-20190823	486	0.0689	-	7470A	PPB	27 Aug 2019 12:53:02	U	SMPL
K4543-17 - 1	TP-2-GW	18499	4.3781	-	7470A	PPB	27 Aug 2019 12:55:10	U	SMPL
PB122583TB - 1	PB122521TB	-21	-0.0524	-	7470A	PPB	27 Aug 2019 12:57:17	U	SMPL
CCV96 - 1	CCV96	21082	4.9960	-	7470A	PPB	27 Aug 2019 12:59:24	U	SMPL
CCB96 - 1	CCB96	-34	-0.0555	-	7470A	PPB	27 Aug 2019 13:01:31	U	SMPL
K4509-11LX5 - 1		5	-0.0462	-	7470A	PPB	27 Aug 2019 13:03:38	U	SMPL
K4509-11A - 1		17422	4.1205	-	7470A	PPB	27 Aug 2019 13:05:45	U	SMPL
K4508-01LX5 - 1		101	-0.0232	-	7470A	PPB	27 Aug 2019 13:07:52	U	SMPL
K4508-01A - 1		14461	3.4121	-	7470A	PPB	27 Aug 2019 13:15:06	U	SMPL
K4508-01MS - 1	BLK-A4MS	16723	3.9532	-	7470A	PPB	27 Aug 2019 13:17:13	U	SMPL
K4508-01MSD - 1	BLK-A4MSD	16571	3.9169	-	7470A	PPB	27 Aug 2019 13:19:21	U	SMPL
CCV97 - 1	CCV97	20815	4.9322	-	7470A	PPB	27 Aug 2019 13:21:29	U	SMPL
CCB97 - 1	CCB97	-23	-0.0529	-	7470A	PPB	27 Aug 2019 13:23:37	U	SMPL