



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB104823

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	08/27/2019	13:32
2	K4543-18		1	9	100.00	>212.0	>212.0	08/27/2019	13:58
3	K4543-18DUP		1	9	100.00	>212.0	>212.0	08/27/2019	14:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB-1048/23

WorkList Name : FP82719 WorkList ID : 130283 Department : Wet-Chemistry Date : 08-27-2019 08:44:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/02/2019	Water	K4543-18	Flash Point	Cool 4 deg C	PSEG01	J12	TP-2-GW	08/26/2019	1010A

Date/Time 08/27/19 13:00
 Received by: KS
 Relinquished by: JP

Date/Time 08/27/19 14:30
 Received by: JP
 Relinquished by: KS