



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB104828

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	08/27/2019	17:10
2	CAL2	1	Water	NA	NA	20.2	7.00	08/27/2019	17:11
3	CAL3	1	Water	NA	NA	20.2	10.03	08/27/2019	17:12
4	ICV	1	Water	NA	NA	20.2	6.99	08/27/2019	17:12
5	CCV1	1	Water	NA	NA	20.2	2.02	08/27/2019	17:15
6	K4567-22	1	Water	NA	NA	17.4	6.43	08/27/2019	17:16
7	K4567-25	1	Water	NA	NA	19.6	6.12	08/27/2019	17:17
8	CCV2	1	Water	NA	NA	20.2	12.01	08/27/2019	17:20

WORKLIST(Hardcopy Internal Chain)

LB 104828

WorkList Name : CORROSIONITY K4567 WorkList ID : 130301 Department : Wet-Chemistry Date : 08-27-2019 16:59:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/27/2019	Water	K4567-22	Corrosivity	Cool 4 deg C	PSEG01	J13	TP-7-GW	08/27/2019	9040C
08/27/2019	Water	K4567-25	Corrosivity	Cool 4 deg C	PSEG01	J13	TP-8-GW	08/27/2019	9040C

Date/Time 08/27/19 17:00
 Received by: Jb
 Relinquished by: _____

Date/Time 08/27/19 18:20
 Received by: cm
 Relinquished by: Jb