


LB104886

ISM02.4

INSTRUMENT ID : CV1



Linear

A= 0.0000e+000

B= 2.3306e-004

C= 5.2687e-004

Rho= 0.9999133

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.012	0.012	50	0.000	50					
0.05	0.050										
0.20	0.200	0.204	0.004	873	0.0 %	873					
2.50	2.500	2.543	0.043	10910	0.0 %	10910					2
5.00	5.000	4.937	-0.063	21179	0.0 %	21179					2
7.50	7.500	7.438	-0.062	31913	0.0 %	31913					-1
10.0	10.000	10.066	0.066	43186	0.0 %	43186					-1

2
2
-1
-1
1

LB104836
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	50	-	0	ISM02.4	PPB	8/28/2019 9:15	Std
	0.2 S0.2	873	-	0.2	ISM02.4	PPB	8/28/2019 9:17	Std
	2.5 S2.5	10910	-	2.5	ISM02.4	PPB	8/28/2019 9:20	Std
	5 S5.0	21179	-	5	ISM02.4	PPB	8/28/2019 9:36	Std
	7.5 S7.5	31913	-	7.5	ISM02.4	PPB	8/28/2019 9:38	Std
	10 S10.0	43186	-	10	ISM02.4	PPB	8/28/2019 9:40	Std
ICV	ICV	16692	3.8908 -		ISM02.4	PPB	8/28/2019 9:43	SMPL
ICB	ICB	2	0.001 -		ISM02.4	PPB	8/28/2019 9:45	SMPL
CCV008	CCV008	20869	4.8644 -		ISM02.4	PPB	8/28/2019 9:48	SMPL
CCB008	CCB008	-3	-0.0002 -		ISM02.4	PPB	8/28/2019 9:50	SMPL
PB122578BL	PBW578	14	0.0038 -		ISM02.4	PPB	8/28/2019 9:52	SMPL
K4209-01	MCOAA1	18	0.0047 -		ISM02.4	PPB	8/28/2019 9:54	SMPL
K4209-02	MCOAA2	21	0.0054 -		ISM02.4	PPB	8/28/2019 9:57	SMPL
K4209-03	MCOAA3	22	0.0057 -		ISM02.4	PPB	8/28/2019 9:59	SMPL
K4209-04	MCOAA6	124	0.0294 -		ISM02.4	PPB	8/28/2019 10:01	SMPL
K4209-05	MCOAA7	63	0.0152 -		ISM02.4	PPB	8/28/2019 10:03	SMPL
K4209-06	MCOAA8	64	0.0154 -		ISM02.4	PPB	8/28/2019 10:06	SMPL
K4209-07	MCOAA0	46	0.0112 -		ISM02.4	PPB	8/28/2019 10:08	SMPL
K4209-08	MCOAE2	35	0.0087 -		ISM02.4	PPB	8/28/2019 10:10	SMPL
K4209-09	MCOAA9	99	0.0236 -		ISM02.4	PPB	8/28/2019 10:13	SMPL
K4209-10	MCOAA4	69	0.0166 -		ISM02.4	PPB	8/28/2019 10:15	SMPL
K4209-11	MCOAA4D	75	0.018 -		ISM02.4	PPB	8/28/2019 10:17	SMPL
K4209-12	MCOAA4S	4867	1.1349 -		ISM02.4	PPB	8/28/2019 10:19	SMPL
K4209-13	MCOAA5	78	0.0187 -		ISM02.4	PPB	8/28/2019 10:22	SMPL
K4209-14	MCOAE3	36	0.0089 -		ISM02.4	PPB	8/28/2019 10:24	SMPL
PB122579BL	PBW579	17	0.0045 -		ISM02.4	PPB	8/28/2019 10:26	SMPL
K4374-07	MX2H69	19583	4.5646 -		ISM02.4	PPB	8/28/2019 10:28	SMPL
K4374-08	MX2H69D	19015	4.4322 -		ISM02.4	PPB	8/28/2019 10:31	SMPL
K4374-09	MX2H69S	23438	5.4631 -		ISM02.4	PPB	8/28/2019 10:33	SMPL
CCV009	CCV009	20790	4.8459 -		ISM02.4	PPB	8/28/2019 10:35	SMPL
CCB009	CCB009	8	0.0024 -		ISM02.4	PPB	8/28/2019 10:38	SMPL