

LB10483A


7471B

INSTRUMENT ID: CVI

Linear

 μ Abs.:

59939



A= 0.0000e+000

B= 1.6760e-004

C= -6.8735e-002

Rho= 0.9998900

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.027	-0.027	249	0.000	249				
0.2	0.200	0.165	-0.035	1394	0.0 %	1394				
2.5	2.500	2.530	0.030	15503	0.0 %	15503				
5.0	5.000	5.109	0.109	30891	0.0 %	30891				
7.5	7.500	7.447	-0.053	44843	0.0 %	44843				
10.0	10.000	9.977	-0.023	59939	0.0 %	59939				

LB104839 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 28 Aug 2019 11:16:29				
Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type
0.0 - 1	50.0	249	-	0.0000	7471B	PPB	28 Aug 2019 11:19:47	S Std
0.2 - 1	50.2	1394	-	0.2000	7471B	PPB	28 Aug 2019 11:22:05	S Std
2.5 - 1	52.5	15503	-	2.5000	7471B	PPB	28 Aug 2019 11:24:23	S Std
5.0 - 1	55.0	30891	-	5.0000	7471B	PPB	28 Aug 2019 11:26:41	S Std
7.5 - 1	57.5	44843	-	7.5000	7471B	PPB	28 Aug 2019 11:28:58	S Std
10.0 - 1	510.0	59939	-	10.0000	7471B	PPB	28 Aug 2019 11:31:15	S Std
ICV89 - 1	ICV89	24004	3.9543	-	7471B	PPB	28 Aug 2019 11:36:11	U SMPL
ICB89 - 1	ICB89	189	-0.0371	-	7471B	PPB	28 Aug 2019 11:38:28	U SMPL
CCV10 - 1	CCV10	30628	5.0645	-	7471B	PPB	28 Aug 2019 11:40:44	U SMPL
CCB10 - 1	CCB10	189	-0.0371	-	7471B	PPB	28 Aug 2019 11:42:59	U SMPL
CRA - 1	CRA	1411	0.1677	-	7471B	PPB	28 Aug 2019 11:45:16	U SMPL
HighStd - 1	HighStd	59900	9.9705	-	7471B	PPB	28 Aug 2019 11:47:31	U SMPL
ChkStd - 1	ChkStd	41247	6.8442	-	7471B	PPB	28 Aug 2019 11:49:48	U SMPL
PB122632BL - 1	PBS	239	-0.0287	-	7471B	PPB	28 Aug 2019 11:52:05	U SMPL
PB122632BS - 1	LCSS	22148	3.6433	-	7471B	PPB	28 Aug 2019 11:54:23	U SMPL
K4562-01 - 1	SB-02-15-17	910	0.0838	-	7471B	PPB	28 Aug 2019 11:56:40	U SMPL
K4562-01DUP - 1	SB-02-15-17DUP	934	0.0878	-	7471B	PPB	28 Aug 2019 11:58:58	U SMPL
K4562-01MS - 1	SB-02-15-17MS	20656	3.3932	-	7471B	PPB	28 Aug 2019 12:01:15	U SMPL
K4562-01MSD - 1	SB-02-15-17MSD	22135	3.6411	-	7471B	PPB	28 Aug 2019 12:03:33	U SMPL
K4562-03 - 1	SB-02-20-22	399	-0.0019	-	7471B	PPB	28 Aug 2019 12:05:50	U SMPL
CCV11 - 1	CCV11	30858	5.1031	-	7471B	PPB	28 Aug 2019 12:08:05	U SMPL
CCB11 - 1	CCB11	172	-0.0399	-	7471B	PPB	28 Aug 2019 12:10:22	U SMPL
K4562-05 - 1	SB-02-25-27	732	0.0539	-	7471B	PPB	28 Aug 2019 12:12:38	U SMPL
K4562-07 - 1	SB-02-30-32	1187	0.1302	-	7471B	PPB	28 Aug 2019 12:14:54	U SMPL
K4567-05 - 1	TP-7-S	1557	0.1922	-	7471B	PPB	28 Aug 2019 12:17:09	U SMPL
K4567-11 - 1	TP-7-D	844	0.0727	-	7471B	PPB	28 Aug 2019 12:19:25	U SMPL
K4567-15 - 1	TP-8-S	2057	0.2760	-	7471B	PPB	28 Aug 2019 12:21:42	U SMPL
K4567-19 - 1	TP-8-D	1223	0.1362	-	7471B	PPB	28 Aug 2019 12:23:59	U SMPL
K4571-01 - 1	RT-3425	1927	0.2542	-	7471B	PPB	28 Aug 2019 12:26:16	U SMPL
K4571-04 - 1	RBR-200028	1631	0.2046	-	7471B	PPB	28 Aug 2019 12:28:33	U SMPL
K4571-07 - 1	RBR-200052	1194	0.1314	-	7471B	PPB	28 Aug 2019 12:30:51	U SMPL
K4572-01 - 1	RBR-200059-200035	1312	0.1512	-	7471B	PPB	28 Aug 2019 12:33:08	U SMPL
CCV12 - 1	CCV12	30614	5.0622	-	7471B	PPB	28 Aug 2019 12:35:24	U SMPL
CCB12 - 1	CCB12	197	-0.0357	-	7471B	PPB	28 Aug 2019 12:37:40	U SMPL
K4576-01 - 1	RT-1868-4221	1870	0.2447	-	7471B	PPB	28 Aug 2019 12:39:56	U SMPL
K4576-02 - 1	RBR-200034-RBR-200036	3360	0.4944	-	7471B	PPB	28 Aug 2019 12:42:12	U SMPL
PB122633BL - 1	PBS	284	-0.0211	-	7471B	PPB	28 Aug 2019 12:44:28	U SMPL
PB122633BS - 1	LCSS	22269	3.6635	-	7471B	PPB	28 Aug 2019 12:46:44	U SMPL
K4517-05 - 1	T2-S-A1-A4	5601	0.8700	-	7471B	PPB	28 Aug 2019 12:49:01	U SMPL
K4517-12 - 1	T4-S-D1-D4	1007	0.1000	-	7471B	PPB	28 Aug 2019 12:51:18	U SMPL
K4517-15 - 1	T4-S-C	3127	0.4553	-	7471B	PPB	28 Aug 2019 12:53:35	U SMPL
K4517-18 - 1	T4-S-E	3710	0.5531	-	7471B	PPB	28 Aug 2019 12:55:53	U SMPL
K4522-02 - 1	SIDEWALL-5	4414	0.6710	-	7471B	PPB	28 Aug 2019 12:58:12	U SMPL
K4544-01 - 1	T16-17-B1-B4-(0-2.25)	20890	3.4324	-	7471B	PPB	28 Aug 2019 13:00:28	U SMPL
CCV13 - 1	CCV13	30971	5.1220	-	7471B	PPB	28 Aug 2019 13:02:44	U SMPL
CCB13 - 1	CCB13	174	-0.0396	-	7471B	PPB	28 Aug 2019 13:05:01	U SMPL
K4544-10 - 1	T16-17-C1-C4-(0-2.25)	9764	1.5677	-	7471B	PPB	28 Aug 2019 13:07:18	U SMPL
K4544-19 - 1	T16-17-E1-E4-(0-2.25)	24408	4.0220	-	7471B	PPB	28 Aug 2019 13:09:35	U SMPL
K4546-09 - 1	T3-S-C1-C4	3497	0.5174	-	7471B	PPB	28 Aug 2019 13:11:50	U SMPL
K4546-12 - 1	T3-S-A	5650	0.8782	-	7471B	PPB	28 Aug 2019 13:14:08	U SMPL
K4547-01 - 1	T3-S-B1-B4	5691	0.8851	-	7471B	PPB	28 Aug 2019 13:16:25	U SMPL
K4547-04 - 1	T3-S-D	9475	1.5193	-	7471B	PPB	28 Aug 2019 13:18:41	U SMPL
K4547-07 - 1	T3-S-G	5230	0.8078	-	7471B	PPB	28 Aug 2019 13:20:59	U SMPL
K4547-10 - 1	T3-S-F	6369	0.9987	-	7471B	PPB	28 Aug 2019 13:23:16	U SMPL
K4547-15 - 1	T7-S-G1-G4	5105	0.7869	-	7471B	PPB	28 Aug 2019 13:25:34	U SMPL
K4548-09 - 1	T8-S-G1-G4(0)	2738	0.3902	-	7471B	PPB	28 Aug 2019 13:27:51	U SMPL
CCV14 - 1	CCV14	31062	5.1372	-	7471B	PPB	28 Aug 2019 13:30:07	U SMPL
CCB14 - 1	CCB14	190	-0.0369	-	7471B	PPB	28 Aug 2019 13:32:24	U SMPL
K4551-01 - 1	S401-K1-1.0-1.5-082619-FD	2712	0.3858	-	7471B	PPB	28 Aug 2019 13:34:41	U SMPL
K4551-08 - 1	S400-K1-1.0-1.5-082619	3949	0.5931	-	7471B	PPB	28 Aug 2019 13:36:58	U SMPL
K4551-17 - 1	S402-K1-1.0-1.5-082619	824	0.0694	-	7471B	PPB	28 Aug 2019 13:39:14	U SMPL
K4551-17DUP - 1	S402-K1-1.0-1.5-082619DUP	762	0.0590	-	7471B	PPB	28 Aug 2019 13:41:30	U SMPL
K4551-18 - 1	K4551-17MS	23521	3.8734	-	7471B	PPB	28 Aug 2019 13:43:47	U SMPL
K4551-20 - 1	K4551-17MSD	23168	3.8142	-	7471B	PPB	28 Aug 2019 13:46:03	U SMPL
K4560-01 - 1	S404-K1-1.0-1.5-082619	3239	0.4741	-	7471B	PPB	28 Aug 2019 13:48:21	U SMPL
K4560-10 - 1	S401-K1-1.0-1.5-082619	3521	0.5214	-	7471B	PPB	28 Aug 2019 13:50:38	U SMPL
K4562-01LX5 - 1		326	-0.0141	-	7471B	PPB	28 Aug 2019 13:52:56	U SMPL
K4562-01A - 1		21443	3.5251	-	7471B	PPB	28 Aug 2019 13:55:13	U SMPL
CCV15 - 1	CCV15	31077	5.1398	-	7471B	PPB	28 Aug 2019 13:57:29	U SMPL
CCB15 - 1	CCB15	289	-0.0203	-	7471B	PPB	28 Aug 2019 13:59:45	U SMPL
K4551-17LX5 - 1		240	-0.0285	-	7471B	PPB	28 Aug 2019 14:02:02	U SMPL
K4551-17A - 1		23463	3.8637	-	7471B	PPB	28 Aug 2019 14:04:19	U SMPL
CCV16 - 1	CCV16	30609	5.0613	-	7471B	PPB	28 Aug 2019 14:06:35	U SMPL
CCB16 - 1	CCB16	181	-0.0384	-	7471B	PPB	28 Aug 2019 14:08:52	U SMPL