

LB104846


7470A

INSTRUMENT ID : CV1

Linear

 μ Abs. :

40468



A= 0.0000e+000

B= 2.4706e-004

C= -5.9846e-002

Rho= 0.9999197

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.040	-0.040	80	0.000	80				
0.2	0.200	0.170	-0.030	931	0.0 %	931				
2.5	2.500	2.542	0.042	10530	0.0 %	10530				
5.0	5.000	5.067	0.067	20753	0.0 %	20753				
7.5	7.500	7.523	0.023	30691	0.0 %	30691				
10.0	10.000	9.938	-0.062	40468	0.0 %	40468				

LB104846 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 28 Aug 2019 14:28:07					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	80	-	0.0000	7470A	PPB	28 Aug 2019 14:29:31	S	Std
0.2 - 1	50.2	931	-	0.2000	7470A	PPB	28 Aug 2019 14:31:40	S	Std
2.5 - 1	52.5	10530	-	2.5000	7470A	PPB	28 Aug 2019 14:33:49	S	Std
5.0 - 1	55.0	20753	-	5.0000	7470A	PPB	28 Aug 2019 14:35:58	S	Std
7.5 - 1	57.5	30691	-	7.5000	7470A	PPB	28 Aug 2019 14:38:06	S	Std
10.0 - 1	510.0	40468	-	10.0000	7470A	PPB	28 Aug 2019 14:42:47	S	Std
ICV90 - 1	ICV90	16190	3.9401	-	7470A	PPB	28 Aug 2019 14:45:19	U	SMPL
ICB90 - 1	ICB90	8	-0.0579	-	7470A	PPB	28 Aug 2019 14:47:26	U	SMPL
CCV17 - 1	CCV17	20745	5.0654	-	7470A	PPB	28 Aug 2019 14:49:32	U	SMPL
CCB17 - 1	CCB17	8	-0.0579	-	7470A	PPB	28 Aug 2019 14:51:39	U	SMPL
CRA - 1	CRA	883	0.1583	-	7470A	PPB	28 Aug 2019 14:55:17	U	SMPL
HighStd - 1	HighStd	39606	9.7252	-	7470A	PPB	28 Aug 2019 14:57:23	U	SMPL
ChkStd - 1	ChkStd	26840	6.5712	-	7470A	PPB	28 Aug 2019 14:59:29	U	SMPL
PB122630BL - 1	PBW	37	-0.0507	-	7470A	PPB	28 Aug 2019 15:01:38	U	SMPL
PB122630BS - 1	LCSW	16693	4.0643	-	7470A	PPB	28 Aug 2019 15:03:46	U	SMPL
K4086-08 - 1	NB-07-082619	74	-0.0416	-	7470A	PPB	28 Aug 2019 15:05:53	U	SMPL
K4086-08DUP - 1	NB-07-082619DUP	100	-0.0351	-	7470A	PPB	28 Aug 2019 15:08:01	U	SMPL
K4086-08MS - 1	NB-07-082619MS	16645	4.0525	-	7470A	PPB	28 Aug 2019 15:10:09	U	SMPL
K4086-08MSD - 1	NB-07-082619MSD	16401	3.9922	-	7470A	PPB	28 Aug 2019 15:12:17	U	SMPL
K4088-14 - 1	CL-02-082619	36	-0.0510	-	7470A	PPB	28 Aug 2019 15:14:24	U	SMPL
CCV18 - 1	CCV18	20016	4.8853	-	7470A	PPB	28 Aug 2019 15:16:31	U	SMPL
CCB18 - 1	CCB18	25	-0.0537	-	7470A	PPB	28 Aug 2019 15:18:37	U	SMPL
K4092-11 - 1	EO-02-082619	133	-0.0270	-	7470A	PPB	28 Aug 2019 15:20:44	U	SMPL
K4093-10 - 1	PL-02-082619	88	-0.0381	-	7470A	PPB	28 Aug 2019 15:22:51	U	SMPL
K4506-02 - 1	RO-COMP	107	-0.0334	-	7470A	PPB	28 Aug 2019 15:24:58	U	SMPL
K4516-27 - 1	T1-S-B1-B4	94	-0.0366	-	7470A	PPB	28 Aug 2019 15:27:05	U	SMPL
K4516-30 - 1	T1-S-E	48	-0.0480	-	7470A	PPB	28 Aug 2019 15:29:12	U	SMPL
K4516-37 - 1	T2-S-B1-B4	99	-0.0354	-	7470A	PPB	28 Aug 2019 15:31:21	U	SMPL
K4516-40 - 1	T2-S-D	82	-0.0396	-	7470A	PPB	28 Aug 2019 15:33:30	U	SMPL
K4525-21 - 1	T9-12-13-A1-A4(0-1.9)	81	-0.0398	-	7470A	PPB	28 Aug 2019 15:35:40	U	SMPL
K4525-30 - 1	T9-12-13-B1-B4(0-1.9)	91	-0.0374	-	7470A	PPB	28 Aug 2019 15:37:58	U	SMPL
K4525-40 - 1	T9-12-13-C(0)	57	-0.0458	-	7470A	PPB	28 Aug 2019 15:40:04	U	SMPL
CCV19 - 1	CCV19	20527	5.0116	-	7470A	PPB	28 Aug 2019 15:42:11	U	SMPL
CCB19 - 1	CCB19	45	-0.0487	-	7470A	PPB	28 Aug 2019 15:44:17	U	SMPL
K4528-27 - 1	T9-12-13-R(0-1.9)	29	-0.0527	-	7470A	PPB	28 Aug 2019 15:46:24	U	SMPL
K4528-30 - 1	T9-12-13-S1-S4(0-1.9)	101	-0.0349	-	7470A	PPB	28 Aug 2019 15:48:32	U	SMPL
K4528-39 - 1	T9-12-13-T1-T4(0-1.9)	99	-0.0354	-	7470A	PPB	28 Aug 2019 15:50:40	U	SMPL
K4543-04 - 1	TP-2-S	88	-0.0381	-	7470A	PPB	28 Aug 2019 15:52:47	U	SMPL
K4543-08 - 1	TP-2-D	62	-0.0445	-	7470A	PPB	28 Aug 2019 15:54:54	U	SMPL
K4543-12 - 1	TP-3-S	76	-0.0411	-	7470A	PPB	28 Aug 2019 15:57:02	U	SMPL
K4543-16 - 1	TP-9-S	94	-0.0366	-	7470A	PPB	28 Aug 2019 15:59:10	U	SMPL
PB122630TB - 1	PB122561TB	72	-0.0421	-	7470A	PPB	28 Aug 2019 16:01:18	U	SMPL
K4086-08LX5 - 1		71	-0.0423	-	7470A	PPB	28 Aug 2019 16:03:27	U	SMPL
K4086-08A - 1		16567	4.0332	-	7470A	PPB	28 Aug 2019 16:05:33	U	SMPL
CCV20 - 1	CCV20	20606	5.0311	-	7470A	PPB	28 Aug 2019 16:10:52	U	SMPL
CCB20 - 1	CCB20	48	-0.0480	-	7470A	PPB	28 Aug 2019 16:12:59	U	SMPL