

TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: JIGNESH

Date: 08/29/2019

OVEN TEMP1: 98 °C TEMP: 104 °C

DATE & TIME IN: 08/28/2019 16:00 OUT: 08/29/2019 07:15

OVEN TEMP2: 104 °C TEMP: 104 °C

DATE & TIME IN: 08/29/2019 09:30 OUT: 08/29/2019 12:35

Run Number: LB104853

BalanceID: WC SC-3

OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB104853BL	LB104853BL	85.6774	85.6774	100	85.6774	85.6774	85.6774	0	0
2	K4577-01	GW-1	76.7443	76.7443	100	76.7607	76.7607	76.7607	0.0164	164
3	K4577-01DUP	GW-1DUP	79.5125	79.5125	100	79.5289	79.5289	79.5289	0.0164	164

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

VB 104853

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS K4577 WorkList ID : 130350 Department : Wet-Chemistry Date : 08-28-2019 14:24:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/03/2019	Water	K4577-01- F	TS	Cool 4 deg C	DVIR01	I22	GW-1	08/27/2019	SM2540 B

Date/Time 08/28/19 14:40
Received by: JS
Relinquished by: CD

Date/Time 08/28/19 16:15
Received by: JS
Relinquished by: JS