



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104854

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	08/28/2019	16:10
2	CAL2	1	Water	NA	NA	20.1	7.01	08/28/2019	16:11
3	CAL3	1	Water	NA	NA	20.1	10.03	08/28/2019	16:12
4	ICV	1	Water	NA	NA	20.1	6.99	08/28/2019	16:15
5	CCV1	1	Water	NA	NA	20.1	2.02	08/28/2019	16:16
6	K4587-01	1	Solid	20.02	20	22.3	5.65	08/28/2019	16:20
7	K4587-01DUP	1	Solid	20.03	20	22.4	5.66	08/28/2019	16:21
8	CCV2	1	Water	NA	NA	20.1	12.02	08/28/2019	16:23

2B 104854

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4587 WorkList ID : 130352 Department : Wet-Chemistry Date : 08-28-2019 15:48:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/28/2019	Solid	K4587-01 - X	pH	Cool 4 deg C	TULL02		FK-PS-01-057	08/28/2019	9045D

Date/Time 08/28/19 16:00
Received by: JA
Relinquished by: JA

Date/Time 08/28/19 17:00
Received by: CS
Relinquished by: JA