



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104866

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	08/29/2019	13:10
2	CAL2	1	Water	NA	NA	20.1	6.99	08/29/2019	13:11
3	CAL3	1	Water	NA	NA	20.1	10.02	08/29/2019	13:12
4	ICV	1	Water	NA	NA	20.1	7.00	08/29/2019	13:14
5	CCV1	1	Water	NA	NA	20.2	2.01	08/29/2019	13:15
6	K4604-01	1	Water	NA	NA	21.5	5.22	08/29/2019	13:16
7	K4604-01DUP	1	Water	NA	NA	21.6	5.23	08/29/2019	13:17
8	CCV2	1	Water	NA	NA	20.1	12.02	08/29/2019	13:19

1B104866

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4604 WorkList ID : 130398 Department : Wet-Chemistry Date : 08-29-2019 12:53:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/22/2019	Water	K4604-01-1	pH	Cool 4 deg C	STER10	I23	CELL-A-POOL-WATER-(DE-IK	08/22/2019	SM 4500-H B

Date/Time 08/29/19 13:00
Received by: JP
Relinquished by: cd

Date/Time 08/29/19 14:00
Received by: cd
Relinquished by: JP