

7471B


LB104869

INSTRUMENT ID: CV1

Linear

 μ Abs.:

59730



A= 0.0000e+000

B= 1.6754e-004

C= -1.5170e-002

Rho= 0.9998695

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.007	0.007	133	0.000	133				
0.2	0.200	0.205	0.005	1315	0.0 %	1315				
2.5	2.500	2.424	-0.076	14559	0.0 %	14559				
5.0	5.000	5.116	0.116	30627	0.0 %	30627				
7.5	7.500	7.456	-0.044	44593	0.0 %	44593				
10.0	10.000	9.992	-0.008	59730	0.0 %	59730				

LB104869 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 29 Aug 2019 13:44:29					
Sample ID	Extended ID	µ Abs	Conc.	Std CondMethod	Units	Date	Type	Type	
0.0 - 1	50-0	133	-	0.00007471B	PPB	29 Aug 2019 13:48:58	S	Std	
0.2 - 1	50-2	1315	-	0.20007471B	PPB	29 Aug 2019 13:51:16	S	Std	
2.5 - 1	52-5	14559	-	2.50007471B	PPB	29 Aug 2019 13:53:35	S	Std	
5.0 - 1	55-0	30627	-	5.00007471B	PPB	29 Aug 2019 13:55:54	S	Std	
7.5 - 1	57-5	44593	-	7.50007471B	PPB	29 Aug 2019 13:58:12	S	Std	
10.0 - 1	510-0	59730	-	10.00007471B	PPB	29 Aug 2019 14:00:30	S	Std	
ICV92 - 1	ICV92	24196	4.0386	-7471B	PPB	29 Aug 2019 14:05:08	U	SMPL	
ICB92 - 1	ICB92	73	-0.0029	-7471B	PPB	29 Aug 2019 14:07:24	U	SMPL	
CCV28 - 1	CCV28	30516	5.0974	-7471B	PPB	29 Aug 2019 14:09:40	U	SMPL	
CCB28 - 1	CCB28	81	-0.0016	-7471B	PPB	29 Aug 2019 14:11:58	U	SMPL	
CRA - 1	CRA	1320	0.2060	-7471B	PPB	29 Aug 2019 14:14:16	U	SMPL	
HighStd - 1	HighStd	60145	10.0614	-7471B	PPB	29 Aug 2019 14:16:32	U	SMPL	
ChkStd - 1	ChkStd	41657	6.9640	-7471B	PPB	29 Aug 2019 14:18:48	U	SMPL	
PB122678BL - 1	PBS	49	-0.0070	-7471B	PPB	29 Aug 2019 14:21:06	U	SMPL	
PB122678BS - 1	LCSS	22338	3.7273	-7471B	PPB	29 Aug 2019 14:23:23	U	SMPL	
K4095-13 - 1	SU-01-082819	1285	0.2001	-7471B	PPB	29 Aug 2019 14:25:40	U	SMPL	
K4095-13DUP - 1	SU-01-082819DUP	1249	0.1941	-7471B	PPB	29 Aug 2019 14:27:58	U	SMPL	
K4095-13MS - 1	SU-01-082819MS	24242	4.0463	-7471B	PPB	29 Aug 2019 14:30:15	U	SMPL	
K4095-13MSD - 1	SU-01-082819MSD	23954	3.9980	-7471B	PPB	29 Aug 2019 14:32:33	U	SMPL	
K4439-02 - 1	S-1-(1-3)	84019	14.0612	-7471B	PPB	29 Aug 2019 14:34:50	U	SMPL	
CCV29 - 1	CCV29	30576	5.1075	-7471B	PPB	29 Aug 2019 14:37:05	U	SMPL	
CCB29 - 1	CCB29	68	-0.0038	-7471B	PPB	29 Aug 2019 14:39:23	U	SMPL	
K4439-03 - 1	S-2-(1-3)	58847	9.8439	-7471B	PPB	29 Aug 2019 14:41:39	U	SMPL	
K4439-04 - 1	S-1-(5-7)	4083	0.6689	-7471B	PPB	29 Aug 2019 14:43:55	U	SMPL	
K4439-05 - 1	S-4A-(1-3)	6388	1.0551	-7471B	PPB	29 Aug 2019 14:46:12	U	SMPL	
K4439-06 - 1	S-3-(1-3)	15723	2.6190	-7471B	PPB	29 Aug 2019 14:48:29	U	SMPL	
K4439-07 - 1	S-5-(1-3)	8960	1.4860	-7471B	PPB	29 Aug 2019 14:50:45	U	SMPL	
K4439-08 - 1	S-2-(5-7)	3854	0.6305	-7471B	PPB	29 Aug 2019 14:53:02	U	SMPL	
K4439-09 - 1	S-4A-(5-7)	1707	0.2708	-7471B	PPB	29 Aug 2019 14:55:20	U	SMPL	
K4439-10 - 1	S-5-(5-7)	4852	0.7977	-7471B	PPB	29 Aug 2019 14:57:38	U	SMPL	
K4579-01 - 1	PAR-12-A1-A4-(0-3.5)	4256	0.6979	-7471B	PPB	29 Aug 2019 14:59:56	U	SMPL	
K4579-08 - 1	PAR-12-B1-B4-(0-3.5)	4966	0.8168	-7471B	PPB	29 Aug 2019 15:02:12	U	SMPL	
CCV30 - 1	CCV30	30218	5.0475	-7471B	PPB	29 Aug 2019 15:04:28	U	SMPL	
CCB30 - 1	CCB30	49	-0.0070	-7471B	PPB	29 Aug 2019 15:06:44	U	SMPL	
K4580-03 - 1	T12-B1-B4-(0-3.5)	7433	1.2301	-7471B	PPB	29 Aug 2019 15:09:00	U	SMPL	
K4580-14 - 1	T12-A1-A4-(0-3.5)	2043	0.3271	-7471B	PPB	29 Aug 2019 15:11:17	U	SMPL	
K4581-02 - 1	T12-C1-C4-(0-3.5)	1858	0.2961	-7471B	PPB	29 Aug 2019 15:13:33	U	SMPL	
K4581-11 - 1	T12-D1-D4-(0-3.5)	6686	1.1050	-7471B	PPB	29 Aug 2019 15:15:49	U	SMPL	
K4582-01 - 1	RT-3422	2131	0.3419	-7471B	PPB	29 Aug 2019 15:18:05	U	SMPL	
K4587-01 - 1	FK-PS-01-057	4437	0.7282	-7471B	PPB	29 Aug 2019 15:20:22	U	SMPL	
K4593-01 - 1	S-3-(5-7)	2628	0.4251	-7471B	PPB	29 Aug 2019 15:22:39	U	SMPL	
PB122680BL - 1	PBS	84	-0.0011	-7471B	PPB	29 Aug 2019 15:24:57	U	SMPL	
PB122680BS - 1	LCSS	22393	3.7365	-7471B	PPB	29 Aug 2019 15:27:15	U	SMPL	
K4514-05 - 1	S602-M-2.0-2.5-082319	2762	0.4476	-7471B	PPB	29 Aug 2019 15:29:32	U	SMPL	
CCV31 - 1	CCV31	29072	4.8555	-7471B	PPB	29 Aug 2019 15:31:48	U	SMPL	
CCB31 - 1	CCB31	82	-0.0014	-7471B	PPB	29 Aug 2019 15:34:04	U	SMPL	
K4514-05DUP - 1	S602-M-2.0-2.5-082319DUP	2848	0.4620	-7471B	PPB	29 Aug 2019 15:36:22	U	SMPL	
K4514-05MS - 1	S602-M-2.0-2.5-082319MS	23172	3.8670	-7471B	PPB	29 Aug 2019 15:50:58	U	SMPL	
K4514-05MSD - 1	S602-M-2.0-2.5-082319MSD	24186	4.0369	-7471B	PPB	29 Aug 2019 15:53:14	U	SMPL	
K4583-01 - 1	GS-STOCKPILE	4826	0.7934	-7471B	PPB	29 Aug 2019 15:55:31	U	SMPL	
K4591-01 - 1	SB-3-25-27	671	0.0972	-7471B	PPB	29 Aug 2019 15:57:48	U	SMPL	
K4591-03 - 1	SB-3-30-32	1858	0.2961	-7471B	PPB	29 Aug 2019 16:00:05	U	SMPL	
K4591-06 - 1	SB-3-35-37	2082	0.3336	-7471B	PPB	29 Aug 2019 16:02:23	U	SMPL	
K4596-01 - 1	S403-K1-1.0-1.5-082819	1698	0.2693	-7471B	PPB	29 Aug 2019 16:04:40	U	SMPL	
K4596-05 - 1	S403-K1-1.0-1.5-082819-FD	2185	0.3509	-7471B	PPB	29 Aug 2019 16:06:58	U	SMPL	
K4095-13LX5 - 1		449	0.0601	-7471B	PPB	29 Aug 2019 16:09:14	U	SMPL	
CCV32 - 1	CCV32	29078	4.8565	-7471B	PPB	29 Aug 2019 16:11:31	U	SMPL	
CCB32 - 1	CCB32	328	0.0398	-7471B	PPB	29 Aug 2019 16:13:47	U	SMPL	
K4095-13A - 1		24082	4.0195	-7471B	PPB	29 Aug 2019 16:16:03	U	SMPL	
K4514-05LX5 - 1		529	0.0735	-7471B	PPB	29 Aug 2019 16:18:20	U	SMPL	
K4514-05A - 1		24699	4.1228	-7471B	PPB	29 Aug 2019 16:20:37	U	SMPL	
K4439-02DLX3 - 1		26444	4.4152	-7471B	PPB	29 Aug 2019 16:22:54	U	SMPL	
CCV33 - 1	CCV33	29260	4.8870	-7471B	PPB	29 Aug 2019 16:25:12	U	SMPL	
CCB33 - 1	CCB33	70	-0.0034	-7471B	PPB	29 Aug 2019 16:27:30	U	SMPL	