

LB104871

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : CT

Instrument ID : Konelab

8/29/2019 14:50

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.983	0.0	0.124	
ICB1	0.002	0.0	0.016	
CCV1	0.999	0.0	0.126	
CCB1	0.003	0.0	0.016	
PB122639BL	0.008	0.0	0.017	
PB122639BS	0.987	0.0	0.124	
K4577-01	0.003	0.0	0.016	
K4577-01DUP	0.004	0.0	0.016	
K4577-01MS	0.865	0.0	0.111	
K4577-01MSD	0.869	0.0	0.111	
CCV2	0.984	0.0	0.124	
CCB2	0.000	0.0	0.016	

N	12
Mean	0.476
SD	0.4949
CV%	104.08

CT
8/29/19

Aquakem v. 7.2AQ1

Results from time period:

Thu Aug 29 14:11:20 2019

Thu Aug 29 14:41:56 2019

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0549	mg/l	8/29/2019 14:11	
0.1PPM	A	Phenolics-	P	0.103	mg/l	8/29/2019 14:11	
0.25PPM	A	Phenolics-	P	0.253	mg/l	8/29/2019 14:11	
0.50PPM	A	Phenolics-	P	0.5033	mg/l	8/29/2019 14:11	
1.0PPM	A	Phenolics-	P	0.9746	mg/l	8/29/2019 14:11	
2.0PPM	A	Phenolics-	P	2.0112	mg/l	8/29/2019 14:11	
ICV1	S	Phenolics-	P	0.9829	mg/l	8/29/2019 14:37	
ICB1	S	Phenolics-	P	0.0022	mg/l	8/29/2019 14:37	
CCV1	S	Phenolics-	P	0.9991	mg/l	8/29/2019 14:37	
CCB1	S	Phenolics-	P	0.0028	mg/l	8/29/2019 14:38	
PB122639BL	S	Phenolics-	P	0.008	mg/l	8/29/2019 14:38	
PB122639BS	S	Phenolics-	P	0.9867	mg/l	8/29/2019 14:38	
K4577-01	S	Phenolics-	P	0.0034	mg/l	8/29/2019 14:38	
K4577-01DUP	S	Phenolics-	P	0.0036	mg/l	8/29/2019 14:38	
K4577-01MS	S	Phenolics-	P	0.8646	mg/l	8/29/2019 14:38	
K4577-01MSD	S	Phenolics-	P	0.8688	mg/l	8/29/2019 14:38	
CCV2	S	Phenolics-	P	0.9836	mg/l	8/29/2019 14:41	
CCB2	S	Phenolics-	P	0.0005	mg/l	8/29/2019 14:41	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : CT

Instrument ID : Konelab

8/29/2019 14:56

Test Phenolics-


Accepted 8/29/2019 14:13

Factor 9.103

Bias 0.016

Coeff. of det. 0.999706

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.022	0.0549	0.0500	
2	Phenol-2	0.027	0.1030	0.1000	
3	Phenol-2	0.044	0.2530	0.2500	
4	Phenol-2	0.071	0.5033	0.5000	
5	Phenol-2	0.123	0.9746	1.0000	
6	Phenol-2	0.237	2.0112	2.0000	