



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104877

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	08/29/2019	16:40
2	CAL2	1	Water	NA	NA	20.1	7.00	08/29/2019	16:41
3	CAL3	1	Water	NA	NA	20.1	10.03	08/29/2019	16:43
4	ICV	1	Water	NA	NA	20.1	7.01	08/29/2019	16:45
5	CCV1	1	Water	NA	NA	20.1	2.02	08/29/2019	16:46
6	K4615-01	1	Water	NA	NA	21.2	7.80	08/29/2019	17:00
7	K4615-01DUP	1	Water	NA	NA	21.3	7.81	08/29/2019	17:01
8	CCV2	1	Water	NA	NA	20.1	12.02	08/29/2019	17:02

28106877

WORKLIST(Hardcopy Internal Chain)

WorkList Name :	PH K4615	WorkList ID :	130406	Department :	Wet-Chemistry	Date :	08-29-2019 14:51:41		
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/29/2019	Water	K4615-01	pH	Cool 4 deg C	HOME01	B12	NR-KIT-TAP-NEW-FILTER	08/29/2019	SM 4500-H B

Date/Time 08/29/19 15:00
Received by: J9
Relinquished by: CA

Date/Time 08/29/19 16:00
Received by: CF
Relinquished by: JD