



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB104878

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2445
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2391
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	08/29/2019	16:40
2	CAL2	1	Water	NA	NA	20.1	7.00	08/29/2019	16:41
3	CAL3	1	Water	NA	NA	20.1	10.03	08/29/2019	16:43
4	ICV	1	Water	NA	NA	20.1	7.01	08/29/2019	16:45
5	CCV1	1	Water	NA	NA	20.1	2.02	08/29/2019	16:46
6	K4613-10	1	Water	NA	NA	16.9	6.17	08/29/2019	16:47
7	K4613-10DUP	1	Water	NA	NA	16.8	6.18	08/29/2019	16:48
8	CCV2	1	Water	NA	NA	20.1	12.02	08/29/2019	16:50

19104878

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY K4613 WorkList ID : 130409 Department : Wet-Chemistry Date : 08-29-2019 16:34:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/29/2019	Water	K4613-10	☒ Corrosivity	Cool 4 deg C	PSEG01	J22	TP-11-GW	08/29/2019	9040C

Date/Time 08/29/19 16:35
Received by: Jg CT
Relinquished by: _____

Date/Time 08/29/19 14:35
Received by: CT
Relinquished by: Jg