



PERCENT SOLID

Supervisor: chirag
Analyst: jignesh
Date: 8/31/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:35
In Date: 08/30/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:27
Out Date: 08/31/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB104887

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
K4409-13	S301-J-4.0-4.5-081919	80	1.13	9.82	8.57	85.6	
K4409-15	S301-S-5.0-5.5-081919	81	1.15	9.72	7.97	79.6	
K4409-16	S301-S-6.0-6.5-081919	82	1.13	9.62	8.37	85.3	
K4409-17	S301-S-7.0-7.5-081919	83	1.15	9.88	8.7	86.5	
K4409-18	S301-S-8.0-8.5-081919	84	1.14	9.84	8.73	87.2	
K4409-19	S301-S-9.0-9.5-081919	85	1.15	9.89	8.79	87.4	
K4514-15	S301A-J-3.0-3.5-082319	86	1.15	9.52	7.79	79.3	
K4514-16	S301A-J-4.0-4.5-082319	87	1.13	9.67	8.15	82.2	
K4554-01	SB-13- (0-2)	92	1.15	9.57	8.46	86.8	
K4554-02	SB-13- (13-15)	93	1.14	9.92	8.89	88.3	
K4554-03	SB-14- (0-2)	94	1.14	9.6	8.28	84.4	
K4554-04	SB-14- (13-15)	95	1.14	9.8	8.81	88.6	
K4554-05	SB-12- (0-2)	96	1.12	9.81	8.96	90.2	
K4554-06	SB-12- (13-15)	97	1.14	9.89	9.04	90.3	
K4554-07	SB-15- (0-2)	98	1.14	9.87	9.11	91.3	
K4554-08	SB-15- (13-15)	99	1.14	9.84	9.3	93.8	
K4554-19	SB-10- (0-2)	100	1.14	9.94	9.15	91.0	
K4554-20	SB-10- (12-14)	101	1.12	9.75	8.6	86.7	
K4554-21	SB-9- (0-2)	102	1.13	9.66	8.74	89.2	
K4554-22	SB-9- (7-9)	103	1.12	9.77	8.71	87.7	
K4554-23	SB-11- (0-2)	104	1.13	9.75	8.82	89.2	
K4554-24	SB-11- (13-15)	105	1.14	9.85	8.9	89.1	
K4554-25	SB-7- (0-2)	106	1.12	9.88	8.25	81.4	
K4554-26	SB-7- (13-15)	107	1.12	9.66	8.88	90.9	
K4554-27	SB-6- (0-2)	108	1.13	9.98	8.99	88.8	
K4554-28	SB-6- (6-8)	109	1.13	9.99	8.97	88.5	
K4554-39	SB-8- (0-2)	110	1.14	9.75	8.48	85.2	
K4554-40	SB-8- (5-7)	111	1.15	9.9	8.86	88.1	



PERCENT SOLID

Supervisor: chirag
Analyst: jignesh
Date: 8/31/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:35
In Date: 08/30/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:27
Out Date: 08/31/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB104887

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
K4554-41	SB-3-(0-2)	112	1.15	9.63	7.91	79.7	
K4554-42	SB-3-(8-10)	113	1.14	9.94	8.35	81.9	
K4554-43	SB-4-(0-2)	114	1.12	9.91	8.69	86.1	
K4554-44	SB-4-(7-9)	115	1.13	9.78	8.67	87.2	
K4554-45	SB-2-(0-2)	116	1.14	9.67	7.94	79.7	
K4554-46	SB-2-(13-15)	117	1.13	9.72	7.44	73.5	
K4554-55	SB-5-(0-2)	118	1.15	9.72	8.7	88.1	
K4554-56	SB-5-(8-10)	119	1.13	9.84	8.09	79.9	
K4554-57	SB-1-(0-2)	120	1.14	9.83	8.76	87.7	
K4554-58	SB-1-(13-15)	121	1.13	9.71	7.36	72.6	
K4554-59	SB-16-(8-10)	122	1.14	9.68	8.56	86.9	
K4573-03	S700B-N-2.0-2.5-082719	88	1.13	9.74	8.22	82.3	
K4574-15	S702B-N-2.0-2.5-082719	89	1.12	9.71	7.35	72.5	
K4575-14	S703A-N-1.0-1.5-082719	90	1.14	9.82	8.64	86.4	
K4594-03	S704A-N-2.0-2.5-082819	91	1.15	9.8	7.79	76.8	
K4620-24	TB-06-082919	123	1.00	2.00	2.00	100.0	TRIP BLANK
K4622-01	PAR-11-A1-A4-(0-4)	1	1.14	9.87	8.16	80.4	
K4622-02	PAR-11-A1-(0)	2	1.12	9.81	8.02	79.4	
K4622-03	PAR-11-A2-(0)	3	1.13	9.71	8.22	82.6	
K4622-04	PAR-11-A3-(0)	4	1.13	9.83	7.7	75.5	
K4622-05	PAR-11-A4-(0)	5	1.13	9.9	7.73	75.3	
K4622-06	PAR-11-A1-(4)	6	1.13	9.82	8.03	79.4	
K4622-07	PAR-11-A2-(4)	7	1.12	9.84	8.27	82.0	
K4622-08	PAR-11-A3-(4)	8	1.13	9.8	7.83	77.3	
K4622-09	PAR-11-A4-(4)	9	1.14	9.71	7.81	77.8	
K4622-10	PAR-11-B1-B4-(0-4)	10	1.13	9.77	8.28	82.8	
K4622-11	PAR-11-B1-(0)	11	1.14	9.75	8.43	84.7	
K4622-12	PAR-11-B2-(0)	12	1.13	9.75	7.88	78.3	



PERCENT SOLID

Supervisor: chirag
Analyst: jignesh
Date: 8/31/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:35
In Date: 08/30/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:27
Out Date: 08/31/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB104887

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
K4622-13	PAR-11-B3- (0)	13	1.12	9.85	8.38	83.2	
K4622-14	PAR-11-B4- (0)	14	1.14	9.57	8.33	85.3	
K4622-15	PAR-11-B1- (4)	15	1.13	9.62	8.54	87.3	
K4622-16	PAR-11-B2- (4)	16	1.14	9.62	7.63	76.5	
K4622-17	PAR-11-B3- (4)	17	1.13	9.7	8.5	86.0	
K4622-18	PAR-11-B4- (4)	18	1.13	9.56	7.91	80.4	
K4622-19	PAR-11-C1-C4- (0-4)	19	1.14	9.59	8.34	85.2	
K4622-20	PAR-11-C1- (0)	20	1.14	9.81	8.38	83.5	
K4623-01	PAR-11-C2- (0)	21	1.14	9.77	8.37	83.8	
K4623-02	PAR-11-C3- (0)	22	1.12	9.67	8.12	81.9	
K4623-03	PAR-11-C4- (0)	23	1.12	9.86	8.65	86.2	
K4623-04	PAR-11-C1- (4)	24	1.13	9.62	8.14	82.6	
K4623-05	PAR-11-C2- (4)	25	1.13	9.74	8.07	80.6	
K4623-06	PAR-11-C3- (4)	26	1.12	9.92	8.5	83.9	
K4623-07	PAR-11-C4- (4)	27	1.12	9.7	8.38	84.6	
K4623-08	PAR-11-D1-D4- (0-4)	28	1.12	9.83	7.47	72.9	
K4623-09	PAR-11-D1- (0)	29	1.13	9.88	8.55	84.8	
K4623-10	PAR-11-D2- (0)	30	1.12	9.92	9.00	89.5	
K4623-11	PAR-11-D3- (0)	31	1.12	9.63	8.66	88.6	
K4623-12	PAR-11-D4- (0)	32	1.12	9.63	8.6	87.9	
K4623-13	PAR-11-D1- (4)	33	1.13	9.75	8.42	84.6	
K4623-14	PAR-11-D2- (4)	34	1.13	9.58	8.86	91.5	
K4623-15	PAR-11-D3- (4)	35	1.14	9.89	8.85	88.1	
K4623-16	PAR-11-D4- (4)	36	1.14	9.62	8.48	86.6	
K4623-17	PAR-11-E1-E4- (0-4)	37	1.13	9.73	8.53	86.0	
K4623-18	PAR-11-E1- (0)	38	1.12	9.72	9.03	92.0	
K4623-19	PAR-11-E2- (0)	39	1.13	9.7	9.00	91.8	
K4623-20	PAR-11-E3- (0)	40	1.13	9.71	8.16	81.9	



PERCENT SOLID

Supervisor: chirag
Analyst: jignesh
Date: 8/31/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:35
In Date: 08/30/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:27
Out Date: 08/31/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB104887

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
K4624-01	PAR-11-E4- (0)	41	1.14	9.83	8.95	89.9	
K4624-02	PAR-11-E1- (4)	42	1.13	9.81	8.75	87.8	
K4624-03	PAR-11-E2- (4)	43	1.12	9.6	8.73	89.7	
K4624-04	PAR-11-E3- (4)	44	1.12	9.72	8.55	86.4	
K4624-05	PAR-11-E4- (4)	45	1.13	9.71	9.2	94.1	
K4624-06	T-11-A1-A4- (0-4)	46	1.12	9.74	9.00	91.4	
K4624-07	T-11-A1- (0)	47	1.13	9.85	8.52	84.7	
K4624-08	T-11-A2- (0)	48	1.14	9.58	8.7	89.6	
K4624-09	T-11-A3- (0)	49	1.13	9.66	8.72	89.0	
K4624-10	T-11-A4- (0)	50	1.12	9.95	9.14	90.8	
K4624-11	T-11-A1- (4)	51	1.13	9.91	8.31	81.8	
K4624-12	T-11-A2- (4)	52	1.14	9.77	9.05	91.7	
K4624-13	T-11-A3- (4)	53	1.13	9.72	8.93	90.8	
K4624-14	T-11-A4- (4)	54	1.12	9.83	9.33	94.3	
K4625-01	HF-01	76	1.13	9.69	7.61	75.7	
K4625-02	HF-01	77	1.13	9.81	8.61	86.2	
K4625-03	MM-01	78	1.13	9.62	7.72	77.6	
K4625-04	MM-01	79	1.12	9.6	6.3	61.1	
K4626-01	1-2-EWC-1	55	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100 % SOLIDS
K4626-02	1-2-EWC-2	56	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100 % SOLIDS
K4626-03	1-2-EWC-3	57	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100 % SOLIDS
K4626-04	1-2-OWS-1	58	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100 % SOLIDS
K4626-05	1-2-SWG-1	59	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100 % SOLIDS
K4626-06	1-2-EXFC-1	60	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100 % SOLIDS
K4626-07	3-WC-1	61	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100 % SOLIDS
K4626-08	4-WDC-1	62	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100 % SOLIDS
K4627-01	7-6-EXJC-1	63	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100 % SOLIDS
K4627-02	6-SWG-1	64	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100 % SOLIDS



PERCENT SOLID

Supervisor: chirag
Analyst: jignesh
Date: 8/31/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:35
In Date: 08/30/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:27
Out Date: 08/31/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB104887

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
K4627-03	6-EWC-1	65	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100 % SOLIDS
K4627-04	7-6-EWC-2	66	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100 % SOLIDS
K4627-05	6-EWG-1	67	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100 % SOLIDS
K4627-06	7-6-DC-1	68	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100 % SOLIDS
K4627-07	7-6-IWC-1	69	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100 % SOLIDS
K4627-08	6-IWG-1	70	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100 % SOLIDS
K4627-09	6-BFJC-1	71	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100 % SOLIDS
K4627-10	7-6-IDC-1	72	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100 % SOLIDS
K4627-11	6-IEJ-1	73	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100 % SOLIDS
K4627-12	7-6-FDS-1	74	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100 % SOLIDS
K4627-13	7-6-FSC-1	75	1.00	2.00	2.00	100.0	CAULKING SAMPLE 100 % SOLIDS
K4629-01	390821755.1	124	1.12	9.64	1.42	3.5	SLUDGE SAMPLE
K4637-05	SVOC-GPC-BLANK	125	1.00	2.00	2.00	100.0	
K4637-06	PEST-GPC-BLANK	126	1.00	2.00	2.00	100.0	
K4637-07	PEST-GPC-BLANK-SPIKE	127	1.00	2.00	2.00	100.0	
K4637-08	PCB-GPC-BLANK	128	1.00	2.00	2.00	100.0	
K4637-09	PCB-GPC-BLANK-SPIKE	129	1.00	2.00	2.00	100.0	
K4637-10	SVOC-GPC2-BLANK	130	1.00	2.00	2.00	100.0	
K4637-11	PEST-GPC2-BLANK	131	1.00	2.00	2.00	100.0	
K4637-12	PEST-GPC2-BLANK-SPIKE	132	1.00	2.00	2.00	100.0	
K4637-13	PCB-GPC2-BLANK	133	1.00	2.00	2.00	100.0	
K4637-14	PCB-GPC2-BLANK -SPIKE	134	1.00	2.00	2.00	100.0	
K4645-01	S700A1-N-0.0-0.5-083019	135	1.11	9.6	8.17	83.2	
K4645-11	S700B1-N-0.0-0.5-083019	136	1.15	9.53	9.3	97.3	
K4646-01	S602A-M-1.0-1.5-083019	137	1.13	9.74	9.07	92.2	
K4646-14	EB-7-08302019	138	1.00	2.00	2.00	100.0	T.B.
K4647-01	S401-K1-1.0-1.5-082619-FD	139	1.14	9.84	7.22	69.9	
K4647-02	S400-K1-1.0-1.5-082619	140	1.14	9.86	7.76	75.9	



PERCENT SOLID

Supervisor: chirag
Analyst: jignesh
Date: 8/31/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 17:35
In Date: 08/30/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:27
Out Date: 08/31/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB104887

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
K4647-03	S602-M-1.0-1.5-082319	141	1.14	9.82	8.82	88.5	
K4647-04	S402-K1-1.0-1.5-082619	142	1.13	9.83	8.7	87.0	
K4647-05	S401-K1-1.0-1.5-082619	143	1.15	9.64	7.32	72.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-083019 WorkList ID : 130417 Department : Wet-Chemistry Date : 08-30-2019 07:45:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/20/2019	Solid	K4409-13	Percent Solids	Cool 4 deg C	JACO05	K11	S301-J-4.0-4.5-081919	08/19/2019	Chemtech -SO
08/20/2019	Solid	K4409-15	Percent Solids	Cool 4 deg C	JACO05	K11	S301-S-5.0-5.5-081919	08/19/2019	Chemtech -SO
08/20/2019	Solid	K4409-16	Percent Solids	Cool 4 deg C	JACO05	K11	S301-S-6.0-6.5-081919	08/19/2019	Chemtech -SO
08/20/2019	Solid	K4409-17	Percent Solids	Cool 4 deg C	JACO05	K11	S301-S-7.0-7.5-081919	08/19/2019	Chemtech -SO
08/20/2019	Solid	K4409-18	Percent Solids	Cool 4 deg C	JACO05	K11	S301-S-8.0-8.5-081919	08/19/2019	Chemtech -SO
08/20/2019	Solid	K4409-19	Percent Solids	Cool 4 deg C	JACO05	K11	S301-S-9.0-9.5-081919	08/19/2019	Chemtech -SO
08/26/2019	Solid	K4514-15	Percent Solids	Cool 4 deg C	JACO05	I42	S301A-J-3.0-3.5-082319	08/23/2019	Chemtech -SO
08/26/2019	Solid	K4514-16	Percent Solids	Cool 4 deg C	JACO05	I42	S301A-J-4.0-4.5-082319	08/23/2019	Chemtech -SO
09/02/2019	Solid	K4554-01	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-13-(0-2)	08/26/2019	Chemtech -SO
09/02/2019	Solid	K4554-02	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-13-(13-15)	08/26/2019	Chemtech -SO
09/02/2019	Solid	K4554-03	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-14-(0-2)	08/26/2019	Chemtech -SO
09/02/2019	Solid	K4554-04	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-14-(13-15)	08/26/2019	Chemtech -SO
09/02/2019	Solid	K4554-05	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-12-(0-2)	08/26/2019	Chemtech -SO
09/02/2019	Solid	K4554-06	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-12-(13-15)	08/26/2019	Chemtech -SO
09/02/2019	Solid	K4554-07	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-15-(0-2)	08/26/2019	Chemtech -SO
09/02/2019	Solid	K4554-08	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-15-(13-15)	08/26/2019	Chemtech -SO
09/03/2019	Solid	K4554-19	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-10-(0-2)	08/27/2019	Chemtech -SO
09/03/2019	Solid	K4554-20	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-10-(12-14)	08/27/2019	Chemtech -SO
09/03/2019	Solid	K4554-21	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-9-(0-2)	08/27/2019	Chemtech -SO
09/03/2019	Solid	K4554-22	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-9-(7-9)	08/27/2019	Chemtech -SO
09/03/2019	Solid	K4554-23	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-11-(0-2)	08/27/2019	Chemtech -SO
09/03/2019	Solid	K4554-24	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-11-(13-15)	08/27/2019	Chemtech -SO

Date/Time 08/30/19 18:35 Date/Time 08/30/19 18:00
 Received by: [Signature] Received by: [Signature]
 Relinquished by: [Signature] Relinquished by: [Signature]

WORKLIST(Hardcopy Internal Chain)

15104887

WorkList Name : %1-083019

WorkList ID : 130417

Department : Wet-Chemistry

Date : 08-30-2019 07:45:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/03/2019	Solid	K4554-25	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-7-(0-2)	08/27/2019	Chemtech -SO
09/03/2019	Solid	K4554-26	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-7-(13-15)	08/27/2019	Chemtech -SO
09/03/2019	Solid	K4554-27	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-6-(0-2)	08/27/2019	Chemtech -SO
09/03/2019	Solid	K4554-28	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-6-(6-8)	08/27/2019	Chemtech -SO
09/04/2019	Solid	K4554-39	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-8-(0-2)	08/27/2019	Chemtech -SO
09/04/2019	Solid	K4554-40	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-8-(5-7)	08/27/2019	Chemtech -SO
09/04/2019	Solid	K4554-41	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-3-(0-2)	08/28/2019	Chemtech -SO
09/04/2019	Solid	K4554-42	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-3-(8-10)	08/28/2019	Chemtech -SO
09/04/2019	Solid	K4554-43	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-4-(0-2)	08/28/2019	Chemtech -SO
09/04/2019	Solid	K4554-44	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-4-(7-9)	08/28/2019	Chemtech -SO
09/04/2019	Solid	K4554-45	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-2-(0-2)	08/28/2019	Chemtech -SO
09/04/2019	Solid	K4554-46	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-2-(13-15)	08/28/2019	Chemtech -SO
09/05/2019	Solid	K4554-55	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-5-(0-2)	08/29/2019	Chemtech -SO
09/05/2019	Solid	K4554-56	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-5-(8-10)	08/29/2019	Chemtech -SO
09/05/2019	Solid	K4554-57	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-1-(0-2)	08/29/2019	Chemtech -SO
09/05/2019	Solid	K4554-58	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-1-(13-15)	08/29/2019	Chemtech -SO
09/05/2019	Solid	K4554-59	Percent Solids	Cool 4 deg C	AKRF02	I33	SB-16-(8-10)	08/29/2019	Chemtech -SO
08/28/2019	Solid	K4573-03	Percent Solids	Cool 4 deg C	JACO05	I21	S700B-N-2.0-2.5-082719	08/27/2019	Chemtech -SO
08/28/2019	Solid	K4574-15	Percent Solids	Cool 4 deg C	JACO05	I21	S702B-N-2.0-2.5-082719	08/27/2019	Chemtech -SO
08/28/2019	Solid	K4575-14	Percent Solids	Cool 4 deg C	JACO05	I21	S703A-N-1.0-1.5-082719	08/27/2019	Chemtech -SO
08/29/2019	Solid	K4594-03	Percent Solids	Cool 4 deg C	JACO05	I42	S704A-N-2.0-2.5-082819	08/28/2019	Chemtech -SO
08/30/2019	Solid	K4620-24	Percent Solids	Cool 4 deg C	JACO05	I41	TB-06-082919	08/29/2019	Chemtech -SO

Date/Time 08/30/19 15:35
 Received by: JP
 Relinquished by: JP

Date/Time 08/30/19 18:00
 Received by: JP
 Relinquished by: JP

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-083019 WorkList ID : 130417 Department : Wet-Chemistry Date : 08-30-2019 07:45:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/08/2019	Solid	K4622-01	Percent Solids	Cool 4 deg C	RAIL01	I41	PAR-11-A1-A4-(0-4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4622-02	Percent Solids	Cool 4 deg C	RAIL01	I41	PAR-11-A1-(0)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4622-03	Percent Solids	Cool 4 deg C	RAIL01	I41	PAR-11-A2-(0)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4622-04	Percent Solids	Cool 4 deg C	RAIL01	I41	PAR-11-A3-(0)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4622-05	Percent Solids	Cool 4 deg C	RAIL01	I41	PAR-11-A4-(0)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4622-06	Percent Solids	Cool 4 deg C	RAIL01	I41	PAR-11-A1-(4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4622-07	Percent Solids	Cool 4 deg C	RAIL01	I41	PAR-11-A2-(4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4622-08	Percent Solids	Cool 4 deg C	RAIL01	I41	PAR-11-A3-(4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4622-09	Percent Solids	Cool 4 deg C	RAIL01	I41	PAR-11-A4-(4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4622-10	Percent Solids	Cool 4 deg C	RAIL01	I41	PAR-11-B1-B4-(0-4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4622-11	Percent Solids	Cool 4 deg C	RAIL01	I41	PAR-11-B1-(0)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4622-12	Percent Solids	Cool 4 deg C	RAIL01	I41	PAR-11-B2-(0)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4622-13	Percent Solids	Cool 4 deg C	RAIL01	I41	PAR-11-B3-(0)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4622-14	Percent Solids	Cool 4 deg C	RAIL01	I41	PAR-11-B4-(0)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4622-15	Percent Solids	Cool 4 deg C	RAIL01	I41	PAR-11-B1-(4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4622-16	Percent Solids	Cool 4 deg C	RAIL01	I41	PAR-11-B2-(4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4622-17	Percent Solids	Cool 4 deg C	RAIL01	I41	PAR-11-B3-(4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4622-18	Percent Solids	Cool 4 deg C	RAIL01	I41	PAR-11-B4-(4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4622-19	Percent Solids	Cool 4 deg C	RAIL01	I41	PAR-11-C1-C4-(0-4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4622-20	Percent Solids	Cool 4 deg C	RAIL01	I41	PAR-11-C1-(0)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4623-01	Percent Solids	Cool 4 deg C	RAIL01	I61	PAR-11-C2-(0)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4623-02	Percent Solids	Cool 4 deg C	RAIL01	I61	PAR-11-C3-(0)	08/29/2019	Chemtech -SO

Date/Time 08/30/19 15:35
 Received by: J9
 Relinquished by: CS

Date/Time 08/31/19 18:00
 Received by: ES
 Relinquished by: JS

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-083019 WorkList ID : 130417 Department : Wet-Chemistry Date : 08-30-2019 07:45:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/08/2019	Solid	K4623-03	Percent Solids	Cool 4 deg C	RAIL01	I61	PAR-11-C4-(0)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4623-04	Percent Solids	Cool 4 deg C	RAIL01	I61	PAR-11-C1-(4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4623-05	Percent Solids	Cool 4 deg C	RAIL01	I61	PAR-11-C2-(4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4623-06	Percent Solids	Cool 4 deg C	RAIL01	I61	PAR-11-C3-(4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4623-07	Percent Solids	Cool 4 deg C	RAIL01	I61	PAR-11-C4-(4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4623-08	Percent Solids	Cool 4 deg C	RAIL01	I61	PAR-11-D1-D4-(0-4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4623-09	Percent Solids	Cool 4 deg C	RAIL01	I61	PAR-11-D1-(0)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4623-10	Percent Solids	Cool 4 deg C	RAIL01	I61	PAR-11-D2-(0)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4623-11	Percent Solids	Cool 4 deg C	RAIL01	I61	PAR-11-D3-(0)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4623-12	Percent Solids	Cool 4 deg C	RAIL01	I61	PAR-11-D4-(0)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4623-13	Percent Solids	Cool 4 deg C	RAIL01	I61	PAR-11-D1-(4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4623-14	Percent Solids	Cool 4 deg C	RAIL01	I61	PAR-11-D2-(4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4623-15	Percent Solids	Cool 4 deg C	RAIL01	I61	PAR-11-D3-(4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4623-16	Percent Solids	Cool 4 deg C	RAIL01	I61	PAR-11-D4-(4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4623-17	Percent Solids	Cool 4 deg C	RAIL01	I61	PAR-11-E1-E4-(0-4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4623-18	Percent Solids	Cool 4 deg C	RAIL01	I61	PAR-11-E1-(0)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4623-19	Percent Solids	Cool 4 deg C	RAIL01	I61	PAR-11-E2-(0)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4623-20	Percent Solids	Cool 4 deg C	RAIL01	I61	PAR-11-E3-(0)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4624-01	Percent Solids	Cool 4 deg C	RAIL01	I23	PAR-11-E4-(0)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4624-02	Percent Solids	Cool 4 deg C	RAIL01	I23	PAR-11-E1-(4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4624-03	Percent Solids	Cool 4 deg C	RAIL01	I23	PAR-11-E2-(4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4624-04	Percent Solids	Cool 4 deg C	RAIL01	I23	PAR-11-E3-(4)	08/29/2019	Chemtech -SO

Date/Time 08/30/19 15:35 Date/Time 08/30/19 18:00
 Received by: JP Received by: ad
 Relinquished by: JP Relinquished by: JP

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-083019

WorkList ID : 130417

Department : Wet-Chemistry

Date : 08-30-2019 07:45:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/08/2019	Solid	K4624-05	Percent Solids	Cool 4 deg C	RAIL01	I23	PAR-11-E4-(4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4624-06	Percent Solids	Cool 4 deg C	RAIL01	I23	T-11-A1-A4-(0-4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4624-07	Percent Solids	Cool 4 deg C	RAIL01	I23	T-11-A1-(0)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4624-08	Percent Solids	Cool 4 deg C	RAIL01	I23	T-11-A2-(0)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4624-09	Percent Solids	Cool 4 deg C	RAIL01	I23	T-11-A3-(0)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4624-10	Percent Solids	Cool 4 deg C	RAIL01	I23	T-11-A4-(0)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4624-11	Percent Solids	Cool 4 deg C	RAIL01	I23	T-11-A1-(4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4624-12	Percent Solids	Cool 4 deg C	RAIL01	I23	T-11-A2-(4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4624-13	Percent Solids	Cool 4 deg C	RAIL01	I23	T-11-A3-(4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4624-14	Percent Solids	Cool 4 deg C	RAIL01	I23	T-11-A4-(4)	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4625-01	Percent Solids	Cool 4 deg C	AREC01	I33	HF-01	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4625-02	Percent Solids	Cool 4 deg C	AREC01	I33	HF-01	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4625-03	Percent Solids	Cool 4 deg C	AREC01	I33	MM-01	08/29/2019	Chemtech -SO
09/08/2019	Solid	K4625-04	Percent Solids	Cool 4 deg C	AREC01	I33	MM-01	08/29/2019	Chemtech -SO
09/05/2019	Solid	K4626-01	Percent Solids	Cool 4 deg C	LANG01	I33	1-2-EWC-1	08/29/2019	Chemtech -SO
09/05/2019	Solid	K4626-02	Percent Solids	Cool 4 deg C	LANG01	I33	1-2-EWC-2	08/29/2019	Chemtech -SO
09/05/2019	Solid	K4626-03	Percent Solids	Cool 4 deg C	LANG01	I33	1-2-EWC-3	08/29/2019	Chemtech -SO
09/05/2019	Solid	K4626-04	Percent Solids	Cool 4 deg C	LANG01	I33	1-2-OWS-1	08/29/2019	Chemtech -SO
09/05/2019	Solid	K4626-05	Percent Solids	Cool 4 deg C	LANG01	I33	1-2-SWG-1	08/29/2019	Chemtech -SO
09/05/2019	Solid	K4626-06	Percent Solids	Cool 4 deg C	LANG01	I33	1-2-EXFC-1	08/29/2019	Chemtech -SO
09/05/2019	Solid	K4626-07	Percent Solids	Cool 4 deg C	LANG01	I33	3-WC-1	08/29/2019	Chemtech -SO
09/05/2019	Solid	K4626-08	Percent Solids	Cool 4 deg C	LANG01	I33	4-WDC-1	08/29/2019	Chemtech -SO

Date/Time

08/30/19 15:35

Received by:

Jy

Relinquished by:

CS

Date/Time

08/30/19 18:00

Received by:

CS

Relinquished by:

Jy

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-083019 WorkList ID : 130417 Department : Wet-Chemistry Date : 08-30-2019 07:45:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/05/2019	Solid	K4627-01	Percent Solids	Cool 4 deg C	LANG01	I33	7-6-EXJC-1	08/29/2019	Chemtech -SO
09/05/2019	Solid	K4627-02	Percent Solids	Cool 4 deg C	LANG01	I33	6-SWG-1	08/29/2019	Chemtech -SO
09/05/2019	Solid	K4627-03	Percent Solids	Cool 4 deg C	LANG01	I33	6-EWC-1	08/29/2019	Chemtech -SO
09/05/2019	Solid	K4627-04	Percent Solids	Cool 4 deg C	LANG01	I33	7-6-EWC-2	08/29/2019	Chemtech -SO
09/05/2019	Solid	K4627-05	Percent Solids	Cool 4 deg C	LANG01	I33	6-EWG-1	08/29/2019	Chemtech -SO
09/05/2019	Solid	K4627-06	Percent Solids	Cool 4 deg C	LANG01	I33	7-6-DC-1	08/29/2019	Chemtech -SO
09/05/2019	Solid	K4627-07	Percent Solids	Cool 4 deg C	LANG01	I33	7-6-IWC-1	08/29/2019	Chemtech -SO
09/05/2019	Solid	K4627-08	Percent Solids	Cool 4 deg C	LANG01	I33	6-IWG-1	08/29/2019	Chemtech -SO
09/05/2019	Solid	K4627-09	Percent Solids	Cool 4 deg C	LANG01	I33	6-BFJC-1	08/29/2019	Chemtech -SO
09/05/2019	Solid	K4627-10	Percent Solids	Cool 4 deg C	LANG01	I33	7-6-IDC-1	08/29/2019	Chemtech -SO
09/05/2019	Solid	K4627-11	Percent Solids	Cool 4 deg C	LANG01	I33	6-IEXJ-1	08/29/2019	Chemtech -SO
09/05/2019	Solid	K4627-12	Percent Solids	Cool 4 deg C	LANG01	I33	7-6-FDS-1	08/29/2019	Chemtech -SO
09/05/2019	Solid	K4627-13	Percent Solids	Cool 4 deg C	LANG01	I33	7-6-FSC-1	08/29/2019	Chemtech -SO
08/31/2019	Solid	K4629-01	Percent Solids	Cool 4 deg C	GARD04	I41	390821755.1	08/21/2019	Chemtech -SO
09/02/2019	Solid	K4637-05	Percent Solids	Cool 4 deg C	CHEM02	I41	SVOC-GPC-BLANK	08/23/2019	Chemtech -SO
09/02/2019	Solid	K4637-06	Percent Solids	Cool 4 deg C	CHEM02	I41	PEST-GPC-BLANK	08/23/2019	Chemtech -SO
09/02/2019	Solid	K4637-07	Percent Solids	Cool 4 deg C	CHEM02	I41	PEST-GPC-BLANK-SPIKE	08/23/2019	Chemtech -SO
09/02/2019	Solid	K4637-08	Percent Solids	Cool 4 deg C	CHEM02	I41	PCB-GPC-BLANK	08/23/2019	Chemtech -SO
09/02/2019	Solid	K4637-09	Percent Solids	Cool 4 deg C	CHEM02	I41	PCB-GPC-BLANK-SPIKE	08/23/2019	Chemtech -SO
09/02/2019	Solid	K4637-10	Percent Solids	Cool 4 deg C	CHEM02	I41	SVOC-GPC2-BLANK	08/23/2019	Chemtech -SO
09/02/2019	Solid	K4637-11	Percent Solids	Cool 4 deg C	CHEM02	I41	PEST-GPC2-BLANK	08/23/2019	Chemtech -SO
09/02/2019	Solid	K4637-12	Percent Solids	Cool 4 deg C	CHEM02	I41	PEST-GPC2-BLANK-SPIKE	08/23/2019	Chemtech -SO

Date/Time 08/30/19 15:35
 Received by: Jg
 Relinquished by: Jg

Date/Time 08/30/19 18:00
 Received by: ce
 Relinquished by: Jg

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-083019

WorkList ID : 130417

Department : Wet-Chemistry

Date : 08-30-2019 07:45:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/02/2019	Solid	K4637-13	Percent Solids	Cool 4 deg C	CHEM02	I41	PCB-GPC2-BLANK	08/23/2019	Chemtech -SO
09/02/2019	Solid	K4637-14	Percent Solids	Cool 4 deg C	CHEM02	I41	PCB-GPC2-BLANK -SPIKE	08/23/2019	Chemtech -SO
09/02/2019	Solid	K4645-01	Percent Solids	Cool 4 deg C	JACO05	I41	S700A1-N-0.0-0.5-083019	08/30/2019	Chemtech -SO
09/02/2019	Solid	K4645-11	Percent Solids	Cool 4 deg C	JACO05	I41	S700B1-N-0.0-0.5-083019	08/30/2019	Chemtech -SO
09/02/2019	Solid	K4646-01	Percent Solids	Cool 4 deg C	JACO05	I23	S602A-M-1.0-1.5-083019	08/30/2019	Chemtech -SO
09/02/2019	Solid	K4646-14	Percent Solids	Cool 4 deg C	JACO05	I23	EB-7-08302019	08/30/2019	Chemtech -SO
08/27/2019	Solid	K4647-01	Percent Solids	Cool 4 deg C	JACO05	I23	S401-K1-1.0-1.5-082619-FD	08/26/2019	Chemtech -SO
08/27/2019	Solid	K4647-02	Percent Solids	Cool 4 deg C	JACO05	I23	S400-K1-1.0-1.5-082619	08/26/2019	Chemtech -SO
08/26/2019	Solid	K4647-03	Percent Solids	Cool 4 deg C	JACO05	I23	S602-M-1.0-1.5-082319	08/23/2019	Chemtech -SO
08/27/2019	Solid	K4647-04	Percent Solids	Cool 4 deg C	JACO05	I23	S402-K1-1.0-1.5-082619	08/26/2019	Chemtech -SO
08/27/2019	Solid	K4647-05	Percent Solids	Cool 4 deg C	JACO05	I23	S401-K1-1.0-1.5-082619	08/26/2019	Chemtech -SO

Date/Time 08/30/19 15:35
 Received by: JY
 Relinquished by: CB

Date/Time 08/30/19 18:00
 Received by: CB
 Relinquished by: JY