

LB104889


ISM02.4

INSTRUMENT ID : CVI

Linear

 μ Abs.:

48522



A= 0.0000e+000

B= 2.0437e-004

C= 2.6081e-002

Rho= 0.9997658

Accept=Accepted

slope
y intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/D
0.00	0.000	0.009	0.009	-86	0.000	-86					—
0.05	0.050										
0.20	0.200	0.239	0.039	1041	0.0 %	1041					20
2.50	2.500	2.458	-0.042	11899	0.0 %	11899					-2
5.00	5.000	4.903	-0.097	23864	0.0 %	23864					-2
7.50	7.500	7.649	0.149	37300	0.0 %	37300					2
10.0	10.000	9.943	-0.057	48522	0.0 %	48522					-1

LB104889

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	-86	-		0 ISM02.4	PPB	8/30/2019 10:51	Std
	0.2 S0.2	1041	-		0.2 ISM02.4	PPB	8/30/2019 10:54	Std
	2.5 S2.5	11899	-		2.5 ISM02.4	PPB	8/30/2019 10:56	Std
	5 S5.0	23864	-		5 ISM02.4	PPB	8/30/2019 10:58	Std
	7.5 S7.5	37300	-		7.5 ISM02.4	PPB	8/30/2019 11:01	Std
	10 S10.0	48522	-		10 ISM02.4	PPB	8/30/2019 11:06	Std
ICV	ICV	20481	4.2118 -		ISM02.4	PPB	8/30/2019 11:09	SMPL
ICB	ICB	-38	0.0183 -		ISM02.4	PPB	8/30/2019 11:12	SMPL
CCV034	CCV034	23556	4.8402 -		ISM02.4	PPB	8/30/2019 11:14	SMPL
CCB034	CCB034	-27	0.0206 -		ISM02.4	PPB	8/30/2019 11:16	SMPL
PB122708BL	PBS708	18	0.0298 -		ISM02.4	PPB	8/30/2019 11:18	SMPL
K4374-19	MX2H73	41693	8.5469 -		ISM02.4	PPB	8/30/2019 11:29	SMPL
K4374-20	MX2H73D	42050	8.6198 -		ISM02.4	PPB	8/30/2019 11:31	SMPL
K4374-21	MX2H73S	60204	12.33 -		ISM02.4	PPB	8/30/2019 11:33	SMPL
K4374-19DLX2		20073	4.1284 -		ISM02.4	PPB	8/30/2019 11:35	SMPL
K4374-20DLX2		20708	4.2582 -		ISM02.4	PPB	8/30/2019 11:38	SMPL
K4374-21DLX2		30305	6.2195 -		ISM02.4	PPB	8/30/2019 11:40	SMPL
CCV035	CCV035	24116	4.9547 -		ISM02.4	PPB	8/30/2019 11:42	SMPL
CCB035	CCB035	-38	0.0183 -		ISM02.4	PPB	8/30/2019 11:45	SMPL