



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB104904

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	08/30/2019	13:55
2	K4613-10		1	8	100.00	>212.0	>212.0	08/30/2019	14:25
3	K4613-10DUP		1	8	100.00	>212.0	>212.0	08/30/2019	14:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 104904

WorkList Name : FP83019 WorkList ID : 130434 Department : Wet-Chemistry Date : 08-30-2019 08:43:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/05/2019	Water	K4613-10	Flash Point	Cool 4 deg C	PSEG01	J22	TP-11-GW	08/29/2019	1010A

Date/Time 08/30/19 13:30
 Received by: KS
 Relinquished by: JP

Date/Time 08/30/19 14:53
 Received by: JP
 Relinquished by: KS