

LB104912


7471B

INSTRUMENT ID: CUI

Linear

 μ Abs.:

59568



A= 0.0000e+000

B= 1.6815e-004

C= 2.7858e-002

Rho= 0.9999218

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.029	0.029	5	0.000	5				
0.2	0.200	0.248	0.048	1308	0.0 %	1308				
2.5	2.500	2.417	-0.083	14209	0.0 %	14209				
5.0	5.000	4.989	-0.011	29507	0.0 %	29507				
7.5	7.500	7.473	-0.027	44276	0.0 %	44276				
10.0	10.000	10.044	0.044	59568	0.0 %	59568				

LB104912 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 31 Aug 2019 09:13:54				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	500	5	-	0.0000	7471B	PPB	31 Aug 2019 09:26:18	S Std
0.2 - 1	50-2	1308	-	0.2000	7471B	PPB	31 Aug 2019 09:28:36	S Std
2.5 - 1	52-5	14209	-	2.5000	7471B	PPB	31 Aug 2019 09:30:53	S Std
5.0 - 1	55-0	29507	-	5.0000	7471B	PPB	31 Aug 2019 09:33:11	S Std
7.5 - 1	57-5	44276	-	7.5000	7471B	PPB	31 Aug 2019 09:35:28	S Std
10.0 - 1	510-0	59568	-	10.0000	7471B	PPB	31 Aug 2019 09:37:45	S Std
ICV97 - 1	ICV97	24746	4.1889	-	7471B	PPB	31 Aug 2019 09:43:57	U SMPL
ICB97 - 1	ICB97	-30	0.0228	-	7471B	PPB	31 Aug 2019 09:46:12	U SMPL
CCV51 - 1	CCV51	29688	5.0199	-	7471B	PPB	31 Aug 2019 09:48:28	U SMPL
CCB51 - 1	CCB51	-47	0.0200	-	7471B	PPB	31 Aug 2019 09:50:44	U SMPL
CRA - 1	CRA	1348	0.2545	-	7471B	PPB	31 Aug 2019 09:53:00	U SMPL
HighStd - 1	HighStd	59723	10.0702	-	7471B	PPB	31 Aug 2019 09:55:16	U SMPL
ChkStd - 1	ChkStd	41773	7.0519	-	7471B	PPB	31 Aug 2019 09:57:32	U SMPL
PB122734BL - 1	PBS	-49	0.0196	-	7471B	PPB	31 Aug 2019 09:59:50	U SMPL
K4618-01 - 1	SB-4-12-14	42533	7.1797	-	7471B	PPB	31 Aug 2019 10:04:25	U SMPL
K4618-01DUP - 1	SB-4-12-14DUP	43393	7.3243	-	7471B	PPB	31 Aug 2019 10:06:42	U SMPL
K4618-01MS - 1	SB-4-12-14MS	64230	10.8281	-	7471B	PPB	31 Aug 2019 10:09:00	U SMPL
K4618-01MSD - 1	SB-4-12-14MSD	61997	10.4526	-	7471B	PPB	31 Aug 2019 10:16:06	U SMPL
K4618-03 - 1	SB-4-14-16	300064	HIGH	-	7471B	PPB	31 Aug 2019 10:18:23	X
CCV52 - 1	CCV52	30406	5.1406	-	7471B	PPB	31 Aug 2019 10:20:20	U SMPL
CCB52 - 1	CCB52	-72	0.0158	-	7471B	PPB	31 Aug 2019 10:22:38	U SMPL
K4618-05 - 1	SB-4-18-20	196	0.0608	-	7471B	PPB	31 Aug 2019 10:24:54	U SMPL
K4618-07 - 1	SB-4-20-22	219	0.0647	-	7471B	PPB	31 Aug 2019 10:27:10	U SMPL
K4618-09 - 1	SB-4-22-24	1734	0.3194	-	7471B	PPB	31 Aug 2019 10:29:26	U SMPL
K4618-11 - 1	SB-4-26-28	1005	0.1968	-	7471B	PPB	31 Aug 2019 10:31:42	U SMPL
K4618-13 - 1	SB-4-28-30	1452	0.2720	-	7471B	PPB	31 Aug 2019 10:33:59	U SMPL
K4619-03 - 1	TP1-COMP	594	0.1277	-	7471B	PPB	31 Aug 2019 10:36:16	U SMPL
K4619-06 - 1	TP2-COMP	1310	0.2481	-	7471B	PPB	31 Aug 2019 10:38:34	U SMPL
K4619-09 - 1	TP3-COMP	872	0.1745	-	7471B	PPB	31 Aug 2019 10:40:51	U SMPL
K4619-12 - 1	TP4-COMP	1054	0.2051	-	7471B	PPB	31 Aug 2019 10:43:10	U SMPL
K4619-15 - 1	TP5-COMP	358	0.0881	-	7471B	PPB	31 Aug 2019 10:45:26	U SMPL
CCV53 - 1	CCV53	30615	5.1757	-	7471B	PPB	31 Aug 2019 10:47:42	U SMPL
CCB53 - 1	CCB53	-52	0.0191	-	7471B	PPB	31 Aug 2019 10:49:58	U SMPL
K4619-18 - 1	TP6-COMP	505	0.1128	-	7471B	PPB	31 Aug 2019 10:52:15	U SMPL
K4619-21 - 1	TP7-COMP	756	0.1550	-	7471B	PPB	31 Aug 2019 10:54:31	U SMPL
K4619-24 - 1	TP8-COMP	645	0.1363	-	7471B	PPB	31 Aug 2019 10:56:47	U SMPL
K4619-27 - 1	TP9-COMP	253	0.0704	-	7471B	PPB	31 Aug 2019 10:59:05	U SMPL
K4619-30 - 1	TP10-COMP	1185	0.2271	-	7471B	PPB	31 Aug 2019 11:01:21	U SMPL
K4619-33 - 1	TP11-COMP	1075	0.2086	-	7471B	PPB	31 Aug 2019 11:03:38	U SMPL
K4619-36 - 1	TP12-COMP	1300	0.2465	-	7471B	PPB	31 Aug 2019 11:05:55	U SMPL
PB122735BL - 1	PBS	10	0.0295	-	7471B	PPB	31 Aug 2019 11:08:13	U SMPL
PB122735BS - 1	LCSS	24155	4.0895	-	7471B	PPB	31 Aug 2019 11:10:30	U SMPL
K4089-06 - 1	OK-02-082919	4940	0.8585	-	7471B	PPB	31 Aug 2019 11:12:46	U SMPL
CCV54 - 1	CCV54	31965	5.4027	-	7471B	PPB	31 Aug 2019 11:15:04	U SMPL
CCB54 - 1	CCB54	-41	0.0210	-	7471B	PPB	31 Aug 2019 11:17:20	U SMPL
K4090-17 - 1	OR-02-082919	2388	0.4294	-	7471B	PPB	31 Aug 2019 11:19:36	U SMPL
K4601-01 - 1	T12-E1-E4-(0-2.7)	4051	0.7090	-	7471B	PPB	31 Aug 2019 11:21:52	U SMPL
K4601-10 - 1	T12-F1-F4-(0-2.7)	4401	0.7679	-	7471B	PPB	31 Aug 2019 11:24:09	U SMPL
K4601-19 - 1	T12-G1-G4-(0-2.7)	5967	1.0312	-	7471B	PPB	31 Aug 2019 11:26:25	U SMPL
K4602-08 - 1	PAR-C1-C4-(0-3.11)	1768	0.3251	-	7471B	PPB	31 Aug 2019 11:28:42	U SMPL
K4602-13 - 1	PAR-D1-D4-(0-3.11)	2686	0.4795	-	7471B	PPB	31 Aug 2019 11:30:58	U SMPL
K4603-02 - 1	PAR-E1-E4-(0-3.11)	5737	0.9925	-	7471B	PPB	31 Aug 2019 11:33:16	U SMPL
K4603-11 - 1	PAR-12-F1-F4(0-3.11)	2886	0.5131	-	7471B	PPB	31 Aug 2019 11:35:34	U SMPL
K4613-03 - 1	TP-10-S	1770	0.3255	-	7471B	PPB	31 Aug 2019 11:37:52	U SMPL
K4613-03DUP - 1	TP-10-SDUP	1734	0.3194	-	7471B	PPB	31 Aug 2019 11:40:08	U SMPL
CCV55 - 1	CCV55	30455	5.1488	-	7471B	PPB	31 Aug 2019 11:42:25	U SMPL
CCB55 - 1	CCB55	-100	0.0110	-	7471B	PPB	31 Aug 2019 11:44:41	U SMPL
K4613-03MS - 1	TP-10-SMS	24260	4.1072	-	7471B	PPB	31 Aug 2019 11:46:58	U SMPL
K4613-03MSD - 1	TP-10-SMSD	24371	4.1258	-	7471B	PPB	31 Aug 2019 11:49:15	U SMPL
K4613-07 - 1	TP-10-D	933	0.1847	-	7471B	PPB	31 Aug 2019 11:51:31	U SMPL
K4618-01LX5 - 1		7942	1.3633	-	7471B	PPB	31 Aug 2019 11:53:47	U SMPL
K4618-01A - 1		61726	10.4070	-	7471B	PPB	31 Aug 2019 12:02:52	U SMPL
K4613-03LX5 - 1		235	0.0674	-	7471B	PPB	31 Aug 2019 12:05:09	U SMPL
K4613-03A - 1		24324	4.1179	-	7471B	PPB	31 Aug 2019 12:07:28	U SMPL
PB122734BS - 1	LCSS	23062	3.9057	-	7471B	PPB	31 Aug 2019 12:09:46	U SMPL
K4618-03DLX10 - 1		40926	6.9095	-	7471B	PPB	31 Aug 2019 12:12:04	U SMPL
K4618-01DLX2 - 1		21235	3.5985	-	7471B	PPB	31 Aug 2019 12:14:21	U SMPL
CCV56 - 1	CCV56	30496	5.1557	-	7471B	PPB	31 Aug 2019 12:16:38	U SMPL
CCB56 - 1	CCB56	-57	0.0183	-	7471B	PPB	31 Aug 2019 12:18:55	U SMPL
K4618-01DUPDLX2 - 1		21277	3.6056	-	7471B	PPB	31 Aug 2019 12:21:12	U SMPL
K4618-01MSDLX2 - 1		31716	5.3609	-	7471B	PPB	31 Aug 2019 12:23:30	U SMPL
K4618-01MSDDLX2 - 1		28156	4.7623	-	7471B	PPB	31 Aug 2019 12:25:47	U SMPL
K4618-01DLX10 - 1		4189	0.7322	-	7471B	PPB	31 Aug 2019 12:28:04	U SMPL
K4618-01ADLX2 - 1		31232	5.2795	-	7471B	PPB	31 Aug 2019 12:30:21	U SMPL
CCV57 - 1	CCV57	30910	5.2253	-	7471B	PPB	31 Aug 2019 12:32:38	U SMPL
CCB57 - 1	CCB57	-39	0.0213	-	7471B	PPB	31 Aug 2019 12:34:56	U SMPL