



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB104916

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	08/31/2019	11:55
2	K4555-23		1	8	100.00	>212.0	>212.0	08/31/2019	12:15
3	K4555-23DUP		1	8	100.00	>212.0	>212.0	08/31/2019	12:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP83119 WorkList ID : 130476 Department : Wet-Chemistry Date : 08-31-2019 08:17:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/05/2019	Water	K4555-23	H Flash Point	Cool 4 deg C	AKRF02	I31	TW-10-DEP	08/29/2019	1010A

Date/Time 08/31/19 11:30
 Received by: KS
 Relinquished by: CJ

Date/Time 08/31/19 12:50
 Received by: CJ
 Relinquished by: KS

LB 104916