



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB104923
Analysis Date: 09/03/2019
BalanceID: WC SC-3
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: apatel
Extraction Date: 09/03/2019
Extraction IN Time: 07:25
Extraction OUT Time: 08:00

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB104923BL	LB104923BL	WATER	2	1000	100	2.4744	2.4744	3.02	2.4744	2.4744	0.0000	0
2	LB104923BS	LB104923BS	WATER	2	1000	100	2.8957	2.8957	3.01	2.9147	2.9147	0.0190	19
3	LB104923BSD	LB104923BSD	WATER	2	1000	100	2.1023	2.1023	3.03	2.1215	2.1215	0.0192	19.2
4	K4610-01	GRAB	WATER	2	1000	100	3.0441	3.0441	3.04	3.0694	3.0694	0.0253	25.3
5	K4613-09	TP-11-GW	WATER	2	1000	100	2.9580	2.9580	3.02	2.9584	2.9584	0.0004	0.4



QC Batch# LB104923 Test: TPH Analysis Date: 09/03/2019

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2566
pH Paper 0-14	M4240
Sodium Sulfate	EP1977
1:1 HCL	WP74116
Silica Gel	W2546
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP74117
LCSWD	10.00 ML	WP74118
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0019</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>09:26</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time:	<u>08:25</u>		
Bal Check Time:	<u>07:35</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>10:00</u>
		Out Time:	<u>09:25</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>11:15</u>
1.0000 gram Balance:	<u>1.0004</u>	In Time:	<u>10:30</u>		
Bal Check Time:	<u>12:15</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>11:40</u>
		Out Time:	<u>10:32</u>		

2B104923

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH K4610 WorkList ID : 130490 Department : Wet-Chemistry Date : 09-02-2019 09:53:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/04/2019	Water	K4610-01 ✓	TPH	Conc H2SO4 4c	ARAM01	I31	GRAB	08/28/2019	1664A
09/05/2019	Water	K4613-09 -12	TPH	Conc H2SO4 4c	PSEG01	J22	TP-11-GW	08/29/2019	1664A

Date/Time 09.09.2019 17:20
 Received by: Jg
 Relinquished by: cd

Date/Time 09.02.19 9:42
 Received by: cd
 Relinquished by: Jg