

LB104929


7471B

INSTRUMENT ID : CUI

Linear

 μ Abs.:

60762



A= 0.0000e+000

B= 1.6385e-004

C= -1.9700e-002

Rho= 0.9999291

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.006	-0.006	84	0.000	84				
0.2	0.200	0.173	-0.027	1177	0.0 %	1177				
2.5	2.500	2.485	-0.015	15286	0.0 %	15286				
5.0	5.000	5.067	0.067	31046	0.0 %	31046				
7.5	7.500	7.545	0.045	46168	0.0 %	46168				
10.0	10.000	9.936	-0.064	60762	0.0 %	60762				

LB104929 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 03 Sep 2019 11:19:51					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	84	-	0.00007471B	PPB		03 Sep 2019 11:29:29	S	Std
0.2 - 1	50-2	1177	-	0.20007471B	PPB		03 Sep 2019 11:31:47	S	Std
2.5 - 1	52-5	15286	-	2.50007471B	PPB		03 Sep 2019 11:34:05	S	Std
5.0 - 1	55-0	31046	-	5.00007471B	PPB		03 Sep 2019 11:36:22	S	Std
7.5 - 1	57-5	46168	-	7.50007471B	PPB		03 Sep 2019 11:38:40	S	Std
10.0 - 1	510-0	60762	-	10.00007471B	PPB		03 Sep 2019 11:40:57	S	Std
ICV98 - 1	ICV98	24397	3.9777	-7471B	PPB		03 Sep 2019 11:44:46	U	SMPL
ICB98 - 1	ICB98	65	-0.0090	-7471B	PPB		03 Sep 2019 11:47:02	U	SMPL
CCV58 - 1	CCV58	31449	5.1331	-7471B	PPB		03 Sep 2019 11:49:18	U	SMPL
CCB58 - 1	CCB58	28	-0.0151	-7471B	PPB		03 Sep 2019 11:51:34	U	SMPL
CRA - 1	CRA	1182	0.1740	-7471B	PPB		03 Sep 2019 11:53:50	U	SMPL
HighStd - 1	HighStd	61239	10.0142	-7471B	PPB		03 Sep 2019 11:56:06	U	SMPL
ChkStd - 1	ChkStd	43449	7.0993	-7471B	PPB		03 Sep 2019 11:58:24	U	SMPL
PB122768BL - 1	PB122768BL	59	-0.0100	-7471B	PPB		03 Sep 2019 12:01:13	U	SMPL
PB122768BS - 1	PB122768BS	20482	3.3362	-7471B	PPB		03 Sep 2019 12:04:06	U	SMPL
K4554-01 - 1	SB-13-(0-2)	843	0.1184	-7471B	PPB		03 Sep 2019 12:06:22	U	SMPL
K4554-01DUP - 1	SB-13-(0-2)DUP	798	0.1111	-7471B	PPB		03 Sep 2019 12:08:41	U	SMPL
K4554-01MS - 1	SB-13-(0-2)MS	22900	3.7324	-7471B	PPB		03 Sep 2019 12:10:58	U	SMPL
K4554-01MSD - 1	SB-13-(0-2)MSD	23652	3.8556	-7471B	PPB		03 Sep 2019 12:13:15	U	SMPL
K4554-02 - 1	SB-13-(13-15)	441	0.0526	-7471B	PPB		03 Sep 2019 12:15:31	U	SMPL
CCV59 - 1	CCV59	31197	5.0918	-7471B	PPB		03 Sep 2019 12:17:48	U	SMPL
CCB59 - 1	CCB59	51	-0.0113	-7471B	PPB		03 Sep 2019 12:20:03	U	SMPL
K4554-03 - 1	SB-14-(0-2)	2071	0.3196	-7471B	PPB		03 Sep 2019 12:22:19	U	SMPL
K4554-04 - 1	SB-14-(13-15)	5139	0.8223	-7471B	PPB		03 Sep 2019 12:24:35	U	SMPL
K4554-05 - 1	SB-12-(0-2)	389	0.0440	-7471B	PPB		03 Sep 2019 12:26:51	U	SMPL
K4554-06 - 1	SB-12-(13-15)	397	0.0453	-7471B	PPB		03 Sep 2019 12:29:07	U	SMPL
K4554-07 - 1	SB-15-(0-2)	469	0.0571	-7471B	PPB		03 Sep 2019 12:31:24	U	SMPL
K4554-08 - 1	SB-15-(13-15)	330	0.0344	-7471B	PPB		03 Sep 2019 12:33:40	U	SMPL
K4554-19 - 1	SB-10-(0-2)	1280	0.1900	-7471B	PPB		03 Sep 2019 12:35:58	U	SMPL
K4554-20 - 1	SB-10-(12-14)	13421	2.1793	-7471B	PPB		03 Sep 2019 12:38:15	U	SMPL
K4554-21 - 1	SB-9-(0-2)	6433	1.0343	-7471B	PPB		03 Sep 2019 12:40:33	U	SMPL
K4554-22 - 1	SB-9-(7-9)	1128	0.1651	-7471B	PPB		03 Sep 2019 12:42:49	U	SMPL
CCV60 - 1	CCV60	31658	5.1674	-7471B	PPB		03 Sep 2019 12:45:05	U	SMPL
CCB60 - 1	CCB60	12	-0.0177	-7471B	PPB		03 Sep 2019 12:47:21	U	SMPL
K4554-23 - 1	SB-11-(0-2)	1520	0.2293	-7471B	PPB		03 Sep 2019 12:49:36	U	SMPL
K4554-24 - 1	SB-11-(13-15)	273	0.0250	-7471B	PPB		03 Sep 2019 12:51:53	U	SMPL
K4554-25 - 1	SB-7-(0-2)	131163	21.4710	-7471B	PPB		03 Sep 2019 12:54:08	U	SMPL
K4554-26 - 1	SB-7-(13-15)	-14	-0.0220	-7471B	PPB		03 Sep 2019 12:56:26	U	SMPL
K4554-27 - 1	SB-6-(0-2)	512	0.0642	-7471B	PPB		03 Sep 2019 12:58:45	U	SMPL
K4554-28 - 1	SB-6-(6-8)	441	0.0526	-7471B	PPB		03 Sep 2019 13:01:01	U	SMPL
K4554-39 - 1	SB-8-(0-2)	13456	2.1850	-7471B	PPB		03 Sep 2019 13:03:20	U	SMPL
K4554-40 - 1	SB-8-(5-7)	816	0.1140	-7471B	PPB		03 Sep 2019 13:05:39	U	SMPL
PB122769BL - 1	PB122769BL	52	-0.0112	-7471B	PPB		03 Sep 2019 13:07:57	U	SMPL
PB122769BS - 1	PB122769BS	20682	3.3690	-7471B	PPB		03 Sep 2019 13:10:13	U	SMPL
CCV61 - 1	CCV61	31549	5.1495	-7471B	PPB		03 Sep 2019 13:12:29	U	SMPL
CCB61 - 1	CCB61	-15	-0.0222	-7471B	PPB		03 Sep 2019 13:14:46	U	SMPL
K4554-41 - 1	SB-3-(0-2)	11543	1.8716	-7471B	PPB		03 Sep 2019 13:17:02	U	SMPL
K4554-41DUP - 1	SB-3-(0-2)DUP	11678	1.8937	-7471B	PPB		03 Sep 2019 13:19:18	U	SMPL
K4554-41MS - 1	SB-3-(0-2)MS	31022	5.0632	-7471B	PPB		03 Sep 2019 13:21:35	U	SMPL
K4554-41MSD - 1	SB-3-(0-2)MSD	33464	5.4633	-7471B	PPB		03 Sep 2019 13:23:52	U	SMPL
K4554-42 - 1	SB-3-(8-10)	5333	0.8541	-7471B	PPB		03 Sep 2019 13:26:09	U	SMPL
K4554-43 - 1	SB-4-(0-2)	116361	19.0458	-7471B	PPB		03 Sep 2019 13:28:26	U	SMPL
K4554-44 - 1	SB-4-(7-9)	422	0.0494	-7471B	PPB		03 Sep 2019 13:30:43	U	SMPL
K4554-45 - 1	SB-2-(0-2)	14004	2.2748	-7471B	PPB		03 Sep 2019 13:33:03	U	SMPL
K4554-46 - 1	SB-2-(13-15)	767	0.1060	-7471B	PPB		03 Sep 2019 13:35:21	U	SMPL
K4554-55 - 1	SB-5-(0-2)	35615	5.8157	-7471B	PPB		03 Sep 2019 13:37:38	U	SMPL
CCV62 - 1	CCV62	31270	5.1038	-7471B	PPB		03 Sep 2019 13:39:54	U	SMPL
CCB62 - 1	CCB62	2	-0.0194	-7471B	PPB		03 Sep 2019 13:42:12	U	SMPL
K4554-56 - 1	SB-5-(8-10)	528	0.0668	-7471B	PPB		03 Sep 2019 13:44:28	U	SMPL
K4554-57 - 1	SB-1-(0-2)	3806	0.6039	-7471B	PPB		03 Sep 2019 13:46:45	U	SMPL
K4554-58 - 1	SB-1-(13-15)	450	0.0540	-7471B	PPB		03 Sep 2019 13:49:02	U	SMPL
K4622-01 - 1	PAR-11-A1-A4-(0-4)	4497	0.7171	-7471B	PPB		03 Sep 2019 13:51:19	U	SMPL
K4622-10 - 1	PAR-11-B1-B4-(0-4)	1422	0.2133	-7471B	PPB		03 Sep 2019 13:53:35	U	SMPL
K4622-19 - 1	PAR-11-C1-C4-(0-4)	13698	2.2247	-7471B	PPB		03 Sep 2019 14:06:04	U	SMPL
K4623-08 - 1	PAR-11-D1-D4-(0-4)	2849	0.4471	-7471B	PPB		03 Sep 2019 14:08:22	U	SMPL
K4623-17 - 1	PAR-11-E1-E4-(0-4)	3862	0.6131	-7471B	PPB		03 Sep 2019 14:50:40	U	SMPL
K4624-06 - 1	T-11-A1-A4-(0-4)	4847	0.7745	-7471B	PPB		03 Sep 2019 14:52:58	U	SMPL
K4625-02 - 1	HF-01	32108	5.2411	-7471B	PPB		03 Sep 2019 14:55:14	U	SMPL
CCV63 - 1	CCV63	31751	5.1826	-7471B	PPB		03 Sep 2019 14:57:31	U	SMPL
CCB63 - 1	CCB63	96	-0.0040	-7471B	PPB		03 Sep 2019 14:59:48	U	SMPL
K4625-04 - 1	MM-01	789	0.1096	-7471B	PPB		03 Sep 2019 15:02:05	U	SMPL
K4554-01LX5 - 1		142	0.0036	-7471B	PPB		03 Sep 2019 15:04:22	U	SMPL
K4554-01A - 1		23680	3.8602	-7471B	PPB		03 Sep 2019 15:06:38	U	SMPL
K4554-41LX5 - 1		2260	0.3506	-7471B	PPB		03 Sep 2019 15:08:55	U	SMPL
K4554-41A - 1		31808	5.1920	-7471B	PPB		03 Sep 2019 15:11:11	U	SMPL
K4554-25DLX5 - 1		26980	4.4009	-7471B	PPB		03 Sep 2019 15:13:28	U	SMPL
K4554-43DLX5 - 1		23205	3.7824	-7471B	PPB		03 Sep 2019 15:15:46	U	SMPL
CCV64 - 1	CCV64	31891	5.2056	-7471B	PPB		03 Sep 2019 15:18:04	U	SMPL
CCB64 - 1	CCB64	-81	-0.0330	-7471B	PPB		03 Sep 2019 15:20:22	U	SMPL