



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104931

Slope : 97.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/03/2019	12:30
2	CAL2	1	Water	NA	NA	20.1	6.99	09/03/2019	12:31
3	CAL3	1	Water	NA	NA	20.1	10.02	09/03/2019	12:33
4	ICV	1	Water	NA	NA	20.1	7.00	09/03/2019	12:35
5	CCV1	1	Water	NA	NA	20.1	2.02	09/03/2019	12:36
6	K4651-02	1	Water	NA	NA	18.7	7.48	09/03/2019	12:37
7	K4651-02DUP	1	Water	NA	NA	18.6	7.49	09/03/2019	12:38
8	CCV2	1	Water	NA	NA	20.1	12.02	09/03/2019	12:40

LB104931

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4651 WorkList ID : 130520 Department : Wet-Chemistry Date : 09-03-2019 12:24:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/30/2019	Water	K4651-02-A	pH	Cool 4 deg C	TETR06	I41	TT-006-IDWGW-20190830	08/30/2019	9040C

Date/Time 09/03/19 12:25
Received by: JA
Relinquished by: JA

Date/Time 09/03/2019 13:30
Received by: CD
Relinquished by: JA