



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB104934

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/03/2019	16:25
2	CAL2	1	Water	NA	NA	20.1	7.01	09/03/2019	16:26
3	CAL3	1	Water	NA	NA	20.1	10.03	09/03/2019	16:27
4	ICV	1	Water	NA	NA	20.1	6.99	09/03/2019	16:30
5	CCV1	1	Water	NA	NA	20.1	2.02	09/03/2019	16:32
6	K4094-03	1	Water	NA	NA	19.5	6.39	09/03/2019	16:33
7	K4094-03DUP	1	Water	NA	NA	19.6	6.40	09/03/2019	16:35
8	CCV2	1	Water	NA	NA	20.1	12.02	09/03/2019	16:36

287106934

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K4094 W WorkList ID : 130532 Department : Wet-Chemistry Date : 09-03-2019 16:21:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/03/2019	Water	K4094-03- 7	Corrosivity	Cool 4 deg C	PSEG01	J31	OILY-WATER	09/03/2019	9040C

Date/Time 09-03-19 16:22
 Received by: Jn
 Relinquished by: _____

Date/Time 09-03-19 17:02
 Received by: Jn
 Relinquished by: Jn