



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB104935

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/03/2019	16:25
2	CAL2	1	Water	NA	NA	20.1	7.01	09/03/2019	16:26
3	CAL3	1	Water	NA	NA	20.1	10.03	09/03/2019	16:27
4	ICV	1	Water	NA	NA	20.1	6.99	09/03/2019	16:30
5	CCV1	1	Water	NA	NA	20.2	2.01	09/03/2019	16:40
6	K4094-01	1	Solid	20.02	20	22.1	6.36	09/03/2019	16:41
7	K4653-04	1	Solid	20.03	20	22.2	10.90	09/03/2019	16:42
8	K4653-04DUP	1	Solid	20.04	20	22.3	10.91	09/03/2019	16:43
9	K4654-01	1	Solid	20.01	20	21.6	5.10	09/03/2019	16:45
10	K4655-04	1	Solid	20.04	20	22.5	7.12	09/03/2019	16:47
11	K4655-08	1	Solid	20.02	20	22.2	6.51	09/03/2019	16:48
12	K4668-03	1	Solid	20.01	20	22.5	8.57	09/03/2019	16:50
13	CCV2	1	Water	NA	NA	20.2	12.02	09/03/2019	17:00

LD 104935

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K4094 S WorkList ID : 130531 Department : Wet-Chemistry Date : 09-03-2019 16:20:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/03/2019	Solid	K4094-01	Corrosivity	Cool 4 deg C	PSEG01	J31	OILY-DEBRIS	09/03/2019	9045D
09/03/2019	Solid	K4653-04	Corrosivity	Cool 4 deg C	PSEG01	J31	287	09/03/2019	9045D
09/03/2019	Solid	K4654-01	Corrosivity	Cool 4 deg C	PSEG01	J31	TR-OILY-DEBRIS	09/03/2019	9045D
09/03/2019	Solid	K4655-04	Corrosivity	Cool 4 deg C	PSEG01	J31	TP-15-S	09/03/2019	9045D
09/03/2019	Solid	K4655-08	Corrosivity	Cool 4 deg C	PSEG01	J31	TP-15-D	09/03/2019	9045D
09/03/2019	Solid	K4668-03	Corrosivity	Cool 4 deg C	PSEG05	J31	SU-04-090319	09/03/2019	9045D

Date/Time 09.03.19 16.25
 Received by: Jg
 Relinquished by: Jg

Date/Time 09-03-19 18.00
 Received by: Jg
 Relinquished by: Jg