

7470A


2B104944

INSTRUMENT ID : CU1

Linear

 μ Abs.:

41883



A= 0.0000e+000

B= 2.4041e-004

C= 1.8250e-004

Rho= 0.9994233

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.021	0.021	88	0.000	88				
0.2	0.200	0.235	0.035	976	0.0 %	976				
2.5	2.500	2.607	0.107	10844	0.0 %	10844				
5.0	5.000	4.727	-0.273	19661	0.0 %	19661				
7.5	7.500	7.541	0.041	31365	0.0 %	31365				
10.0	10.000	10.069	0.069	41883	0.0 %	41883				

LB104944 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 04 Sep 2019 10:59:14				
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50-0	88	-	0.0000	7470A	PPB	04 Sep 2019 11:10:48	S Std
0.2 - 1	50-2	976	-	0.2000	7470A	PPB	04 Sep 2019 11:12:57	S Std
2.5 - 1	52-5	10844	-	2.5000	7470A	PPB	04 Sep 2019 11:15:06	S Std
5.0 - 1	55-0	19661	-	5.0000	7470A	PPB	04 Sep 2019 11:17:14	S Std
7.5 - 1	57-0	31365	-	7.5000	7470A	PPB	04 Sep 2019 11:19:22	S Std
10.0 - 1	510-0	41883	-	10.0000	7470A	PPB	04 Sep 2019 11:21:30	S Std
ICV99 - 1	ICV99	17242	4.1453	-	7470A	PPB	04 Sep 2019 11:25:01	U SMPL
ICB99 - 1	ICB99	47	0.0115	-	7470A	PPB	04 Sep 2019 11:27:08	U SMPL
CCV65 - 1	CCV65	19511	4.6908	-	7470A	PPB	04 Sep 2019 11:29:15	U SMPL
CCB65 - 1	CCB65	60	0.0146	-	7470A	PPB	04 Sep 2019 11:31:21	U SMPL
CRA - 1	CRA	957	0.2303	-	7470A	PPB	04 Sep 2019 11:33:28	U SMPL
HighStd - 1	HighStd	42617	10.2457	-	7470A	PPB	04 Sep 2019 11:35:34	U SMPL
ChkStd - 1	ChkStd	28508	6.8537	-	7470A	PPB	04 Sep 2019 11:37:41	U SMPL
PB122803BL - 1	PBW	17	0.0043	-	7470A	PPB	04 Sep 2019 11:39:48	U SMPL
PB122803BS - 1	LCSW	15774	3.7924	-	7470A	PPB	04 Sep 2019 11:41:56	U SMPL
K4622-21 - 1	PAR-11-A1-A4-(0-4)	56	0.0136	-	7470A	PPB	04 Sep 2019 11:44:03	U SMPL
K4622-21DUP - 1	PAR-11-A1-A4-(0-4)DUP	93	0.0225	-	7470A	PPB	04 Sep 2019 11:46:11	U SMPL
K4622-21MS - 1	PAR-11-A1-A4-(0-4)MS	17068	4.1035	-	7470A	PPB	04 Sep 2019 11:48:20	U SMPL
K4622-21MSD - 1	PAR-11-A1-A4-(0-4)MSD	17269	4.1518	-	7470A	PPB	04 Sep 2019 11:50:28	U SMPL
K4622-30 - 1	PAR-11-B1-B4-(0-4)	179	0.0432	-	7470A	PPB	04 Sep 2019 11:52:35	U SMPL
CCV66 - 1	CCV66	19934	4.7925	-	7470A	PPB	04 Sep 2019 11:54:43	U SMPL
CCB66 - 1	CCB66	60	0.0146	-	7470A	PPB	04 Sep 2019 11:56:50	U SMPL
K4622-39 - 1	PAR-11-C1-C4-(0-4)	213	0.0514	-	7470A	PPB	04 Sep 2019 11:58:57	U SMPL
K4623-28 - 1	PAR-11-D1-D4-(0-4)	140	0.0338	-	7470A	PPB	04 Sep 2019 12:01:04	U SMPL
K4623-37 - 1	PAR-11-E1-E4-(0-4)	61	0.0148	-	7470A	PPB	04 Sep 2019 12:03:11	U SMPL
K4624-20 - 1	T-11-A1-A4-(0-4)	85	0.0206	-	7470A	PPB	04 Sep 2019 12:05:18	U SMPL
K4625-06 - 1	HF-01	106	0.0257	-	7470A	PPB	04 Sep 2019 12:07:26	U SMPL
K4625-08 - 1	MM-01	204	0.0492	-	7470A	PPB	04 Sep 2019 12:09:34	U SMPL
K4643-21 - 1	T11-B1-B4-(0-4)	225	0.0543	-	7470A	PPB	04 Sep 2019 12:11:42	U SMPL
K4643-30 - 1	T11-C1-C4-(0-4)	197	0.0475	-	7470A	PPB	04 Sep 2019 12:13:50	U SMPL
K4643-39 - 1	T11-D1-D4-(0-4)	146	0.0353	-	7470A	PPB	04 Sep 2019 12:15:59	U SMPL
K4644-24 - 1	T11-E1-E4-(0-4)	113	0.0273	-	7470A	PPB	04 Sep 2019 12:18:06	U SMPL
CCV67 - 1	CCV67	19778	4.7550	-	7470A	PPB	04 Sep 2019 12:20:13	U SMPL
CCB67 - 1	CCB67	27	0.0067	-	7470A	PPB	04 Sep 2019 12:22:20	U SMPL
PB122804BL - 1	PBW	99	0.0240	-	7470A	PPB	04 Sep 2019 12:24:26	U SMPL
PB122804BS - 1	LCSW	16054	3.8597	-	7470A	PPB	04 Sep 2019 12:26:34	U SMPL
K4613-10 - 1	TP-11-GW	40	0.0098	-	7470A	PPB	04 Sep 2019 12:28:41	U SMPL
K4613-10DUP - 1	TP-11-GWDUP	82	0.0199	-	7470A	PPB	04 Sep 2019 12:30:48	U SMPL
K4613-10MS - 1	TP-11-GWMS	17064	4.1025	-	7470A	PPB	04 Sep 2019 12:32:56	U SMPL
K4613-10MSD - 1	TP-11-GWMSD	17076	4.1054	-	7470A	PPB	04 Sep 2019 12:35:05	U SMPL
PB122805BL - 1	PBW	59	0.0144	-	7470A	PPB	04 Sep 2019 12:37:13	U SMPL
PB122805BS - 1	LCSW	15787	3.7955	-	7470A	PPB	04 Sep 2019 12:39:22	U SMPL
K4555-24 - 1	TW-2	29	0.0072	-	7470A	PPB	04 Sep 2019 12:41:31	U SMPL
K4555-25 - 1	TW-12	101	0.0245	-	7470A	PPB	04 Sep 2019 12:43:37	U SMPL
CCV68 - 1	CCV68	19570	4.7050	-	7470A	PPB	04 Sep 2019 12:45:44	U SMPL
CCB68 - 1	CCB68	31	0.0076	-	7470A	PPB	04 Sep 2019 12:47:52	U SMPL
K4651-02 - 1	TT-006-IDWGW-20190830	416	0.1002	-	7470A	PPB	04 Sep 2019 12:49:59	U SMPL
K4651-02DUP - 1	TT-006-IDWGW-20190830DUP	387	0.0932	-	7470A	PPB	04 Sep 2019 12:52:06	U SMPL
K4651-02MS - 1	TT-006-IDWGW-20190830MS	20030	4.8156	-	7470A	PPB	04 Sep 2019 12:54:13	U SMPL
K4651-02MSD - 1	TT-006-IDWGW-20190830MSD	19406	4.6656	-	7470A	PPB	04 Sep 2019 12:56:20	U SMPL
PB122803TB - 1	PB122717TB	57	0.0139	-	7470A	PPB	04 Sep 2019 12:58:27	U SMPL
PB122804TB - 1	PB122749TB	108	0.0261	-	7470A	PPB	04 Sep 2019 13:00:34	U SMPL
K4622-21LX5 - 1		54	0.0132	-	7470A	PPB	04 Sep 2019 13:02:42	U SMPL
K4622-21A - 1		17049	4.0989	-	7470A	PPB	04 Sep 2019 13:04:50	U SMPL
K4613-10LX5 - 1		-5	-0.0010	-	7470A	PPB	04 Sep 2019 13:06:59	U SMPL
K4613-10A - 1		17164	4.1266	-	7470A	PPB	04 Sep 2019 13:09:06	U SMPL
CCV69 - 1	CCV69	19514	4.6915	-	7470A	PPB	04 Sep 2019 13:11:14	U SMPL
CCB69 - 1	CCB69	16	0.0040	-	7470A	PPB	04 Sep 2019 13:13:21	U SMPL
K4651-02LX5 - 1		98	0.0237	-	7470A	PPB	04 Sep 2019 13:15:28	U SMPL
K4651-02A - 1		19746	4.7473	-	7470A	PPB	04 Sep 2019 13:17:36	U SMPL
CCV70 - 1	CCV70	19351	4.6523	-	7470A	PPB	04 Sep 2019 13:19:43	U SMPL
CCB70 - 1	CCB70	33	0.0081	-	7470A	PPB	04 Sep 2019 13:21:51	U SMPL