



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB104952

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/04/2019	17:40
2	CAL2	1	Water	NA	NA	20.1	7.00	09/04/2019	17:41
3	CAL3	1	Water	NA	NA	20.1	10.04	09/04/2019	17:42
4	ICV	1	Water	NA	NA	20.1	6.99	09/04/2019	17:45
5	CCV1	1	Water	NA	NA	20.1	2.02	09/04/2019	17:46
6	K4679-09	1	Solid	20.02	20	22.4	8.56	09/04/2019	17:48
7	K4679-09DUP	1	Solid	20.03	20	22.5	8.57	09/04/2019	17:49
8	K4679-10	1	Solid	20.01	20	21.8	8.12	09/04/2019	17:50
9	K4679-11	1	Solid	20.03	20	22.8	7.48	09/04/2019	18:00
10	K4679-12	1	Solid	20.04	20	22.7	7.46	09/04/2019	18:01
11	CCV2	1	Water	NA	NA	20.1	12.02	09/04/2019	18:03

12104952

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K4680 WorkList ID : 130573 Department : Wet-Chemistry Date : 09-04-2019 16:23:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/03/2019	Solid	K4679-09- 8	Corrosivity	Cool 4 deg C	LIRO01	I33	WS-SB01-COMP	09/03/2019	9045D
09/03/2019	Solid	K4679-10- 13	Corrosivity	Cool 4 deg C	LIRO01	I33	WS-SB02-COMP	09/03/2019	9045D
09/03/2019	Solid	K4679-11- 13	Corrosivity	Cool 4 deg C	LIRO01	I33	WS-SB03-COMP	09/03/2019	9045D
09/03/2019	Solid	K4679-12- 13	Corrosivity	Cool 4 deg C	LIRO01	I33	WS-SB11-COMP	09/03/2019	9045D

Date/Time 09/04/19 16:25
Received by: Jg
Relinquished by: cd

Date/Time 09/04/19 18:00
Received by: cd
Relinquished by: Jg