



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB104956
Analysis Date: 09/05/2019
BalanceID: WC SC-3
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: apatel
Extraction Date: 09/05/2019
Extraction IN Time: 07:40
Extraction OUT Time: 08:20

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB104956BL	LB104956BL	WATER	1.3	1000	100	2.4178	2.4178	3.02	2.4178	2.4178	0.0000	0
2	LB104956BS	LB104956BS	WATER	1.3	1000	100	2.1366	2.1366	3.01	2.1553	2.1553	0.0187	18.7
3	K4691-01	DRAIN-WATER-TANK-1	WATER	1.6	1000	100	3.0336	3.0336	3.04	3.0367	3.0367	0.0031	3.1
4	K4691-01MS	DRAIN-WATER-TANK-1	WATER	1.6	1000	100	2.9960	2.9960	3.03	3.0170	3.0170	0.0210	21
5	K4691-01MSD	DRAIN-WATER-TANK-1	WATER	1.6	1000	100	2.9758	2.9758	3.05	2.9965	2.9965	0.0207	20.7



QC Batch# LB104956 Test: TPH Analysis Date: 09/05/2019

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2566
pH Paper 0-14	M4240
Sodium Sulfate	EP1977
1:1 HCL	WP74116
Silica Gel	W2546
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP74117
LCSWD	NA	NA
MS/MSD	10.00 ML	WP74118

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0019</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>09:41</u>
1.0000 gram Balance:	<u>1.0004</u>	In Time:	<u>08:40</u>		
Bal Check Time:	<u>08:00</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>10:20</u>
		Out Time:	<u>09:40</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>11:26</u>
1.0000 gram Balance:	<u>1.0003</u>	In Time:	<u>10:50</u>		
Bal Check Time:	<u>12:35</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>11:50</u>
		Out Time:	<u>11:25</u>		

YB104956

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH K4691 WorkList ID : 130577 Department : Wet-Chemistry Date : 09-05-2019 7:30:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/11/2019	Water	K4691-01	TPH	Conc H2SO4 to	MAJO01	L11	DRAIN-WATER-TANK-1	09/04/2019	1664A

C, P, E,

Date/Time 09/05/19 7:35
Received by: Jp
Relinquished by: CQ

Date/Time 09.05.19 10:30
Received by: CQ
Relinquished by: Jp