



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104960

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/04/2019	17:25
2	CAL2	1	Water	NA	NA	20.2	7.01	09/04/2019	17:26
3	CAL3	1	Water	NA	NA	20.2	10.03	09/04/2019	17:27
4	ICV	1	Water	NA	NA	20.2	6.99	09/04/2019	17:30
5	CCV1	1	Water	NA	NA	20.1	2.02	09/04/2019	17:31
6	K4684-01	1	Water	NA	NA	23.1	6.30	09/04/2019	17:33
7	K4684-03	1	Water	NA	NA	23.9	7.51	09/04/2019	17:35
8	K4684-03DUP	1	Water	NA	NA	23.8	7.52	09/04/2019	17:36
9	K4691-01	1	Water	NA	NA	15.4	7.23	09/04/2019	17:37
10	CCV2	1	Water	NA	NA	20.2	12.02	09/04/2019	17:39

VB104960

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4691 WorkList ID : 130574 Department : Wet-Chemistry Date : 09-04-2019 17:19:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/04/2019	Water	K4684-01 - 17	pH	Cool 4 deg C	PSEG01	J32	20841-20842	09/04/2019	9040C
09/04/2019	Water	K4684-03 - C	pH	Cool 4 deg C	PSEG01	J32	20769	09/04/2019	9040C
09/04/2019	Water	K4691-01 - 18	pH	Cool 4 deg C	MAJO01	L11	DRAIN-WATER-TANK-1	09/04/2019	9040C

Date/Time 09/04/19 17:120
Received by: for
Relinquished by: cb

Date/Time 09/04/19 18:00
Received by: cb
Relinquished by: jo