


LBID4963

7470A

INSTRUMENT ID : CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.069	0.069	88	0.000	88				
0.2	0.200	0.242	0.042	824	0.0 %	824				
2.5	2.500	2.553	0.053	10667	0.0 %	10667				
5.0	5.000	4.701	-0.299	19813	0.0 %	19813				
7.5	7.500	7.543	0.043	31916	0.0 %	31916				
10.0	10.000	10.092	0.092	42771	0.0 %	42771				

LB104963 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Sep 2019 10:45:59

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-0	88	-	0.0000	7470A	PPB	05 Sep 2019 10:47:47	S	Std
0.2 - 1	50-2	824	-	0.2000	7470A	PPB	05 Sep 2019 10:49:56	S	Std
2.5 - 1	52-5	10667	-	2.5000	7470A	PPB	05 Sep 2019 10:52:05	S	Std
5.0 - 1	55-0	19813	-	5.0000	7470A	PPB	05 Sep 2019 10:54:14	S	Std
7.5 - 1	57-5	31916	-	7.5000	7470A	PPB	05 Sep 2019 10:56:22	S	Std
10.0 - 1	510-0	42771	-	10.0000	7470A	PPB	05 Sep 2019 10:58:30	S	Std
ICV03 - 1	ICV03	16569	3.9391	-	7470A	PPB	05 Sep 2019 11:05:07	U	SMPL
ICB03 - 1	ICB03	80	0.0670	-	7470A	PPB	05 Sep 2019 11:07:14	U	SMPL
CCV75 - 1	CCV75	19827	4.7042	-	7470A	PPB	05 Sep 2019 11:09:21	U	SMPL
CCB75 - 1	CCB75	13	0.0513	-	7470A	PPB	05 Sep 2019 11:11:28	U	SMPL
CRA - 1	CRA	832	0.2436	-	7470A	PPB	05 Sep 2019 11:13:35	U	SMPL
HighStd - 1	HighStd	41666	9.8327	-	7470A	PPB	05 Sep 2019 11:26:13	U	SMPL
ChkStd - 1	ChkStd	29225	6.9112	-	7470A	PPB	05 Sep 2019 11:28:20	U	SMPL
PB122824BL - 1	PBW	73	0.0654	-	7470A	PPB	05 Sep 2019 11:30:27	U	SMPL
PB122824BS - 1	LCSW	16114	3.8323	-	7470A	PPB	05 Sep 2019 11:32:34	U	SMPL
K4653-04 - 1	287	29	0.0550	-	7470A	PPB	05 Sep 2019 11:34:41	U	SMPL
K4655-04 - 1	TP-15-S	72	0.0651	-	7470A	PPB	05 Sep 2019 11:36:50	U	SMPL
K4655-08 - 1	TP-15-D	35	0.0565	-	7470A	PPB	05 Sep 2019 11:38:58	U	SMPL
K4668-03 - 1	SU-04-090319	96	0.0708	-	7470A	PPB	05 Sep 2019 11:41:07	U	SMPL
K4668-03DUP - 1	SU-04-090319DUP	63	0.0630	-	7470A	PPB	05 Sep 2019 11:43:14	U	SMPL
CCV76 - 1	CCV76	19622	4.6561	-	7470A	PPB	05 Sep 2019 11:45:20	U	SMPL
CCB76 - 1	CCB76	36	0.0567	-	7470A	PPB	05 Sep 2019 11:47:27	U	SMPL
K4668-03MS - 1	SU-04-090319MS	16327	3.8823	-	7470A	PPB	05 Sep 2019 11:49:34	U	SMPL
K4668-03MSD - 1	SU-04-090319MSD	16689	3.9673	-	7470A	PPB	05 Sep 2019 11:56:20	U	SMPL
PB122824TB - 1	PB122783TB	49	0.0597	-	7470A	PPB	05 Sep 2019 12:01:13	U	SMPL
K4668-03LX5 - 1		22	0.0534	-	7470A	PPB	05 Sep 2019 12:03:20	U	SMPL
K4668-03A - 1		16156	3.8422	-	7470A	PPB	05 Sep 2019 12:05:28	U	SMPL
CCV77 - 1	CCV77	20331	4.8226	-	7470A	PPB	05 Sep 2019 12:07:36	U	SMPL
CCB77 - 1	CCB77	11	0.0508	-	7470A	PPB	05 Sep 2019 12:09:44	U	SMPL