



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB104965

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	09/05/2019	10:55
2	K4684-01		1	8	100.00	>212.0	>212.0	09/05/2019	11:25
3	K4684-01DUP		1	8	100.00	>212.0	>212.0	09/05/2019	11:50
4	K4684-03		1	9	100.00	>212.0	>212.0	09/05/2019	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP9519 WorkList ID : 130607 Department : Wet-Chemistry Date : 09-05-2019 10:25:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/11/2019	Water	K4684-01	Flash Point	Cool 4 deg C	PSEG01	J32	20841-20842	09/04/2019	1010A
09/11/2019	Water	K4684-03	Flash Point	Cool 4 deg C	PSEG01	J32	20769	09/04/2019	1010A

Date/Time 09/05/19 10:30
 Received by: KS
 Relinquished by: JP

Date/Time 09/05/19 12:25
 Received by: JP
 Relinquished by: KS