



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104978

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/05/2019	17:15
2	CAL2	1	Water	NA	NA	20.1	6.99	09/05/2019	17:16
3	CAL3	1	Water	NA	NA	20.1	10.02	09/05/2019	17:17
4	ICV	1	Water	NA	NA	20.1	7.00	09/05/2019	17:19
5	CCV1	1	Water	NA	NA	20.1	2.02	09/05/2019	17:20
6	K4700-02	1	Water	NA	NA	21.3	7.26	09/05/2019	18:00
7	K4700-02DUP	1	Water	NA	NA	21.4	7.27	09/05/2019	18:01
8	K4714-01	1	Water	NA	NA	17.2	6.00	09/05/2019	18:02
9	CCV2	1	Water	NA	NA	20.1	12.02	09/05/2019	18:05

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4714
WorkList ID : 130624
Department : Wet-Chemistry
Date : 09-05-2019 16:57:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/05/2019	Water	K4700-02	✓ pH	Cool 4 deg C	DEWB01	J33	SINK	09/05/2019	9040C
09/05/2019	Water	K4714-01	✓ pH	Cool 4 deg C	MAJO01	L23	DRAIN-WATER-TANK-1	09/05/2019	9040C

Date/Time 06/05/19 17:10

Received by: A

Relinquished by: [Signature]

Date/Time 09/05/19 18.00
Received by: *CS*
Relinquished by: *JS*