



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB104990

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	09/06/2019	11:10
2	CAL2	1	Water	NA	NA	20.1	7.00	09/06/2019	11:11
3	CAL3	1	Water	NA	NA	20.1	10.03	09/06/2019	11:12
4	ICV	1	Water	NA	NA	20.1	6.99	09/06/2019	11:13
5	CCV1	1	Water	NA	NA	20.1	2.02	09/06/2019	11:15
6	K4723-01	1	Water	NA	NA	19.6	7.24	09/06/2019	11:16
7	K4723-01DUP	1	Water	NA	NA	19.7	7.25	09/06/2019	11:17
8	CCV2	1	Water	NA	NA	20.2	12.02	09/06/2019	11:20

22104990

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4723 WorkList ID : 130656 Department : Wet-Chemistry Date : 09-06-2019 10:59:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/06/2019	Water	K4723-01- E	pH	Cool 4 deg C	UNIO02	L11	FRAC-A32780T	09/06/2019	9040C

Date/Time 09.06.19 11:11:00
 Received by: JS
 Relinquished by: CB

Date/Time 09.06.19 11:40
 Received by: CB
 Relinquished by: JS