



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 9/24/2019

OVENTEMP IN Celsius(°C): 108
Time IN: 14:15
In Date: 09/06/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:42
Out Date: 09/07/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB104995

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K4718-01	C0AC4	1	1.14	9.92	8.71	86.2	
K4718-02	C0AC5	2	1.13	9.73	8.69	87.9	
K4718-03	C0AC6	3	1.14	9.69	8.36	84.4	
K4718-04	C0AC7	4	1.13	9.64	8.57	87.4	
K4718-05	C0AC8	5	1.15	9.98	8.7	85.5	
K4718-06	C0AC9	6	1.13	9.89	7.71	75.1	
K4718-07	C0AD0	7	1.14	9.7	8.65	87.7	
K4718-08	C0AD2	8	1.14	9.8	8.38	83.6	
K4718-09	C0AD3	9	1.14	9.69	8.2	82.6	
K4718-10	C0AD4	10	1.12	9.88	8.25	81.4	
K4718-11	C0AD6	11	1.15	9.78	2.59	16.7	
K4718-12	C0AD7	12	1.14	9.96	8.41	82.4	
K4718-13	C0AD8	13	1.13	9.64	8.4	85.4	
K4718-14	C0AD9	14	1.12	9.57	7.83	79.4	
K4718-15	C0AE0	15	1.14	9.8	8.21	81.6	
K4718-16	C0AE0MS	16	1.14	9.8	8.21	81.6	
K4718-17	C0AE0MSD	17	1.14	9.8	8.21	81.6	
K4718-18	C0AE1	18	1.12	9.9	7.98	78.1	
K4718-19	C0AE2	19	1.12	9.73	8.26	82.9	
K4718-20	C0AE3	20	1.12	9.79	8.47	84.8	
K4718-21	C0AE4	21	1.13	9.68	8.2	82.7	
K4718-22	C0AE5	22	1.13	9.63	8.36	85.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$