


LB104999

7470A

INSTRUMENT ID : CU1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	1	0.000	1				
0.2	0.200	0.169	-0.031	712	0.0 %	712				
2.5	2.500	2.558	0.058	10943	0.0 %	10943				
5.0	5.000	5.134	0.134	21978	0.0 %	21978				
7.5	7.500	7.213	-0.287	30885	0.0 %	30885				
10.0	10.000	10.122	0.122	43344	0.0 %	43344				

LB104999 INSTRUMENT ID : CV1

Method: 7470A	Operator: Admin	Date of Analysis: 06 Sep 2019 14:21:53							
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD-0	1	-	0.0000	7470A	PPB	06 Sep 2019 15:03:15	S	Std
0.2 - 1	SD-2	712	-	0.2000	7470A	PPB	06 Sep 2019 15:05:24	S	Std
2.5 - 1	SD-5	10943	-	2.5000	7470A	PPB	06 Sep 2019 15:07:33	S	Std
5.0 - 1	SD-10	21978	-	5.0000	7470A	PPB	06 Sep 2019 15:09:41	S	Std
7.5 - 1	SD-15	30885	-	7.5000	7470A	PPB	06 Sep 2019 15:11:49	S	Std
10.0 - 1	SD-20	43344	-	10.0000	7470A	PPB	06 Sep 2019 15:17:11	S	Std
ICV05 - 1	ICV05	15837	3.7004	-	7470A	PPB	06 Sep 2019 15:20:35	U	SMPL
ICB05 - 1	ICB05	51	0.0150	-	7470A	PPB	06 Sep 2019 15:22:41	U	SMPL
CCV85 - 1	CCV85	21560	5.0364	-	7470A	PPB	06 Sep 2019 15:24:48	U	SMPL
CCB85 - 1	CCB85	58	0.0166	-	7470A	PPB	06 Sep 2019 15:26:55	U	SMPL
CRA - 1	CRA	694	0.1651	-	7470A	PPB	06 Sep 2019 15:29:01	U	SMPL
HighStd - 1	HighStd	43288	10.1091	-	7470A	PPB	06 Sep 2019 15:31:07	U	SMPL
ChkStd - 1	ChkStd	30098	7.0297	-	7470A	PPB	06 Sep 2019 15:33:14	U	SMPL
PB122870BL - 1	PBW	7	0.0047	-	7470A	PPB	06 Sep 2019 15:35:22	U	SMPL
PB122870BS - 1	LCSW	16359	3.8222	-	7470A	PPB	06 Sep 2019 15:37:30	U	SMPL
K4679-09 - 1	WS-SB01-COMP	15	0.0066	-	7470A	PPB	06 Sep 2019 15:39:38	U	SMPL
K4679-10 - 1	WS-SB02-COMP	46	0.0138	-	7470A	PPB	06 Sep 2019 15:41:47	U	SMPL
K4679-11 - 1	WS-SB03-COMP	52	0.0152	-	7470A	PPB	06 Sep 2019 15:43:55	U	SMPL
K4679-12 - 1	WS-SB11-COMP	56	0.0161	-	7470A	PPB	06 Sep 2019 15:46:04	U	SMPL
K4680-20 - 1	T1-1-4-(0-1.5)	50	0.0147	-	7470A	PPB	06 Sep 2019 15:48:11	U	SMPL
CCV86 - 1	CCV86	21645	5.0563	-	7470A	PPB	06 Sep 2019 15:50:17	U	SMPL
CCB86 - 1	CCB86	71	0.0196	-	7470A	PPB	06 Sep 2019 15:52:24	U	SMPL
K4680-27 - 1	T2-3-(2-4)	254	0.0624	-	7470A	PPB	06 Sep 2019 15:54:33	U	SMPL
K4680-30 - 1	T3-1-4-(0-1.5)	-12	0.0003	-	7470A	PPB	06 Sep 2019 15:56:40	U	SMPL
K4694-21 - 1	T1-1-4-(0-1.5)	84	0.0227	-	7470A	PPB	06 Sep 2019 15:58:47	U	SMPL
K4694-30 - 1	T2-1-4-(0-1.5)	91	0.0243	-	7470A	PPB	06 Sep 2019 16:00:53	U	SMPL
K4694-39 - 1	T3-1-3-(0-1.5)	54	0.0157	-	7470A	PPB	06 Sep 2019 16:03:01	U	SMPL
K4695-20 - 1	T5-1-4-(0-2.5)	63	0.0178	-	7470A	PPB	06 Sep 2019 16:05:09	U	SMPL
K4695-20DUP - 1	T5-1-4-(0-2.5)DUP	51	0.0150	-	7470A	PPB	06 Sep 2019 16:07:17	U	SMPL
K4695-20MS - 1	T5-1-4-(0-2.5)MS	16146	3.7725	-	7470A	PPB	06 Sep 2019 16:09:25	U	SMPL
K4695-20MSD - 1	T5-1-4-(0-2.5)MSD	16909	3.9506	-	7470A	PPB	06 Sep 2019 16:11:34	U	SMPL
PB122871BL - 1	PBW	43	0.0131	-	7470A	PPB	06 Sep 2019 16:13:40	U	SMPL
CCV87 - 1	CCV87	22017	5.1431	-	7470A	PPB	06 Sep 2019 16:15:47	U	SMPL
CCB87 - 1	CCB87	52	0.0152	-	7470A	PPB	06 Sep 2019 16:17:54	U	SMPL
PB122871BS - 1	LCSW	16399	3.8316	-	7470A	PPB	06 Sep 2019 16:20:00	U	SMPL
K4656-03 - 1	AU-06-090519	0	0.0031	-	7470A	PPB	06 Sep 2019 16:22:07	U	SMPL
K4656-03DUP - 1	AU-06-090519DUP	26	0.0091	-	7470A	PPB	06 Sep 2019 16:24:13	U	SMPL
K4656-03MS - 1	AU-06-090519MS	16890	3.9462	-	7470A	PPB	06 Sep 2019 16:26:20	U	SMPL
K4656-03MSD - 1	AU-06-090519MSD	17311	4.0445	-	7470A	PPB	06 Sep 2019 16:28:28	U	SMPL
K4661-03 - 1	HD-01-090519	11	0.0056	-	7470A	PPB	06 Sep 2019 16:30:35	U	SMPL
K4663-03 - 1	OK-03-090519	80	0.0217	-	7470A	PPB	06 Sep 2019 16:32:43	U	SMPL
K4669-03 - 1	TR-05-090519	40	0.0124	-	7470A	PPB	06 Sep 2019 16:34:51	U	SMPL
K4692-17 - 1	SB-24-COMP	47	0.0140	-	7470A	PPB	06 Sep 2019 16:37:00	U	SMPL
K4692-18 - 1	SB-31-COMP	39	0.0122	-	7470A	PPB	06 Sep 2019 16:39:07	U	SMPL
CCV88 - 1	CCV88	23145	5.4065	-	7470A	PPB	06 Sep 2019 16:41:14	U	SMPL
CCB88 - 1	CCB88	29	0.0098	-	7470A	PPB	06 Sep 2019 16:43:21	U	SMPL
K4692-19 - 1	SB-30-COMP	33	0.0108	-	7470A	PPB	06 Sep 2019 16:45:28	U	SMPL
K4692-20 - 1	SB-29-COMP	48	0.0143	-	7470A	PPB	06 Sep 2019 16:47:34	U	SMPL
K4692-21 - 1	SB-28-COMP	58	0.0166	-	7470A	PPB	06 Sep 2019 16:49:42	U	SMPL
K4692-22 - 1	SB-25-COMP	25	0.0089	-	7470A	PPB	06 Sep 2019 16:51:49	U	SMPL
K4692-23 - 1	SB-26-COMP	34	0.0110	-	7470A	PPB	06 Sep 2019 16:53:56	U	SMPL
K4692-24 - 1	SB-27-COMP	38	0.0119	-	7470A	PPB	06 Sep 2019 16:56:03	U	SMPL
K4693-04 - 1	TP-18-S	47	0.0140	-	7470A	PPB	06 Sep 2019 16:58:12	U	SMPL
K4693-08 - 1	TP-18-D	31	0.0103	-	7470A	PPB	06 Sep 2019 17:00:21	U	SMPL
K4696-11 - 1	WS-SB04-COMP	46	0.0138	-	7470A	PPB	06 Sep 2019 17:02:30	U	SMPL
K4696-12 - 1	WS-SB05-COMP	15	0.0066	-	7470A	PPB	06 Sep 2019 17:04:37	U	SMPL
CCV89 - 1	CCV89	21117	4.9330	-	7470A	PPB	06 Sep 2019 17:06:44	U	SMPL
CCB89 - 1	CCB89	36	0.0115	-	7470A	PPB	06 Sep 2019 17:08:52	U	SMPL
K4696-13 - 1	WS-SB06-COMP	38	0.0119	-	7470A	PPB	06 Sep 2019 17:10:59	U	SMPL
K4696-14 - 1	WS-SB09-COMP	31	0.0103	-	7470A	PPB	06 Sep 2019 17:13:06	U	SMPL
K4696-15 - 1	WS-SB08-COMP	33	0.0108	-	7470A	PPB	06 Sep 2019 17:15:14	U	SMPL
PB122872BL - 1	PBW	81	0.0220	-	7470A	PPB	06 Sep 2019 17:17:21	U	SMPL
PB122872BS - 1	LCSW	16384	3.8281	-	7470A	PPB	06 Sep 2019 17:19:29	U	SMPL
K4683-05 - 1	20190667-FIELD-BLANK	1	0.0033	-	7470A	PPB	06 Sep 2019 17:21:36	U	SMPL
K4723-01 - 1	FRAC-A32780T	9104	2.1285	-	7470A	PPB	06 Sep 2019 17:23:44	U	SMPL
K4723-01DUP - 1	FRAC-A32780TDUP	9176	2.1453	-	7470A	PPB	06 Sep 2019 17:25:52	U	SMPL
K4723-01MS - 1	FRAC-A32780TMS	26103	6.0971	-	7470A	PPB	06 Sep 2019 17:28:01	U	SMPL
K4723-01MSD - 1	FRAC-A32780TMSD	26202	6.1202	-	7470A	PPB	06 Sep 2019 17:30:08	U	SMPL
CCV90 - 1	CCV90	20654	4.8249	-	7470A	PPB	06 Sep 2019 17:32:16	U	SMPL
CCB90 - 1	CCB90	24	0.0087	-	7470A	PPB	06 Sep 2019 17:34:23	U	SMPL
PB122870TB - 1	PB122802TB	114	0.0297	-	7470A	PPB	06 Sep 2019 17:36:31	U	SMPL
PB122871TB - 1	PB122835TB	21	0.0080	-	7470A	PPB	06 Sep 2019 17:38:38	U	SMPL
K4695-20LX5 - 1		4	0.0040	-	7470A	PPB	06 Sep 2019 17:40:46	U	SMPL
K4695-20A - 1		16515	3.8586	-	7470A	PPB	06 Sep 2019 17:42:53	U	SMPL
K4656-03LX5 - 1		-23	-0.0023	-	7470A	PPB	06 Sep 2019 17:45:00	U	SMPL
K4656-03A - 1		17348	4.0531	-	7470A	PPB	06 Sep 2019 17:47:08	U	SMPL
K4723-01LX5 - 1		1825	0.4291	-	7470A	PPB	06 Sep 2019 17:49:17	U	SMPL
K4723-01A - 1		26183	6.1157	-	7470A	PPB	06 Sep 2019 17:51:25	U	SMPL
CCV91 - 1	CCV91	21577	5.0404	-	7470A	PPB	06 Sep 2019 17:53:33	U	SMPL
CCB91 - 1	CCB91	-39	-0.0060	-	7470A	PPB	06 Sep 2019 17:55:41	U	SMPL