



PERCENT SOLID

Supervisor: AMANDEEP

Analyst: jignesh

Date: 9/9/2019

OVENTEMP IN Celsius(°C): 107

Time IN: 15:35

In Date: 09/06/2019

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 07:33

Out Date: 09/07/2019

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB105005

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K4749-01	MH12P5	1	1.00	2.00	2.00	100.0	DUST SAMPLE 100% SOLIDS
K4749-02	MH12P6	2	1.00	2.00	2.00	100.0	DUST SAMPLE 100% SOLIDS
K4749-03	MH12P7	3	1.00	2.00	2.00	100.0	DUST SAMPLE 100% SOLIDS
K4749-04	MH12P8	4	1.00	2.00	2.00	100.0	DUST SAMPLE 100% SOLIDS
K4749-05	MH12P8D	5	1.00	2.00	2.00	100.0	DUST SAMPLE 100% SOLIDS
K4749-06	MH12P8S	6	1.00	2.00	2.00	100.0	DUST SAMPLE 100% SOLIDS
K4749-07	MH12P9	7	1.00	2.00	2.00	100.0	DUST SAMPLE 100% SOLIDS
K4749-08	MH12Q0	8	1.00	2.00	2.00	100.0	DUST SAMPLE 100% SOLIDS
K4749-09	MH12Q1	9	1.00	2.00	2.00	100.0	DUST SAMPLE 100% SOLIDS
K4749-10	MH12Q2	10	1.00	2.00	2.00	100.0	DUST SAMPLE 100% SOLIDS
K4749-11	MH12Q3	11	1.00	2.00	2.00	100.0	DUST SAMPLE 100% SOLIDS
K4749-12	MH12Q4	12	1.00	2.00	2.00	100.0	DUST SAMPLE 100% SOLIDS
K4749-13	MH12Q5	13	1.00	2.00	2.00	100.0	DUST SAMPLE 100% SOLIDS
K4749-14	MH12Q6	14	1.00	2.00	2.00	100.0	DUST SAMPLE 100% SOLIDS
K4749-15	MH12Q7	15	1.00	2.00	2.00	100.0	DUST SAMPLE 100% SOLIDS
K4749-16	MH12Q8	16	1.00	2.00	2.00	100.0	DUST SAMPLE 100% SOLIDS
K4749-17	MH12Q9	17	1.00	2.00	2.00	100.0	DUST SAMPLE 100% SOLIDS
K4749-18	MH12R1	18	1.00	2.00	2.00	100.0	DUST SAMPLE 100% SOLIDS
K4749-19	MH12R2	19	1.00	2.00	2.00	100.0	DUST SAMPLE 100% SOLIDS

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$